



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
 website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
 Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

BID BULLETIN NO. BB-2023-013


Date : **January 25,2023**
 TO : **ALL PROSPECTIVE BIDDERS**
 FROM : **BIDS AND AWARDS COMMITTEE**
 SUBJECT : **AMENDMENT AS STATED**

Name of Project : **Supply, Delivery and Installation of Dental Chair and unit with complete accessories (TV monitor, light cure, ultrasonic scaler, led light generator handpiece, low speed handpiece, intra oral camera) ABC = Php 9,000,000.00 (IB 2023-028)**

This Bid Bulletin No. **BB-2023-013** outlines the amendments in the bidding documents for above projects as follows:

FROM	TO
1. Under Section I. Invitation to Bid	
Project Name:	Project Name:
Supply,Delivery and Installation of Dental Chair and unit with complete accessories (TV monitor, light cure, Ultrasonic scaler, fiber optic high speed hand piece , low speed hand piece, intra oral camera)	Supply,Delivery and Installation of Dental Chair and unit with complete accessories (TV monitor, light cure, Ultrasonic scaler, LED high speed hand piece , low speed hand piece, intra oral camera)
2. Under Section VII. Technical Specifications	
A. Specifications:	A. Specifications:
1. Air pressure requirement: 0.56-0.6 Mpa	1. Air pressure requirement: 0.5-0.8 MPa
10. Chair load capacity: 230 kg	10. Chair load capacity: 120-230 kg
11. With original xray intra oral film viewer:	11. With original xray intra oral film viewer:
• Size: 175mmx100mm	• Size: 175mm (±50) x100mm (±50)
12. Dimension of Doctor's Tray: 310mmx415mm	12. Dimension of Doctor's Tray: 310mm (±)50 x415mm (±)50
17. Unit will operate at 230 volts 3-Pin Power Plug/Cable, strictly at 60 hertz	17. Unit will operate at 220-240 volts 3-Pin Power Plug/Cable, strictly at 60 hertz
B. Accessories (Per Unit)	B. Accessories (Per Unit)
5. Monitor: 17" Screen , candor quadrant display setting, with continuous captures images in one display - should be 230V ;60 Hz	5. Monitor: atleast 17" Screen , candor quadrant display setting, with continuous captures images in one display - should be 220-240 V ;60 Hz
Section VI: Schedule of Requirements Thirty (30) to Forty Five (45) calendar days from receipt of Purchase Order	Section VI: Schedule of Requirements Ninety (90) to One Hundred Twenty (120) calendar days from receipt of Purchase Order

For information and guidance of all concerned.


EMMA A.MARIANO, CPA,MGM-ESP
 Chairperson, Bids and Awards Committee