



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

BID BULLETIN NO. BB-2023-062

Date : August 31, 2023
TO : ALL PROSPECTIVE BIDDERS
FROM : BIDS AND AWARDS COMMITTEE
SUBJECT : AMENDMENT AS STATED

Name of Project/s : **IB 2023-114 : One (1) Units Anesthesia Machine with Monitor, MRI compatible (ABC=Php 8,000,000.00)**

This Bid Bulletin No. **BB-2023-062** outlines the amendments in the issued bidding documents for above project as follows:


1. Under Section VII. Technical Specifications

FROM	TO
<i>Name of Project</i>	<i>Name of Project</i>
Supply and Delivery of One (1) Unit Anesthesia Machine with Monitor, MRI Compatible	Supply and Delivery of One (1) Unit Anesthesia Machine, MRI Compatible
<i>A. Specifications :</i>	<i>A. Specifications :</i>
1. Full color touchscreen display	1. Full color touchscreen display (optional), at least 3 inches x 6 inches
2. Gas Supply	2. Gas Supply
• Pipeline input range: 280kPa - 600kPa	• Pipeline input range: at least 240 kPa to 600kPa
• Modes of ventilation: VC, PC, SIMV, PSV PRO	• Modes of ventilation: VCV mode, PCV, PCV-VG, SIMV-PCV-VG, PSV CPAP+PSV, SMV-VCV
• Tidal volume range: 20 - 1500ml	• Tidal volume range: 5 - 1500ml
• Minute volume range: 0 - 99.9 L/min	• Minute volume range: <0.1 - 99.9 L/min
• I:E ratio 4:1 to 1:8	• I:E ratio at least 2:1- 1.8
8. With central brake	8. Single foot lever lock and unlock two front casters
10. Anesthetic agent delivery system compatible to vaporizer Isoflurane and Sevoflurane, 2 positions	10. Anesthetic agent delivery system compatible with Cartridge type vaporizer Isoflurane and Sevoflurane, 2 positions
• Quick change CO2 absorbent	delete
11. Anesthesia Monitor	11. Anesthesia Monitor (not MRI Compatible)

BID BULLETIN NO. BB-2023-062

FROM	TO
<i>B. Accessories (to be provided per unit)</i>	<i>B. Accessories (to be provided per unit)</i>
1. Sampling line	delete
2. D fend water trap pro	delete
3. skin and general. Temperature: Esophageal and rectal probe	3. skin and general.
9. Laryngoscope compatible to MRI (Diamond) with blades MAC 2.0, 3.0 and Miller 1.0	Laryngoscope compatible to MRI (Diamond Type) with blades MAC 1.0 2.0, 3.0 and Miller 1.0
10. Battery Back up for transport . Battery time up to 90 minutes when fully charged (2- 3 hrs charge time)	Battery Back up for transport . Battery time at least 45 minutes when fully charged (2-3 hrs charge time)

For information and guidance of all concerned.


FRANCIS S. DELA CUESTA, RN, MAN
Chairperson, Bids and Awards Committee