



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

## BID BULLETIN NO. BB-2023-075

Date : **October 27, 2023**  
TO : **ALL PROSPECTIVE BIDDERS**  
FROM : **BIDS AND AWARDS COMMITTEE**  
SUBJECT : **AMENDMENT AS STATED**

Name of Project/s : **IB 2023-136: Supply and Delivery of One (1) Unit Electric, Hospital Bed ( ABC = Php 500,000.00)**

This Bid Bulletin No. **BB-2023-075** outlines the amendments in the issued bidding documents for above project as follows:

### 1. Under Section VII. Technical Specifications

FROM	TO
<b><u>A. Specifications</u></b>	<b><u>A. Specifications</u></b>
24. Voltage/ Frequency: 220votls/50Hz	24. The unit will be operated at 230 Volts 3-Pin Power Plug / cable, strictly at 60 hertz
	<b><i>Additional Specifications</i></b>
	25. With built-in battery back up system for detection and life saving equipment
	26. No separate volt conversion
	27. Power consumption of the unit ( <i>must be clearly stipulated</i> )
	28. Mechanical parts should be of heavy-duty type
	29. Equipment should be maintainable
<b><u>B. Accessories (to be provided per unit)</u></b>	
<b>1. Bed Side Cabinet</b>	
Overall size: (LxWxH) 500mm X 495mm X 870mm	Overall size: (LXWXH) 500mm X 495-510mm X 550-870mm
Laminated side panels	Laminated side panels or powder coated metal
Protective aluminum corner extrusions	Protective aluminum/PVC corner extrusions
400mm twin castors	<b>40mm twin castors</b>



**BID BULLETIN NO. BB-2023-075**

FROM		TO
	<b>2. Over-bed Table</b>	
	Height adjustable at at least 745mm to 1055mm	Height adjustable <b>at least 610mm to 1055mm</b>
	<b>6. Patient Slide Sheet</b>	6. Patient Slide Sheet - made of soft fabric, foldable, can be wiped clean with chlorine/alcohol based solution and washable
		<i>Additional Specifications</i>
		7. One (1) Unit AVR compatible to the unit

**2. Under Section VI. Schedule of Requirements**

FROM		TO
	Thirty (30) calendar days from receipt of Purchase Order/Notice to Proceed	Thirty to Forty Five (30 - 45) calendar days from receipt of Purchase Order/Notice to Proceed

For information and guidance of all concerned.

  
**FRANCIS S. DELA CUESTA, RN, MAN**  
 Chairperson, Bids and Awards Committee