

# Republic of the Philippines DEPARTMENT OF HEALTH PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: <a href="mailto:www.pcmc.gov.ph">www.pcmc.gov.ph</a> email: <a href="mailto:officeofthedirector@pcmc.gov.ph">officeofthedirector@pcmc.gov.ph</a>
Trunkline: 8588-9900 DirectLine: 8924-0836 Fax No: 8924-0840

## NOTICE OF AWARD NOA-2022-049

### MTC OPTO -MEDIC, INC.

631 Ronquillo St., Quiapo Manila 179 Yakal St., Makati City

Tel. No.: 8813-1351/53/55, 8812-2162

Mobile No. 09178448058

E-mail Addresses: mtc.jhara.cabrera@gmail.com mtcoptomedicsales@gmail.com

#### Dear Sir / Madam:

This is to inform you that based on the result of the conducted Public Bidding for the project stated below under Invitation to Bid No. IB-2021-156 as per BAC Resolution No.R2022-01-064, your proposal was found to be responsive.

| QTY | UNIT | ITEM DESCRIPTION   | TOTAL COST (Php) |
|-----|------|--|------------------|
| 1   | unit | CRYO MACHINE Brand: KEELER/CRYOMATIC MK II (see Annex "A" for detailed specifications) | Php1,048,000.00  |

You are hereby required to provide on or before a 1 0 JAN 2022 the Performance Security in either of the following form:

| FORM OF PERFORMANCE SECURITY   | AMOUNT OF PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price) |
|--|--|
| a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank   | Five percent (5%)  |
| b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank | Php52,400.00   |
| c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.   | Thirty percent (30%)  Php314,400.00  |

## NOTICE OF AWARD NOA-2022-049 MTC OPTO -MEDIC, INC.

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

We will furnish you the approved Purchase Order (PO) upon receipt of funds from the Department of Budget and Management (DBM)

| Very truly yours,   |
|---|
| SONIA B. GONZALEZ, MD MScHSM, MPM OIC, Executive Director   |
| Conforme:   |
| This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto. |
| Authorized Signatory (Signature over printed name)  |
| Designation   |
| Date  |

| UNI  | T ITEM DESCRIPTION  | TOTAL COST      |
|------|---|-----------------|
| unit | CRYO MACHINE  | Php1,048,000.00 |
|      | BRAND: KEELER   |                 |
|      | MAKE / MODEL: KEELER/ CRYOMATIC MK II   |                 |
|      | A. Specifications   |                 |
|      | Will provide Local Trolley Stainless  |                 |
|      | 2. Gas Specification: Medical grade Nitrous Oxide(N₂O) or Medical Grade Carbon dioxide (CO₂) in Non- syphonic cylinders ,         |                 |
|      | 3. Operating range: 3100-4480 kPA (450-650PSI/31-45 Bar)  |                 |
|      | 4. Maximum Cylinder Pressure:8275 kPa(1200PSI/83 Bar)   |                 |
|      | 5. Input Voltage Range: 100-240Vac (50/60Hz)  |                 |
|      | 6. Dimension: Width: 350 mm (14") Depth: 200mm (8") Height: 190mm (7.5")  |                 |
|      | 7. Weight: 4.5 kg (10lbs)   |                 |
|      | 8. Equipment Classification: Class 1, Type BF (Applied Part)  |                 |
|      | 9. Protection Against Ingress Console IpxO Footswitch IPX7  |                 |
|      | 10. Graphical LCD used to provide system information to the user; such as probe information, gas cylinder status and freeze time. |                 |
|      | 11. Storage: +10°C to +55°C Operating:+10°C to +40°C  |                 |
| 1.43 | 12. Temperature Range: -40°C to +70°C   |                 |
|      | 13. Relative Humidity: Transport: 10% to 95% Storage: 10%to 95% Operation: 30% to 90%   |                 |
|      | 14. Atmospheric Pressure: Transport: 500hPa to 1060hPa Storage: 700hPa to 1060hPa Operation: 800hPa to 1060hPa                    |                 |
|      | 15. The unit operates at 100-240VAC (50/60Hz)   |                 |
| 11/  | 16. Power consumption: 100 VA   |                 |
| . 9  | 17. Mechanical Parts are made of heavy duty durable type  |                 |
|      | 18. Equipment is maintainable and serviceable   |                 |
|      | B. ACCESSORIES to be provided per unit  |                 |
|      | 1. One (1) Unit AVR compatible to the unit  |                 |
|      | Brand: STAVOL   |                 |
| v    | Model: 1000VA   |                 |
| d    | 2. Retina Probe #2: 2509-P-8030-Compact Retinal Probe   |                 |
|      | 3. Footswitch   |                 |
|      | C.WARRANTY  |                 |
|      | With two (2) years warranty on services, parts and accessories  | ¥               |
|      | With free quarterly preventive maintenance and calibration  |                 |
|      | E. With duly notarized certificate from Distributor - MTC - Optomedic Inc.  |                 |

| QTY | UNIT                   | ITEM DESCRIPTION   | TOTAL COST      |
|-----|------------------------|--|-----------------|
| 1   | unit                   | CRYO MACHINE   | Php1,048,000.00 |
|     |                        | i. Parts, accessories and consumables are readily available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. Service Center: 3rd Floor OHI Bldg. 179 Yakal St., San Antoio Village, Makati City. Tel Nos: 8813-1351/8813-1355/ Fax No. 8812-2162. Email Address: technical@mtceyecare.com/mtcservice.karen@gmail.com. Contact Person: Karen Parungao / Emilyn Frias |                 |
|     |                        | ii. Ensures availability of competence in-house technical specialist in handling and providing technical support as well as maintenance of the equipment being offered; and  |                 |
|     | Variable of the second | iii. Assures conduct training for proper operation and maintenance to end-users of the equipment upon delivery;  |                 |
|     |                        | iv.shall provide replacement/back-up unit while the delivered unit is being repaired.  |                 |
|     |                        | F. With duly notarized certificate from Principal Manufacturer - KEELER OPTHALMIC INSTRUMENTS  |                 |
|     |                        | Certifies that:  |                 |
|     |                        | i. All the terms and conditions stated in bidding documents per IRR of RA 9184 and corresponding contract for the project will be honored including in the event that a change of distributorship will occurs during the duration of the contract up to the warranty and preventive maintenance period.  |                 |
|     |                        | ii. The expected useful life of the equipment is five (5) to eight (8) years under normal use:   |                 |
|     |                        | <ul> <li>iii. Guarantees availability of all spare parts, accessories and consumables at<br/>least for the next five (5) years from testing, commissioning, acceptance and<br/>delivery;</li> </ul>  |                 |
|     |                        | iv. Guarantees provision of the certificate from the manufacturer for the consumer guidelines regarding disposal of the equipment(information about how and where the used and decomissioned products / parts can be returned for recycling and or dispoal)  |                 |
| 9   |                        | DELIVERY PERIOD: Forty Five - Sixty (45 - 60) Calendar Days  |                 |

| Conforme:                                      |      |
|--|------|
| Authorized Signatory (Signature over printed a | name |
| Designation/Position                           |      |
| Date   |      |

## **CERTIFICATION**

This is to certify that MTC Opto-Medic, Inc. guarantees the provision of the certificate from the manufacturer for the consumer guidelines regarding disposal of the equipment (Information about how and where the used and decommissioned products/parts can be returned for recycling and/or disposal.

This certificate was issued in compliance with the requirements for Supply and Delivery of Various Medical Equipments – Supply and Delivery of Cryo Machine (Rebid) IB-2021-156.

CESAR E. ROXAS

Name of Company of Authorized Representative

January 6, 2022

Date

MTC Opto-Medic, Inc.
Name of Company