



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

BID BULLETIN NO. BB-2020-013

Date : February 03, 2020
TO : ALL PROSPECTIVE BIDDERS
FROM : BIDS AND AWARDS COMMITTEE
SUBJECT : AMENDMENT AS STATED
Name of Project : **Supply and Delivery of Various Medical Equipment for Perinatology under Invitation to Bid Nos. IB-2020-030 to IB-2020-032**
Total ABC = Php 4,100,000.00

This Bid Bulletin No. **BB-2020-013** outlines the amendments in the bidding documents for above projects as follows:


I. Under Section VI. Schedule of Requirements

FROM	TO
Delivery Period Forty-Five to Sixty (45 - 60) calendar days	Delivery Period Sixty to Ninety (60 - 90) calendar days

II. Under Section VII. Technical Specifications

FROM	TO
IB-2020-032 (Operating Room Table)	
A. Specifications:	A. Specifications:
7. Load capacity of 130 to 200 kgs	7. Load capacity: Minimum of 130 kgs
9. Max Table Length: 2050mm (including head and leg plates)	9. Minimum Table Length: 1900mm (including head and leg plates)
10. Max Table Width: 480mm (back plate only)	10. Minimum Table Width: 480mm (back plate only)
11. Lowest Table Level: 690mm	11. Lowest Table Level: 520mm to 690mm
12. Highest Table Level: 1040mm	12. Highest Table Level: $\geq 1040\text{mm}$
13. Lifting Distance: 350mm	Delete
18. Upward/Downward for Head Plate: $\geq 90^\circ / \geq 90^\circ$	18. Upward/Downward for Head Plate: $\geq 40^\circ / \geq 90^\circ$
IB-2020-030 to IB-2020-032	
The unit will be operated at 230 volts, strictly at 60 hertz	The unit should comply with PCMC's power source at 230 volts , strictly at 60 hertz

For information and guidance of all concerned.


SHEILA ANN D. MASANGKAY, MD, MHSA
Chairperson, Bids and Awards Committee

r

