



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
 website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
 Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

BID BULLETIN NO. BB-2020-040

Date : **July 23, 2020**
 TO : **ALL PROSPECTIVE BIDDERS**
 FROM : **BIDS AND AWARDS COMMITTEE**
 SUBJECT : **AMENDMENT AS STATED**
 Name of Project : **Supply and Delivery of One (1) Unit Portable Ultrasound Machine under Invitation to Bid No. IB-2020-080**
ABC = Php 2,000,000.00

This Bid Bulletin No. **BB-2020-040** outlines the amendments in the bidding documents for above projects as follows:

Under Section VII. Technical Specifications

FROM	TO
ITEM DESCRIPTION	ITEM DESCRIPTION
A. SPECIFICATIONS:	A. SPECIFICATIONS:
Table mounted plus movable stand with lock	Table mounted laptop type with trolley
15 inches high resolution LED/LCD monitor	At least 15 inches high resolution LCD/LED Monitor
Dashboard connectivity	DICOM connectivity
Audible and visual alarms	Must have Audible and visual alarms
TDI	delete
Freehand 3D	delete
Continuous Wave Doppler	delete
Stress Echo	delete
➤ Transducer	➤ Transducer
a. Curvilinear Transducer Application: Adult Abdomen, gynecology, obstetrics, 1.76MHz, FOV: 75 degrees.	a. Curvilinear Transducer Application: Adult Abdomen, gynecology, obstetrics, 1.7-6MHz, FOV: 75 degrees.
b. Linear Transducer Application : Small parts, vascular, pediatrics, superficial musculoskeletal, nerve, 3.5-16MHz, FOV: 38mm	b. Linear Transducer Application : Small parts, vascular, pediatrics, superficial musculoskeletal, nerve, 3.5-16MHz, FOV: 38 degrees
B. Accessories	A. SPECIFICATIONS:
Uninterruptible Power Supply (UPS) (compatible to the equipment power requirement) and proper grounding when needed	Uninterruptible Power Supply (UPS), 1 hour (compatible to the equipment power requirement) and proper grounding when needed,

For information and guidance of all concerned.

MARIA ROSARIO S. CRUZ, MD.D
 Chairperson, Bids and Awards Committee

