



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

BID BULLETIN NO. BB-2020-045

Date : August 20, 2020
TO : ALL PROSPECTIVE BIDDERS
FROM : BIDS AND AWARDS COMMITTEE
SUBJECT : AMENDMENT AS STATED

Name of Project : Supply and Delivery of Portable 2D Echo Machine (ABC =
Php2,500,000.00) IB 2020-084

This Bid Bulletin No. **BB-2020-045** outlines the amendments in the bidding documents for above projects as follows:

1. Under Section VII. Technical Specifications

FROM	TO
<i>A. Specifications:</i>	<i>A. Specifications:</i>
7. Transducer Probes	7. Transducer Probes
• At least 3 with the following frequencies:	• At least 3 with the following frequencies:
1. 1 - 4 Mhz phased / cardiac for use on adult / big pedia	1. phased/cardiac transducer probe for use on adult/big pedia with frequency range between 1-4 MHz
2. 2 - 8 MHz phased / cardiac for use on small pedia / neonate	2. phased/cardiac transducer probe for use on small pedia/neonate with frequency range between 2 - 8 MHz
3. 6 - 13 MHz linear for vascular or 1 - 5 MHz convex for fetal echo (end-user's choice)	3. linear transducer probe for vascular imaging with frequency range between 6-13 MHz.

For information and guidance of all concerned.


MARIA ROSARIO S. CRUZ, MD
Chairperson, Bids and Awards Committee

v.