

## Republic of the Philippines DEPARTMENT OF HEALTH PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100 website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

## BID BULLETIN NO. BB-2020-046

Date

: August 20, 2020

TO

: ALL PROSPECTIVE BIDDERS

FROM

: BIDS AND AWARDS COMMITTEE

**SUBJECT** 

: AMENDMENT AS STATED

Name of Project

Supply and Delivery of Three (3) Units Suction Machine

(ABC = Php240,000.00 @ Php 80,000.00 per unit) IB 2020-086

This Bid Bulletin No. BB-2020-046 outlines the amendments in the bidding documents for above projects as follows:

## 1. Under Section VII. Technical Specifications

| FROM   | ТО  |
|--|---|
| A. Specifications:   | A. Specifications:  |
| Quiet and effective operation due to the<br>unique design of WOBL piston | <ul> <li>Quiet and effective operation due to the unique design<br/>of WOBL piston or its equivalent</li> </ul> |
| Maximum vacuum pressure : 600 - 700 mm-<br>Hg                            | Maximum vacuum pressure : 600 - 800 mm-Hg   |
| Sound level : 54 dBA or lower  | • Sound level : 60 dBA or lower   |
| • Weight : 14-30kgs  | • Weight : 14- <b>35kgs</b>   |
| B. Accessories per unit  | B. Accessories per unit   |
| Suction Bottles (one each)   | Two (2) Polycarbonate Suction Bottles - at least 3<br>liters per bottle   |

For information and guidance of all concerned.

MARIA ROSARIO S. CRUZ, MD

maries

Chairperson, Bids and Awards Committee



