



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

BID BULLETIN NO. BB-2020-061

Date : **October 21, 2020**
TO : **ALL PROSPECTIVE BIDDERS**
FROM : **BIDS AND AWARDS COMMITTEE**
SUBJECT : **AMENDMENT AS STATED**
Name of Project : **Supply and Delivery of Portable 2D Echo Machine** (ABC =
Php2,500,000.00) IB 2020-113

This Bid Bulletin No. **BB-2020-061** outlines the amendments in the bidding documents for above projects as follows:

1. Under Section VII. Technical Specifications

| FROM | TO |
|--|---|
| <i>A. Specifications:</i> | <i>A. Specifications:</i> |
| 6. Transducer Probes | 6. Transducer Probes |
| 1. phased/cardiac transducer probe for use on adult/big pedia with frequency range between 1-4 MHz | 1. phased/cardiac transducer probe for use on adult/big pedia with frequency range between 1-5 MHz |
| 2. phased/cardiac transducer probe for use on small pedia/neonate with frequency range between 2 - 8 MHz | 2. phased/cardiac transducer probe for use on small pedia/neonate with frequency range between 2 - 9 MHz |

For information and guidance of all concerned.

MARIA ROSARIO S. CRUZ, MD
Chairperson, Bids and Awards Committee