



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

## BID BULLETIN NO. BB-2021-045

Date : **June 29, 2021**

TO : **ALL PROSPECTIVE BIDDERS**

FROM : **BIDS AND AWARDS COMMITTEE**

SUBJECT : **AMENDMENT AS STATED**

Name of Project : **Supply and Delivery of C-Arm Xray Machine including table under Invitation to Bid No. IB-2021-083**  
**Total ABC = Php 20,000,000.00**

This Bid Bulletin No. **BB-2021-045** outlines the amendments in the bidding documents for above projects as follows:

### A. Under Section VII. Technical Specifications

FROM	TO
I. Specifications:	I. Specifications:
E. Digital Subtraction Angiography / Roadmap: Parameters	E. Digital Subtraction Angiography / Roadmap: Parameters
93. Image acquisition of 0.5 to 30 f/s	93. Image acquisition of 0.5 to at least 30 f/s
III. Warranty:	III. Warranty:
> With at least three (3) years on services, parts (including tube and detector) and accessories	> With at least three (3) years on services, parts (including tube and detector) and one (1) year on accessories

### B. Under Section VI. Schedule of Requirements

FROM	TO
PCMC Requirement DELIVERY PERIOD	PCMC Requirement DELIVERY PERIOD
Forty-Five to Sixty (45 -60) calendar days	Sixty to Ninety (60 – 90) calendar days

BID BULLETIN NO. BB-2021-045

Supply and Delivery of C-Arm Xray Machine including table under Invitation to Bid No. IB-2021-083

C. Under Section VIII. Checklist of Technical and Financial Documents

FROM	TO
I. TECHNICAL COMPONENT ENVELOPE	I. TECHNICAL COMPONENT ENVELOPE
Class "A" Documents	Class "A" Documents
Technical Documents	Technical Documents
9.ii. Duly notarized Certificate from Principal Manufacturer/ Distributor ... model being bided is currently in use in the local/international market with <i>at least two (2) installations</i> in the Philippines, preferably within NCR area ...;	9.ii. Duly notarized Certificate from Principal Manufacturer/ Distributor ... model being bided is currently in use in the local/international market with <u>at least one (1) installation</u> in the Philippines, preferably within NCR area ...;

For information and guidance of all concerned.

  
MARIA ROSARIO S. CRUZ, MD  
Chairperson, Bids and Awards Committee