



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

BID BULLETIN NO. BB-2021-054

Date : August 09, 2021
TO : ALL PROSPECTIVE BIDDERS
FROM : BIDS AND AWARDS COMMITTEE
SUBJECT : AMENDMENT AS STATED
Name of Project : **Supply, Delivery and Installation of Two (2) units HEATING AND COOLING MACHINE**
under Invitation to Bid No. **IB-2021-103**
ABC/ unit = **Php1,850,000.00**
Total ABC=**Php3,700,000.00**

This Bid Bulletin No. **BB-2021-054** outlines the amendments in the bidding documents for above projects as follows:

Under Section VI. Schedule of Requirements

FROM	TO
PCMC Requirement DELIVERY PERIOD	PCMC Requirement DELIVERY PERIOD
Forty-Five (45) to Sixty (60) calendar days	Forty-Five (45) to Ninety (90) calendar days

Under Section VII. Technical Specifications

FROM	TO
ITEM DESCRIPTION	ITEM DESCRIPTION
A. SPECIFICATIONS:	A. SPECIFICATIONS:
3. Weight: Approximate 25kg (empty tank)	Weight: Approximate 25kg (empty water tank)
4. Nominal temperature: 5°C-40°C	Nominal temperature: 4°C-42°C
4.1 It features a temperature range of at least 5°C-40°C so it can cover all areas of application by means of thermo-electrical elements and continuously pumped through the water pad by a centrifugal pump.	It features a temperature range of at least 4°C-42°C so it can cover all areas of application by means of thermo-electrical elements and continuously pumped through the water pad by a centrifugal pump.
5.2 Quick response fuse - cuts all power when heating plate temperature reaches	Quick response fuse - cuts all power when heating plate temperature reaches above 42°C or safety shutdown beyond 42°C
6. Alarm level: ≤ 55 dBA	Alarm level: ≤ 60 dBA
11. Coolant: Water-glycol mixture	Coolant: Water-glycol mixture or its equivalent technology
19. The unit has a connection of two (2) water pads.	The unit has a connection for at least two (2) water pads.
24. Power consumption: 440 W	Power consumption: 440 W - 800 W
C. CONSUMABLES to be Provided per Unit	C. CONSUMABLES to be Provided per Unit
1. Multi-use Water Pad- cover (2 pcs)	Multi-use Water Pad - Adult (2 pcs)
2. Multi-use Water Pad - Intensive (3 pcs)	Multi-use Water Pad - Pedia (3 pcs)
3. Multi-use Water Pad -Baby (2 pcs)	Multi-use Water Pad - Infant (2 pcs)



BID BULLETIN NO. BB-2021-054

Under Section VIII. Checklist of Financial and Technical Document

FROM	TO
PCMC Requirement Technical Document	PCMC Requirement Technical Document
9. ii. Duly notarized Certificate from Principal Manufacturer/ Distributor that the brand has been present for at least five (5) years in the local/ international market and model being bided is currently in use in the local/international market with at least two (2) installations in the Philippines , preferably within NCR area (Please specify and submit supporting documents);	9. ii. Duly notarized Certificate from Principal Manufacturer/ Distributor that the brand has been present for at least five (5) years in the local/ international market and model being bided is currently in use in the local/international market with at least one (1) installation in the Philippines , preferably within NCR area (Please specify and submit supporting documents);

For information and guidance of all concerned.


MARIA ROSARIO S. CRUZ, MD
Chairperson, Bids and Awards Committee 