



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

BID BULLETIN NO. BB-2021-066

Date : **September 14, 2021**

TO : **ALL PROSPECTIVE BIDDERS**

FROM : **BIDS AND AWARDS COMMITTEE**

SUBJECT : **AMENDMENT AS STATED**

Name of Project : **Supply and Delivery of Ultrasound Machine with Microconvex probe and Endocavity (Midrange to High-End) under Invitation to Bid No. IB-2021-125**
Total ABC = Php 7,000,000.00

This Bid Bulletin No. **BB-2021-066** outlines the amendments in the bidding documents for above projects as follows:

Under Section VII. Technical Specifications

FROM	TO
I. Specifications:	I. Specifications:
C. Transducers	C. Transducers
Artifact Reduction Technology	Artifact Reduction Technology
2. System should be capable of displaying 4 live images on touchscreen with different speckle, contrast, and edge parameter settings, and should further allow the user to choose from among them to be displayed on the main monitor, to reduce amount of time spent on adjusting the image parameters. Reduces speckle and enhances contrast to provide a realistic tissue presentation and patient-specific processing that adapts to differences in tissue.	2. System should be capable of displaying 4 live images on touchscreen with different speckle, contrast, and edge parameter settings, and should further allow the user to choose from among them to be displayed on the main monitor, to reduce amount of time spent on adjusting the image parameters. Reduces speckle and enhances contrast to provide a realistic tissue presentation and patient-specific processing that adapts to differences in tissue. OR system should be capable of displaying four (4) quadrants on the main monitor easily controlled with a control knob, with different speckle, contrast and edge parameter settings and should further allow the user to choose from among them, to reduce amount of time spent on adjusting the image parameters.

For information and guidance of all concerned.


MARIA ROSARIO S. CRUZ, MD
Chairperson, Bids and Awards Committee *My*

