



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

## BID BULLETIN NO. BB-2021-067

Date : September 14, 2021  
TO : ALL PROSPECTIVE BIDDERS  
FROM : BIDS AND AWARDS COMMITTEE  
SUBJECT : AMENDMENT AS STATED  
Name of Project : **Supply and Delivery of Ultrasound Machine with curvilinear transducer (portable laptop type with trolley) under Invitation to Bid No. IB-2021-121**  
**Total ABC = Php 7,000,000.00**

This Bid Bulletin No. **BB-2021-067** outlines the amendments in the bidding documents for above projects as follows:

### Under Section VII. Technical Specifications


FROM		TO	
I. Specifications:		I. Specifications:	
A. Main Equipment		A. Main Equipment	
3	Approximately 5 kg-7kg weight (with batteries/accessories)	3	At least 5 kg-8kg weight (with batteries/accessories)
6	8-segment TGC control	6	At least 6 to 8-segment TGC control
8	Built-in mounted back-up battery with continuous use for at least 90 minutes or higher	8	Built-in mounted back-up battery with continuous use for at least 60 minutes or higher
C. Imaging Modes		C. Imaging Modes	
3	M-Mode/Color M-Mode	3	M-Mode/Color M-Mode or its equivalent
4	Anatomical M-Mode	4	Anatomical M-Mode or its equivalent
5	Curved Anatomical M-Mode	5	Curved Anatomical M-Mode or its equivalent
D. Data Storage and Transfer		D. Data Storage and Transfer	
5	> Boot up from Standby: at least 7 to 10seconds	5	> Boot up from Standby: at least less than 30 seconds
E. Transducers		E. Transducers	
	Phased Array Transducer		Phased Array Transducer
3	Footprint : At least 15mm	3	Footprint : At least 9 mm to 15mm
II. Accessories per unit:		II. Accessories per unit:	
4	> Material Type: Magnesium Alloy with locking wheel	4	> Material Type: Magnesium Alloy with locking wheel or its equivalent

**BID BULLETIN NO. BB-2021-067**

**Supply and Delivery of Ultrasound Machine with curvilinear transducer (portable laptop type with trolley) under  
Invitation to Bid No. IB-2021-121**

FROM		TO	
> Gel warmer and probe holder: At least five (5) holders		> Probe holder: At least four (4) holders	
> Height must be adjustable : Up to 25cm		> Height must be adjustable : Up to 20 to 25cm	

For information and guidance of all concerned.

  
**MARIA ROSARIO S. CRUZ, MD**  
Chairperson, Bids and Awards Committee *MD*