



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

BID BULLETIN NO. BB-2021-068

Date : **September 14, 2021**

TO : **ALL PROSPECTIVE BIDDERS**

FROM : **BIDS AND AWARDS COMMITTEE**

SUBJECT : **AMENDMENT AS STATED**

Name of Project : **Supply and Delivery of Ultrasound Machine with multi-frequency transducer (portable laptop type with trolley) under Invitation to Bid No. IB-2021-122**
Total ABC = Php 4,000,000.00

This Bid Bulletin No. **BB-2021-068** outlines the amendments in the bidding documents for above projects as follows:

Under Section VII. Technical Specifications

FROM		TO	
I. Specifications:		I. Specifications:	
A. Ultrasound System Structure		A. Ultrasound System Structure	
Others		Others	
2	Built-in mounted back-up battery with continuous use for at least 90 minutes or higher	2	Built-in mounted back-up battery with continuous use for at least 60 minutes or higher
3	Fully charged battery should lasts more than 24 hours when in standby mode		DELETE
II. Accessories per unit:		II. Accessories per unit:	
4	> Material Type: Magnesium Alloy with locking wheel	4	> Material Type: Magnesium Alloy with locking wheel or its equivalent
	> Gel warmer and probe holder: At least five (5) holders	4	Probe holder: At least four (4) holders
	> Height must be adjustable : Up to 25cm		> Height must be adjustable : Up to 20 to 25cm

For information and guidance of all concerned.


MARIA ROSARIO S. CRUZ, MD
Chairperson, Bids and Awards Committee