

## Republic of the Philippines DEPARTMENT OF HEALTH PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: <a href="mailto:www.pcmc.gov.ph">www.pcmc.gov.ph</a> email: <a href="mailto:officeofthedirector@pcmc.gov.ph">officeofthedirector@pcmc.gov.ph</a>
Trunkline: 8588-9900 DirectLine: 8924-0836 Fax No: 8924-0840

## BID BULLETIN NO. BB-2021-073

Date

: September 29, 2021

TO

: ALL PROSPECTIVE BIDDERS

**FROM** 

: BIDS AND AWARDS COMMITTEE

**SUBJECT** 

: AMENDMENT AS STATED

Name of Project

Supply and Delivery of Nine (9) Units Incubator with Weighing Scale

(ABC = Php 23,400,000.00 @ 2,600,000.00 / unit) IB 2021-135

This Bid Bulletin No. <u>BB-2021-073</u> outlines the amendments in the bidding documents for above projects as follows:

## 1. Under Section VII. Technical Specifications

FROM	ТО
A. Specifications:	A. Specifications:
7. Full color touch display: 10.4 inch LCD with visual and audio alarm	Full color touch display: at least 10 inches LCD with visual and audio alarm
9. Oxygen input flow rate 5 to 15 liters/min or oxygen concentration range 25 to 70%	Oxygen input flow rate 2 to 15 liters/min or oxygen concentration range 25 to 70%
16. Electronic tilting mechanism: 0 to 15 degrees in both trendelenburg and reverse tredelenburg positions	Electronic or manual mechanism: 0 to 15 degrees in both trendelenburg and reverse trendelenburg positions

For information and guidance of all concerned.

MARIA ROSARIO S. CRUZ, MD

Textres

Chairperson, Bids and Awards Committee