



BID BULLETIN NO. BB-2021-075

Date : **October 04, 2021**

TO : **ALL PROSPECTIVE BIDDERS**

FROM : **BIDS AND AWARDS COMMITTEE**

SUBJECT : **AMENDMENT / ADDENDUM AS STATED**

Name of Project : **Supply and Delivery of NINE (9) sets STRETCHER BED with IV POLE and OXYGEN HOLDER under Invitation to Bid No. IB-2021-140**
ABC/ unit = Php300,000.00
Total ABC=Php2,700,000.00

This Bid Bulletin No. **BB-2021-075** outlines the amendments in the bidding documents for above projects as follows:


Under Section VI. Schedule of Requirements

FROM	TO
PCMC Requirement DELIVERY PERIOD	PCMC Requirement DELIVERY PERIOD
Forty-Five to Sixty (45-60) calendar days upon site readiness (PCMC Eight Storey Building) / from the time that the specific site is ready for the delivery of the units.	Sixty to Ninety (60-90) calendar days upon site readiness (PCMC Eight Storey Building) / from the time that the specific site is ready for the delivery of the units.

Under Section VII. Technical Specifications

FROM	TO
ITEM DESCRIPTION	ITEM DESCRIPTION
A. SPECIFICATIONS:	A. SPECIFICATIONS:
2. With height adjustment within 58.5-90.5cm, hydraulic	Height adjustment: Lowest: 53cm-58cm Highest: 89cm-93cm
5. With 5 casters: 4 casters wheels 200mm and central caster is adjustable from 4 point with lock	5 Casters: 4 caster wheels 200mm and central caster is adjustable with lock. All casters are secured with brake lock.
7. With mattress size: 192 x 55 x 5 cm	Mattress size must adhere with the manufacturer's standard bed size.
9. With 2 section waterproof mattress platform	With 2 section waterproof mattress platform where mattress is placed.
11. With patient trap	With patient strap.
15. (additional specification)	Mattress thickness is within 3-4 inches

For information and guidance of all concerned.


MARIA ROSARIO S. CRUZ, MD
Chairperson, Bids and Awards Committee 