



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100

website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph

Trunkline: 8588-9900 DirectLine: 8924-0836 Fax No: 8924-0840

BID BULLETIN NO. BB-2021-087

Date : **November 29, 2021**
TO : **ALL PROSPECTIVE BIDDERS**
FROM : **BIDS AND AWARDS COMMITTEE**
SUBJECT : **AMENDMENT AS STATED**

Name of Project : **Supply and Delivery of Two (2) Units Cardiac Monitor**
(ABC = Php1,000,000.00 @ 500,000.00 / unit) IB 2021-166

This Bid Bulletin No. **BB-2021-087** outlines the amendments in the bidding documents for above projects as follows:

1. Under Section VII. Technical Specifications

FROM	TO
<i>A. Specifications:</i>	<i>A. Specifications:</i>
2.2 Non-invasive Blood Pressure (NIBP)	2.2 Non-invasive Blood Pressure (NIBP)
<ul style="list-style-type: none">• NIBP Monitoring, 3 measurement modes: Auto, Manual and customize series (BP sequence)	<ul style="list-style-type: none">• Able to get BP monitoring by manual, auto and customized meaning to be able to do readings whether manual or auto in a defined interval and still be able to do STAT reading as needed
2.3 Invasive Blood Pressure with waveform	2.3 Invasive Blood Pressure with waveform
<ul style="list-style-type: none">• Pressure monitoring: ABP, ART, CVP, PAP, PPV	<ul style="list-style-type: none">• Measures Arterial Blood Pressure, Pulmonary Artery Pressure, Central Venous Pressure, Right Arterial Pressure, Left Arterial Pressure and Intracranial Pressure
<ul style="list-style-type: none">• Seven (7) cardiac output module inclusive of thermodilution accessories interchangeable across all patient monitors	<ul style="list-style-type: none">• Optional



BID BULLETIN NO. BB-2021-087

	2.4 Pulse Oximetry/SPO2 with waveform	2.4 Pulse Oximetry/SPO2 with waveform
	SPO2 Monitoring, multisite Y-probe or wrap around rubber glove, shockproof, waterproof SPO2 sensor, Pleth waveform - visual indicator of patient's pulse, Perfusion indicator - numerical value for the pulsatile portion	<ul style="list-style-type: none">●The numerical value for the pulsatile portion may be optional
	3. DATA STORAGE	
	With at least 72 hours history saving, recording with at least 400 snapshots	<ul style="list-style-type: none">●With at least 72 hours history saving and all values being measured are recorded, retrievable but does not necessarily screenshot

For information and guidance of all concerned.



MARIA ROSARIO S. CRUZ, MD
Chairperson, Bids and Awards Committee