



Republic of the Philippines  
**DEPARTMENT OF HEALTH**  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
 website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
 Trunkline: 8588-9900 DirectLine: 8924-0836 Fax No: 8924-0840

**BID BULLETIN NO. BB-2022-016**


Date : **March 14, 2022**  
 TO : **ALL PROSPECTIVE BIDDERS**  
 FROM : **BIDS AND AWARDS COMMITTEE**  
 SUBJECT : **AMENDMENT AS STATED**  
 Name of Project : **Supply and Delivery of Seventy Nine (79) units of Syringe Pump  
 (ABC = Php 5,293,000.00) IB 2022-060**

This Bid Bulletin No. **BB-2022-016** outlines the amendments in the bidding documents for above projects as follows:

**1. Under Section VII. Technical Specifications**

	FROM	TO
	<b>A. Specifications:</b>	<b>A. Specifications:</b>
1	1. Can be used with syringes of any brand (universally compatible no need for recalibration) as small as 3ml, 5ml and 10ml size, can also be used larger sizes syringes such as 20ml, 30ml, 50/60ml	1. Can be used with syringes of any brand (universally compatible) as small as 2ml, 5ml and 10ml size, can also be used larger sizes syringes such as 20ml, 30ml, 50/60ml
2	2. Minimal flow rates of 0.1ml/hr with the rate form 0.1ml/hr to 1800ml/hr	2. Minimal flow rates of 0.1ml/hr with the rate form 0.1ml/hr from 1200ml/hr to 1800ml/hr
3	3. Data lock, locks the keypad and prevent unauthorized access to safety relevant function	3. Data lock/Screen lock: locks the keypad and prevent unauthorized access to safety relevant function
4	4. Capacity: 2,000 to 20,000 events, Approximately 1 year of event log storage in normal size	4. Capacity: Atleast minimum of 2,000 to 20,000 event or better can be stored and playback, Approximately 1 year of event log storage in normal size
5	5. Free flow clamp, prevents free flow on initial setting up and/or on changing of syringes	5. Barrel clamp to hold the syringe in place, prevents free flow on initial setting up and/or on changing of syringes

For information and guidance of all concerned.

  
**MARIA ROSARIO S. CRUZ, MD**  
 Chairperson, Bids and Awards Committee