



BID BULLETIN NO. BB-2022-021

Date : March 22, 2022
TO : ALL PROSPECTIVE BIDDERS
FROM : BIDS AND AWARDS COMMITTEE
SUBJECT : AMENDMENT AS STATED
Name of Project : Supply, Delivery and Installation of Two (2) units Operating Light (LED dual ceiling mounted and pendant surgical monitor) under Invitation to Bid No. IB-2022-072
ABC = Php 4,600,000.00 (Php2,300,000.00/unit)

This Bid Bulletin No. **BB-2022-021** outlines the amendments on the following which shall form an integral part of the Bidding Documents for above projects as follows:


Under Section I. Invitation to Bid
II. Bid Data Sheet
III. Special Conditions of Contract
VI. Schedule of Requirements
VII. Technical Specifications
Price Schedule

FROM	TO
ITEM DESCRIPTION	ITEM DESCRIPTION
A. SPECIFICATIONS:	A. SPECIFICATIONS:
Supply, Delivery and Installation of Operating Room Light (LED dual ceiling mounted and pendant surgical monitor)	Supply, Delivery and Installation of Operating Room Light (LED dual ceiling mounted)

Under Section VII. Technical Specifications

FROM	TO
ITEM DESCRIPTION	ITEM DESCRIPTION
A. SPECIFICATIONS:	A. SPECIFICATIONS:
With two lamps, with at least one lamp capable of at least 140,000 lux	With two lamps, with at least 160,000 lux
B. ACCESSORIES PER UNIT	B. ACCESSORIES PER UNIT
Uninterruptable Power Supply (UPS) of at least 2KVA (compatible to the equipment power requirement) and proper grounding when needed	Uninterruptable Power Supply (UPS) of at least 1KVA (compatible to the equipment power requirement) and proper grounding when needed

For information and guidance of all concerned.


MARIA ROSARIO S. CRUZ, MD, FPPS, FPSN, FPNSP
Chairperson, Bids and Awards Committee

