



## BID BULLETIN NO. BB-2022-037

Date : May 24, 2022  
TO : ALL PROSPECTIVE BIDDERS  
FROM : BIDS AND AWARDS COMMITTEE  
SUBJECT : AMENDMENT AS STATED  
Name of Project : **Supply and Delivery of Four (4) units Indbator with Weighing Scale**  
(ABC = Php 6,400,000.00 @ 1,600,000.00 /unit) IB 2022-086

This Bid Bulletin No. **BB-2022-036** outlines the amendments in the bidding documents for above projects as follows:

### 1. Under Section VII. Technical Specifications

| FROM   | TO   |
|--|--|
| <i>A. Specifications:</i>  | <i>A. Specifications:</i>  |
| 23. The unit will be operated at 230 Volts 3-Pin Power Plug / cable, strictly at 60 hertz. | 23. The unit will be operated at 230 Volts 3-Pin Power Plug / cable, strictly at 60 hertz. <b><u>(If offer is 220 Volts, UPS to be provided should have an ouptput voltage of 220 Volts)</u></b> |

For information and guidance of all concerned.

EMMA A. MARIANO, CPA, MGM-ESP  
Chairperson, Bids and Awards Committee

