



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunkline: 8588-9900 DirectLine: 8924-0836 Fax No: 8924-0840

**BID BULLETIN NO. BB-2022-043**

Date : July 19, 2022

TO : ALL PROSPECTIVE BIDDERS

FROM : BIDS AND AWARDS COMMITTEE

SUBJECT : AMENDMENT /ADDENDUM AS STATED

Name of Project : Supply, Delivery and Installation of One (1) unit WASHER DIS-INFECTOR  
under Invitation to Bid No. IB-2022-109  
Total ABC = Php5,000,000.00

This Bid Bulletin No. BB-2022-043 outlines the amendments and addendum in the bidding documents for the above project as follows:

**Under Section VII. Technical Specifications**

AMENDMENTS:		
	FROM	TO
<b>A. SPECIFICATIONS of Main Equipment</b>		
2	Exterior dimensions: at least 720mm x 1980mm x 720mm (W x H x D)	Exterior dimensions: 720mm to 1000mm x 1980mm x 720mm to 1000mm (W x H x D)
<b>F. FIT OUT PLAN</b>		
1	Contractor must submit schematic equipment fit out plan to hospital, to verify the space allotted for the equipment.	<b>DELETE</b>

ADDENDUM:	
<b>A. SPECIFICATIONS of Main Equipment</b>	
16	The equipment must be with a reverse osmosis.
17	The washer and reverse osmosis must fit in the space of 2m x 2.5m.
18	The equipment shall have filtration system when discharging waste water/or air during drying process.

EMMA A. MARIANO, CPA, MGM-ESP  
Chairperson, Bids and Awards Committee

