



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100

website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph

Trunkline: 8588-9900 DirectLine: 8924-0836 Fax No: 8924-0840

BID BULLETIN NO. BB-2022-046

Date : **July 25, 2022**

TO : **ALL PROSPECTIVE BIDDERS**

FROM : **BIDS AND AWARDS COMMITTEE**

SUBJECT : **AMENDMENT AS STATED**

Name of Project : **Supply, Delivery and Installation of One (1) unit WASHER DIS-INFECTOR**

under Invitation to Bid No. IB-2022-109

Total ABC = Php5,000,000.00

This Bid Bulletin No. **BB-2022-046** outlines the amendments and addendum in the bidding documents for the above project as follows:

Under Section VII. Technical Specifications

AMENDMENTS:		
	FROM	TO
A. SPECIFICATIONS of Main Equipment		
1	Chamber Dimensions: -at least 635mm x 645mm x 640mm (W x H x D) -chamber material is made up of stainless steel medical grade 316L.	Chamber Dimensions: -at least 635mm x 645mm x 640mm (W x H x D) -chamber material is made up of stainless steel medical grade 304L or 316L .
3	Door configuration: -single manual drop down glass -double pane tempered glass at least 6mm interior and 3mm exterior glass. (door manual)	Door configuration: single manual drop down glass / automated vertical sliding door ; double pane tempered glass at least 6mm interior and 3mm exterior glass.
4	Loading Height: 755mm above the floor	Loading Height: 755mm above the floor or manufacturer's standard .
7	At least 10 racks capacity. All racks are designed for loading and unloading on all sides.	At least 8-10 DIN trays . All trays are designed for loading and unloading on all sides.
9	Cycle time is at least 35-40 minutes including drying time	Cycle time is at least 30-40 minutes including drying time
B. ACCESSORIES		
3	1 unit 20-minute biological incubator.	DELETE
4	1 unit instrument wash cart	1 unit instrument wash tray / bowl / or manufacturers machine compatible (pre-washing)
C. CONSUMABLES		
1	3 gallons enzymatic cleanser	3 gallons enzymatic cleanser or manufacturer's standard

AMENDMENTS:		
FROM		TO
3	2 container lubricant spray (at least 400mL each)	2 container lubricant spray or manufacturer's standard
5	3 gallons drying agent	DELETE

EMMA A. MARIANO, CPA, MGM-ESP
 Chairperson, Bids and Awards Committee