

# Republic of the Philippines DEPARTMENT OF HEALTH PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: <a href="mailto:www.pcmc.gov.ph">www.pcmc.gov.ph</a> email: <a href="mailto:officeofthedirector@pcmc.gov.ph">officeofthedirector@pcmc.gov.ph</a>
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

### BID BULLETIN NO. BB-2022-048

Date

: August 9, 2022

TO

: ALL PROSPECTIVE BIDDERS

**FROM** 

: BIDS AND AWARDS COMMITTEE

**SUBJECT** 

: AMENDMENT AS STATED

Name of Project/s:

1. Supply and Delivery of One (1) Unit Digital Video Colposcope with Video Digital USB Camera: IB 2022-115 (ABC = Php 2,000,000.00

This Bid Bulletin No. <u>BB-2022-048</u> outlines the amendments in the issued bidding documents for above project as follows:

#### 1. Under Section VII. Technical Specifications

FROM	ТО	
Technical Specifications posted on July 27, 2022	Revised Technical Specifications dated August 9, 2022 (see attached)	

For information and guidance of all concerned.

EMMA A. MARIANO, CPA, MGM-ESP

Chairperson, Bids and Awards Committee

#### PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City

## TECHNICAL SPECIFICATIONS

(as per Bid Bulletin No.: BB-2022-048)

<u>Instruction:</u> Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT			PCMC REQUIREMENT	BIDDERS' OFFER TO COMPLY WITH
QTY	UNIT	Item Description		THE PCMC REQUIREMENTS
1	unit	DIGITAL VIDEO COLPOSCOPE with VIDEO DIGITAL USB CAMERA		
		<u>A. S</u>	pecifications	
		1	Optic colposcope system with digital video and imaging capability	
		2	Equipped with Digital Single Lens Reflex (DSLR) camera to ensure highest image / Video quality documentation	
		3	Zoom knobs conveniently located on both sides of the optical head with vertical and horizontal adjustment capabilities	
		4	Convergent 3 step magnification optical lens design (x7.5/x15/x30) to lessen eye fatigue	
		5	With continuously adjustable white light LED Lighting with green filter lighting that enables detection of vascular patterns	
		6	Camera captures images/ videos can be saved into a preinstalled patient record software	
		7	Objective lens (working distance): 300 mm	
		8	Lamp Power Consumption : 10W	
		9	Illumination: White and Green LED's	
		10	Working Height Adjustment: 780mm - 1200mm	
		11	Intensity of Illumination: 45,000 - 52,000 Lux	
		12	Color Temperature : 5,700 - 6,000 k	
		13	Magnification: x7.5/x15/x30	
		14	Field view: 46/23/11.5	
		15	Eyepieces: straight insight	

PCMC REQUIREMENT			PCMC REQUIREMENT	BIDDERS' OFFER TO COMPLY WITH
QTY	UNIT	Item Description		THE PCMC REQUIREMENTS
1	unit	DIGITAL VIDEO COLPOSCOPE with VIDEO DIGITAL USB CAMERA		
		16	Illuminated field diamter: 58mm	
		17	Horizontal fine adjustment : 40mm	
		18	Vertical fine adjustment : 80mm	
		19	Camera : 24 MP DSLR	
		20	Assembly: on roller base	
		21	Footprint: max 750mm	
		22	Scope Stand/Base Specifications:	
			Height: 50-60 inches (varies by base)	
			Base Diameter: 28 inches	
			Weight: 60 lb (without base: total 100-129 lb varies by base)	
			• 110 V power supply	
		23	Colposcope head on upright wheeled stand	
		24	Pre-installed Patient Record Software	
		25	The unit will be operated at 230 Volts 3-Pin Power Plug / cable, strictly at 60 hertz	
		26	No separate volt conversion	
		27	Power consumption of the unit (must be clearly stipulated)	
		28	Mechanical parts should be of heavy-duty type	
		29	Equipment should be maintainable	
		Othe	r Specifications: (please indicate if applicable)	
		BRAND:		
		MAK	KE / MODEL:	
		B. A.	ccessories (to be provided per unit)	
		1	One (1) Unit UPS with frequency converter or its equivalent	

PCMC REQUIREMENT			PCMC REQUIREMENT	BIDDERS' OFFER TO COMPLY WITH
QTY	UNIT		Item Description	THE PCMC REQUIREMENTS
1	unit	DIGITAL VIDEO COLPOSCOPE with VIDEO DIGITAL USB CAMERA		
			Note: Safety electrical devices should be compatible to the equipment power requirement and proper grounding when needed	
		2	DSLR Camera with Photo Tube	
		3	Footswitch	
		4	Touchscreen, Full ID Tablet PC or Full ID laptop	
		D.W	Varranty	
			Warranty of at least 3 years on parts and 3 years on services and labor	
			Semi Annual preventive maintenance and calibration with certificate and sticker.	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

NAME OF COMPANY	ADDRESS	
SIGNATURE OVER PRINTED NAME	TELEPHONE / FAX NO.	