



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

**BID BULLETIN NO. BB-2022-054**

Date : **October 10, 2022**  
TO : **ALL PROSPECTIVE BIDDERS**  
FROM : **BIDS AND AWARDS COMMITTEE**  
SUBJECT : **AMENDMENT AS STATED**  
Name of Project : **Supply and Delivery of Rapid Antigen Test kit (Covid 19) under Invitation to Bid No. IB-2022-127**  
**Total ABC = Php 8,601,600.00**

This Bid Bulletin No. **BB-2022-054** outlines the amendments and additional requirements in the bidding documents for above projects as follows:

**I. Under Section III. Bid Data Sheet**

FROM	TO
<b>ITB Clause 5.3</b>	<b>ITB Clause 5.3</b>
For this purpose, contracts similar to the Project shall be: a. Supply and delivery of <b>Various Rapid Antigen Test Supplies.</b>	For this purpose, contracts similar to the Project shall be: a. Supply and delivery of Various Laboratory Supplies
<b>ITB Clause 20.2</b>	<b>ITB Clause 20.2</b>
12. With manufacturer and/or products certification by an independent 3rd party certifying body (ISO 14020, 14021, 14024, 14025 or its equivalent), is preferred.	12. With manufacturer and/or products certification by an independent 3rd party certifying body (ISO 14020, 14021, 14024, 14025 or its equivalent), is preferred.
	<b>13. FDA and RITM certified of Rapid Antigen Test Kit (attach proof)</b>
13. Consumer guidelines regarding disposal of the supplies (Information about how and where the used/decommissioned products/ packaging/parts can be returned for recycling and/or disposal e.g. buy-back program)	14. Consumer guidelines regarding disposal of the supplies (Information about how and where the used/decommissioned products/ packaging/parts can be returned for recycling and/or disposal e.g. buy-back program)
14. Other appropriate licenses and permits required by law and stated in the Bidding Documents	15. Other appropriate licenses and permits required by law and stated in the Bidding Documents

**II. Under Section VII. Technical Specifications**

FROM			TO		
<b>PCMC Requirement:</b>			<b>PCMC Requirement:</b>		
<b>QTY</b>	<b>UNIT</b>	<b>ITEM DESCRIPTION</b>	<b>QTY</b>	<b>UNIT</b>	<b>ITEM DESCRIPTION</b>
26,880	pc	Supply and Delivery of Rapid Antigen Test Kit (Covid 19)	26,880	pc	Supply and Delivery of Rapid Antigen Test Kit (Covid 19)
			<b>Specifications:</b>		
			1. Product complied with the required Sensitivity of $\geq 80\%$ and Specificity of $\geq 97\%$		
			2. Results within 15 minutes		
			3. Antigen test with nasopharyngeal swab collection kit.		
			4. Storage of kits at room temperature.		
<b>Additional Requirements:</b>			<b>Additional Requirements:</b>		
Products to be bided shall pass the end-user's evaluation			Products to be bided <b>should have passed</b> the end-user's evaluation		

For information and guidance of all concerned.

**EMMA A. MARIANO, CPA, MGM-ESP**  
Chairperson, Bids and Awards Committee

