



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

## BID BULLETIN NO. BB-2022-055

Date : **October 10, 2022**

TO : **ALL PROSPECTIVE BIDDERS**

FROM : **BIDS AND AWARDS COMMITTEE**

SUBJECT : **AMENDMENT AS STATED**

Name of Project : **Supply and Delivery of two (2) units Portable X-ray Machine DR high-end under Invitation to Bid No. IB-2022-126**  
**Total ABC = Php 25,160,000.00**

This Bid Bulletin No. **BB-2022-055** outlines the amendments in the bidding documents for above projects as follows:

### Under Section VII. Technical Specifications


FROM	TO
<b>I. Specifications</b>	<b>I. Specifications</b>
<b>A. Main Equipment</b>	<b>A. Main Equipment</b>
3. At least 17" to 19" integrated color touch screen monitor	3. At least 17" integrated color touch screen monitor
4. At least a 4.5-meter coiled exposure cable switch	4. At least a 4 meter coiled exposure cable switch
8. Maximum SID not less than 210cm	8. SID not less than 200cm
9. Weight Should not exceed 550 kg.	9. Maximum weight is 560 kg
<b>C. Xray Tube</b>	<b>C. Xray Tube</b>
3. Anode heat capacity of not less than 1,000 HU	3. Anode heat capacity of not less than 1,000 kHU
4. Tube house heat capacity of at least 1,000 HU	4. Tube house heat capacity of at least 1,000 kHU
<b>E. Two (2) Flat Panel Detectors (14 x 17) Magnesium-alloy shell-type frame</b>	<b>E. Two (2) Flat Panel Detectors (14 x 17) Magnesium-alloy shell-type frame</b>
2. Detertor area at least 16.8" x 13.8" pixel	2. Detector area at least 2836 x 2336 pixel
11. Load capacity: at least 310 kg	11. Load capacity: at least 200kg

**BID BULLETIN NO. BB-2022-055**

**Supply and Delivery of two (2) units Portable X-ray Machine DR high-end  
under Invitation to Bid No. IB-2022-126**

<b>FROM</b>	<b>TO</b>
<b>I. Specifications</b>	<b>I. Specifications</b>
<b>F. Two (2) Flat Panel Detectors (10 x 12) Magnesium-alloy shell-type frame</b>	<b>F. Two (2) Flat Panel Detectors (10 x 12) Magnesium-alloy shell-type frame</b>
2. Detertor area at least 16.8" x 13.8" pixel	2. Detector area at least 1920 x 1536 pixel
11. Load capacity: at least 310 kg	11. Load capacity: at least 200kg
<b>II Accessories PER UNIT:</b>	<b>II Accessories PER UNIT:</b>
3. One (1) unit desktop computer with 650VA UPS with license software for work station, 5year warranty, at least 8th generation or higher Intel i7 - 16GB RAM, Keyboard and Mouse, Hard Drive: at least 1TB HDD, at least 23" monitor. The unit and accesories should branded and of the same brand.	3. Branded Laptop, 15.6 in FHD IPS, 15-10500H, 8GB DDR4 RAM, 512 GB SSD, RTX 3050 4 GB GPU, at least windows 10, with bag

For information and guidance of all concerned.



**EMMA A. MARIANO, CPA, MGM-ESP**  
Chairperson, Bids and Awards Committee