



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

**BID BULLETIN NO. BB-2022-057**

Date : **October 18, 2022**  
TO : **ALL PROSPECTIVE BIDDERS**  
FROM : **BIDS AND AWARDS COMMITTEE**  
SUBJECT : **AMENDMENT AS STATED**

Name of Project : **SUPPLY, DELIVERY, INSTALLATION, TESTING, AND COMMISSIONING:**  
**1 unit Single Mobile Operating Room Light under IB-2022-134**  
**ABC = 1,300,000.00**  
**2 units Dual Ceiling Mounted Operating Room Light and Pendant Surgical Monitor under IB-2022-135**  
**ABC = 4,000,000.00 (2,000,000.00/unit)**

This Bid Bulletin No. **BB-2022-057** outlines the amendments on the following which shall form an integral part of the Bidding Documents for above projects as follows:

**IB-2022-134 : One (1) unit Single Mobile Operating Room Light**  
Under Section VII. Technical Specifications

FROM	TO
ITEM DESCRIPTION	ITEM DESCRIPTION
<b>A. SPECIFICATIONS:</b>	<b>A. SPECIFICATIONS:</b>
With two lamps, with at least one lamp capable of at least 160,000 lux	With one lamp capable of at least 160,000 lux
Real-time wireless transmission capability for camera system	Delete
Light head rotates at least 300 degrees	Light head rotates at least 180 degrees
<b>B. ACCESSORIES PER UNIT</b>	<b>B. ACCESSORIES PER UNIT</b>
1 Uninterruptable Power Supply (UPS) of at least 1KVA (compatible to the equipment power requirement) and proper grounding when needed	Uninterruptable Power Supply (UPS) that can regulate voltage to 230V, 60Hz, power supply
2 Automatic Voltage Regulator (AVR) (compatible to the equipment power requirement) and proper grounding when needed	


**IB-2022-135 : Two (2) units Dual Ceiling Mounted Operating Room Light and Pendant Surgical Monitor**  
Under Section VII. Technical Specifications

FROM	TO
ITEM DESCRIPTION	ITEM DESCRIPTION
<b>A. SPECIFICATIONS:</b>	<b>A. SPECIFICATIONS:</b>
With remote and handle/dome controls for illuminance and focus	With remote and handle/dome controls for illuminance and camera system
With at least 80 bulbs or higher	With at least 70 bulbs or higher

### BID BULLETIN NO. BB-2022-057

B. ACCESSORIES PER UNIT		B. ACCESSORIES PER UNIT
1	Uninterruptable Power Supply (UPS) of at least 1KVA (compatible to the equipment power requirement) and proper grounding when needed	Uninterruptable Power Supply (UPS) that can regulate voltage to 230V, 60Hz, power supply
2	Automatic Voltage Regulator (AVR) (compatible to the equipment power requirement) and proper grounding when needed	

For information and guidance of all concerned.



**EMMA A. MARIANO, CPA, MGM-ESP**  
Chairperson, Bids and Awards Committee