



Republic of the Philippines  
DEPARTMENT OF HEALTH  
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

**BID BULLETIN NO. BB-2022-059**

Date : **October 19, 2022**  
TO : **ALL PROSPECTIVE BIDDERS**  
FROM : **BIDS AND AWARDS COMMITTEE**  
SUBJECT : **AMENDMENT AS STATED**  
Name of Project : **Supply and Delivery of One (1) unit Biological Laboratory Reagent Refrigerator under Invitation to Bid No. IB-2022-130**  
**Total ABC = Php 463,903.25**

This Bid Bulletin No. **BB-2022-059** outlines the amendments and additional requirements in the bidding documents for above projects as follows:

**I. Under Section VI. Schedule of Requirements**

FROM	TO
PCMC Requirement Delivery Period	PCMC Requirement Delivery Period
Thirty to <del>Forty-Five</del> (30 - <del>45</del> ) calendar days	Thirty to <u>Sixty</u> (30 - <u>60</u> ) calendar days

**II. Under Section VII. Technical Specifications**

**A. Amendments**

FROM	TO
I. Specifications	I. Specifications
2. Temperature Range: <del>2-8</del> °C	2. Temperature Range: <u>2-15</u> °C
12. External dimension H: <del>158-161 cm</del>	12. External dimension H: <u>up to 164 cm</u>
13. Internal dimension H: <del>129-132 cm</del> W: <del>52-55 cm</del>	13. Internal dimension H: <u>107-132 cm</u> W: <u>50-55 cm</u>
14. Weight: <del>100-128 kgs</del>	14. Weight: <u>96-128 kgs</u>

**B. Additional Specification**

23. Electrical Digital Key Lock

For information and guidance of all concerned.

EMMA A. MARIANO, CPA, MGM-ESP  
Chairperson, Bids and Awards Committee