

## Republic of the Philippines DEPARTMENT OF HEALTH HILIPPINE CHILDREN'S MEDICAL CENTE

PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: <a href="mailto:www.pcmc.gov.ph">www.pcmc.gov.ph</a> email: <a href="mailto:officeofthedirector@pcmc.gov.ph">officeofthedirector@pcmc.gov.ph</a>
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

BID BULLETIN NO. BB-2022-063

DATE

: November 2, 2022

TO

: ALL PROSPECTIVE BIDDERS

FROM

: BIDS AND AWARDS COMMITTEE

**SUBJECT** 

AMENDMENT AS STATED

Name of Project

: Supply and Delivery of:

IB-2022-142: One (1) Unit Videonystanography System IB-2022-143: One (1) Unit Wireless Handheld Ultrasound

This Bid Bulletin No. BB-2022-063 outlines the amendments in the bidding documents for above project as follows:

| FROM  |   | ТО  |
|---|---|---|
| IB-2022-142<br>Under: Section I. In   | evitation to Bid. Section II. Instruction t   | o Bidders, Section III. Bid Data Sheet and                |
| Section VI. Schedu  |   | o Brautis, decasi III. Bia baia sileet ala                |
| Name of Project: Supply and Deliver of One (1) Unit                                 |   | Name of Project: Supply and Deliver of One (1) Unit       |
| Videonystanography System   |   | Videonystagmography System                                |
| IB-2022-143   |   |   |
|   |   | o Bidders, Section III. Bid Data Sheet and                |
| Section VI. Schedule of Requirements  |   |   |
| Name of Project: Supply and Deliver of One (1) Unit<br>Wireless Handheld Ultrasound |   | Name of Project: Supply and Deliver of One (1) Lot        |
|   |   | Wireless Handheld Ultrasound                              |
| Section VII. Technical Specifications Scanners                                      |   | 3. Probes   |
| Multipurpose  | HD Convex: Frequency 2-6 MHz                  | Probe 1. HD Convex: Frequency 2-6 MHz Depth up to         |
| Multiput pose   | Depth up to 40 cm                             | 40 cm   |
|   | Specialty: Abdominal, Lung,                   | Specialty: Abdominal, Lung, OB/GYN, Cardiac,              |
|   | OB/GYN, Cardiac, Superficial,                 | Superficial, Vascular                                     |
|   | Vascular                                      |   |
| Specialty   | HD Linear: Frequency 4-13 MHz                 | Probe 2. HD Linear: Frequency 4-13 MHz Depth up to        |
|   | Depth up to 11 cm                             | 11 cm   |
|   | Specialty: Vascular, Lung, Musculo            | Specialty: Vascular, Lung, Musculo Skeletal, Small Parts, |
|   | Skeletal, Small Parts, Nerve                  | Nerve   |
|   |   |   |
| Probe Transducers Standard Features   |   | Probe Standard Features                                   |
| Secured   |   | has accessible digital storage through iCloud ready a     |
| Storage   | iCloud  | secure web-based portal that allows you to save, review,  |
|   | -   | manage, and share your scanned images                     |
|   | you to save,                                  |   |
|   | review, manage, and share your scanned images |   |
|   | Scamed images                                 |   |

For information and guidance of all concerned.

EMMA A. MARIANO, CPA, MGM-ESP Chairperson, Bids and Awards Committee

