



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

**BID BULLETIN NO. BB-2021-051**

Date : **July 14, 2021**

TO : **ALL PROSPECTIVE BIDDERS**

FROM : **BIDS AND AWARDS COMMITTEE**

SUBJECT : **AMENDMENT AS STATED**

Name of Project : **IB-2021-098: 7 units Cardiac Monitor with Invasive Cables**


This Bid Bulletin No. **BB-2021-051** outlines the amendments in the bidding documents for above project as follows:

**Section VII. Technical Specifications**

FROM	TO
<b>2. MEASUREMENTS</b>	<b>2. MEASUREMENTS</b>
<b>Invasive Blood Pressure with waveform</b>	<b>Invasive Blood Pressure with waveform</b>
Invasive blood pressure, dual invasive blood pressure with dual adapter cable, Pressure monitoring: ABP, ART, CVP, PAP, PPV, Cardiac output monitoring	Invasive blood pressure, dual invasive blood pressure with dual adapter cable, Pressure monitoring: ABP, ART, CVP, PAP, PPV Two (2) cardiac output module inclusive of thermodilution accessories interchangeable across all patient monitors
<b>Pulse Oximetry/SPO2 with waveform</b>	<b>Pulse Oximetry/SPO2 with waveform</b>
SPO2 Monitoring, multisite Y-probe or wrap around rubber glove, shockproof, waterproof SPO2 sensor, Pleth waveform - visual indicator of patient's pulse, Perfusion indicator - numerical value for the pulsatile portion	SPO2 Monitoring, multisite Y-probe or wrap around rubber glove, shockproof, waterproof SPO2 sensor, Pleth waveform - visual indicator of patient's pulse, <i>Perfusion indicator - numerical value or signal for pulsatile portion</i>
<b>3. DATA STORAGE</b>	<b>3. DATA STORAGE</b>
Storage data retrieval at least for 7 days after discharge	Storage data retrieval at least for 5 days after discharge
<b>4. PATIENT MONITOR</b>	<b>4. PATIENT MONITOR</b>
Screen color display resolution minimum 1280 x750 pixel	Screen color display resolution minimum 1024 x 768 pixel

FROM	TO
<b>6. CENTRAL MONITORING SYSTEM</b>	<b>6. CENTRAL MONITORING SYSTEM</b>
<b>CPU</b>	<b>CPU</b>
Processor: Intel i7 6 core 2.9 GHz 10th gen	at least 1.6 GHz Dual-core processor or manufacturer's standard
<b>Operating System</b>	<b>Operating System</b>
Windows 10	Windows Embedded Standard 7 (WSE7)
<b>Main Memory</b>	<b>Main Memory</b>
at least 16 GB RAM	at least 4 GB DDR3 SDRAM
<b>Video card</b>	<b>Video card</b>
Video card: Geforce GTX 1050	Display port 1.1A Output Resolution; 1920x1080, 60Hz or manufacturer's standard
<b>Storage</b>	<b>Storage</b>
1 TB SSD	256 GB SSD or manufacturer's standard
<b>Connection</b>	<b>Connection</b>
HDMI and USB connectivity	HDMI and VGA display port connectivity or manufacturer's standard
<b>Central monitor</b>	<b>Central monitor</b>
With graphical trend data of 10 days	With graphical trend data of at least 6 days
	CPU Operating system compatible with patient monitor software, upgradeable operating system is at no cost once OS platform is not applicable to run the software during the warranty period
<b>7. ACCESSORIES</b>	<b>7. ACCESSORIES</b>
with 2 units removable transport monitor at least 7" screen display attached to cardiac monitor	with 2 units removable transport monitor at least 6.5" screen display attached to cardiac monitor
with 2 units removable printer modules	with 2 units built in thermal printer
<b>10. POWER REQUIREMENT</b>	<b>10. POWER REQUIREMENT</b>
230 VAC/ 60 Hertz	240 VAC/ 60 Hertz

For information and guidance of all concerned.

  
**MARIA ROSARIO S. CRUZ, MD**  
 Chairperson, BAC 