



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Bids and Awards Committee
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: bac@pcmc.gov.ph
Trunkline: 8588-9900 local 361/355 Telefax No.: 8924-0870

SECTION I

Invitation to Bid

Supply and Delivery of Various Medical Equipment

IB 2022-142-IB 2022-154



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

INVITATION TO BID

1. The Philippine Children's Medical Center (PCMC) through the COB CY 2021 intends to apply the sum of Thirty Three Million Seven Hundred Seventy Thousand Pesos (Php 33,770,000.00) being the Approved Budget for the Contract (ABC) to payments under the following Invitation to Bid. Bids received in excess of the ABC shall be automatically rejected at bid opening.

IB NUMBER	Qty	Unit	ITEM	ABC	TOTAL ABC	Cost of Bidding Documents
					(Php)	
IB 2022-142	1	unit	Videostanography System	3,080,000.00	3,080,000.00	5,000.00
IB 2022-143	1	unit	Wireless Handheld Ultrasound	1,000,000.00	1,000,000.00	1,000.00
IB 2022-144	2	unit	Bed Bath Stretcher	150,000.00	300,000.00	500.00
IB 2022-145	1	unit	Pediatric Orthopedic Traction bed Package with bedside table overbed table and footstool with IV pole	350,000.00	350,000.00	500.00
IB 2022-146	11	unit	Stainless Medicine Cabinet Heavy Duty	300,000.00	3,300,000.00	5,000.00
IB 2022-147	4	unit	Phototherapy, Blue Light	160,000.00	640,000.00	1,000.00
IB 2022-148	2	unit	Weighing Scale for Neonates	80,000.00	160,000.00	500.00
IB 2022-149	10	unit	Portable Digital Sphygmomanometer	110,000.00	1,100,000.00	5,000.00
IB 2022-150	14	unit	Refrigerator for Medicine (+2-+8c)	60,000.00	840,000.00	1,000.00
IB 2022-151	3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	1,500,000.00	4,500,000.00	5,000.00
IB 2022-152	1	unit	High-End Cardiac Ultrasound with Advanced Applications & Interchangeable Probes	12,000,000.00	12,000,000.00	25,000.00
IB 2022-153	1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	6,000,000.00	6,000,000.00	10,000.00
IB 2022-154	1	unit	Portable ECG Machine 12 Led with review station	500,000.00	500,000.00	500.00

2. The Philippine Children's Medical Center (PCMC) now invites bids for the above-mentioned project. Delivery of the Goods is required as stated in Section VI Delivery of Requirements. Bidders should have completed, within the past three (3) years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II, Instructions to Bidders.

PhilHealth Accredited

IB-2022-142 to IB-2022-154



3. Bidding will be conducted through open competitive bidding procedures using a nondiscretionary "pass/fail" criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) 9184, otherwise known as the "Government Procurement Reform Act".

- a. Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country
- b. the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA No. 5183.

4. Prospective Bidders may obtain further information from PCMC and inspect the Bidding Documents at the address given below during office hours.

5. A complete set of Bidding Documents may be acquired by interested Bidders starting **October 20, 2022**, upon payment of the applicable fee stated above. It may also be downloaded free of charge from the website of the Philippine Government Electronic Procurement System (PhilGEPS) and the website of PCMC, provided that Bidders shall pay the applicable fee for the Bidding Documents not later than the submission of their bids.

6. The Philippine Children's Medical Center will hold a **Pre-Bid Conference on October 28, 2022 at 10:00 A.M** through video conferencing via *zoom* (Meeting ID: 972 0276 1051 Passcode: 142-154) which shall be open to prospective bidders.

7. Bids must be duly received through manual submission on or before **November 9, 2022, 9:30 A.M., Guard-on-Duty, 3rd Floor, Procurement Division Area, PCMC Main Building**. Late bids shall not be accepted.

8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in **ITB Clause 14**.

9. Bid opening shall be on **November 9, 2022, 10:00 A.M** 3rd Floor, Function Hall 1 & 2, PCMC Main Building. Bids will be opened in the presence of the Bidders' representatives who choose to attend at the aforementioned venue. In compliance to social distancing and to support the government's effort to mitigate, if not contain transmission of COVID -19, we will **strictly allow only one authorized representative** per bidder company to enter the venue during opening of bids. Provided further, that said authorized representative shall follow PCMC's safety protocol by wearing face mask as required prior entering PCMC Premises.

10. The **Philippine Children's Medical Center (PCMC)** reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA No. 9184, without thereby incurring any liability to the affected bidder or bidders.

11. For further information, please refer to:

Procurement Division
3rd Floor, Procurement Division Area
PCMC Main Building
Quezon Avenue, cor. Agham Road Quezon City
Trunk line: 8588-9900 Loc 361 / 355
Fax Number: 924-0870
Email: pcmcbaac@gmail.com

12. You may visit the following websites:

For downloading of Bidding Document: www.pcmc.gov.ph
www.philgeps.gov.ph

October 20, 2022



EMMA A. MARIANO, CPA, MGM-ESP
Chairman, Bids & Awards Committee



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Bids and Awards Committee
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: bac@pcmc.gov.ph
Trunkline: 8588-9900 local 361/355 Telefax No.: 8924-0870

SECTION II

Instructions to Bidders

*Supply and Delivery of
Various Medical Equipment*

IB-2022-142 to 154

1. Scope of Bid

The **Philippine Children's Medical Center (PCMC)** wishes to receive Bids for the following Project/s:

IB Number	Qty	UOM	ITEM DESCRIPTION	ABC per unit	TOTAL ABC
<i>Supply and Delivery of:</i>					
IB 2022-142	1	unit	Videonystanography System	3,080,000.00	3,080,000.00
IB 2022-143	1	unit	Wireless Handheld Ultrasound	1,000,000.00	1,000,000.00
IB 2022-144	2	unit	Bed Bath Stretcher	150,000.00	300,000.00
IB 2022-145	1	unit	Pediatric Orthopedic Traction bed Package with bedside table overbed table and footstool with IV pole	350,000.00	350,000.00
IB 2022-146	11	unit	Stainless Medicine Cabinet Heavy Duty	300,000.00	3,300,000.00
IB 2022-147	4	unit	Phototherapy, Blue Light	160,000.00	640,000.00
IB 2022-148	2	unit	Weighing Scale for Neonates	80,000.00	160,000.00
IB 2022-149	10	unit	Portable Digital Sphygmomanometer	110,000.00	1,100,000.00
IB 2022-150	14	unit	Refrigerator for Medicine (+2-+8c)	60,000.00	840,000.00
IB 2022-151	3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	1,500,000.00	4,500,000.00
IB 2022-152	1	unit	High-End Cardiac Ultrasound with Advanced Applications & Interchangeable Probes	12,000,000.00	12,000,000.00
IB 2022-153	1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	6,000,000.00	6,000,000.00
IB 2022-154	1	unit	Portable ECG Machine 12 Led with review station	500,000.00	500,000.00

The above Procurement Projects, the details of which are described in Section VII (Technical Specifications).

2. Funding Information

2.1. The GOP through the source of funding as indicated below for **COB CY 2022** in the amount of **Thirty Three Million Seven Hundred Seventy Thousand Peso (Php 33,770,000.00)**.

2.2. The source of funding is:

a. GOCC and GFIs, the Corporate Operating Budget.

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2.
 - a. Foreign ownership exceeding those allowed under the rules may participate pursuant to:
 - i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
 - ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
 - iii. When the Goods sought to be procured are not available from local suppliers; or
 - iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:
 - a. For the procurement of Non-expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
 - b. For procurement where the Procuring Entity has determined, after the conduct of market research, that imposition of either (a) or (b) will likely result to failure of bidding or monopoly that will defeat the purpose of public bidding: the Bidder should comply with the following requirements: [*Select either failure or monopoly of bidding based on market research conducted*]
 - i. Completed at least two (2) similar contracts, the aggregate amount of which should be equivalent to at least *fifty percent (50%) in the case of non-expendable supplies and services or twenty-five percent (25%) in the case of expendable supplies*] of the ABC for this Project; and
 - ii. The largest of these similar contracts must be equivalent to at least half of the percentage of the ABC as required above.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

7. Subcontracts

- 7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that:

- a. Subcontracting is not allowed.

8. Pre-Bid Conference

The Philippine Children's Medical Center will hold a **Pre-Bid Conference** on **October 28, 2022** at **10:00 A.M** through video conferencing via *zoom* (**Meeting ID: 972 0276 1051 Passcode: 142-154**) which shall be open to prospective bidders.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, **at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.**

10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within **the past three (3) years** prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
 - a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, ex-warehouse, ex-showroom, or off-the-shelf, as applicable);
 - ii. The cost of all customs duties and sales and other taxes already paid or payable;
 - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
 - iv. The price of other (incidental) services, if any, listed in e.

- b. For Goods offered from abroad:
 - i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
 - ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications)**.

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in:
 - a. Philippine Pesos.

14. Bid Security

- 14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security shall be valid until **120 calendar days**. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

15. Sealing and Marking of Bids

Use of indelible ink color blue shall be used by the authorized signatory in signing the required forms. **Strictly NO using of staple wire and thick materials for tab**

The **First (1st) Envelope**, shall contain the following **Technical Component** accomplished in two (2) sets, **each set filed in a folder**

The **Second (2nd) Envelope** shall contain the **Financial Component** accomplished in three (3) sets, **each set filed in a folder**

All copies should be certified as true copy

COLOR CODING OF FOLDERS/ENVELOPES	BLUE
--	-------------

LABEL ON THE ENVELOPE/S:

Name of PROCURING ENTITY
 Name of CONTRACT TO BE BID
 IB Number
 DATE of Bid Opening
 Name of the Bidder Company
 Address of the Bidder Company

IDENTIFY THE ENVELOPES:

as: > Technical Component Requirements (original and copy 1)
 > Financial Component Requirement (original, copy 1 and copy 2)

16. Deadline for Submission of Bids

- 16.1. The Bidders shall submit on the specified date and time and either at its physical address or through online submission as indicated in paragraph 7 of the **IB**.

17. Opening and Preliminary Examination of Bids

- 17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

- 17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

- 18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "*passed*," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.

- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.

- 19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.

- 19.4. The Project shall be awarded as follows:

Option 3 - One Project having several items, which shall be awarded as separate contracts per item.

- 19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

- 20.2. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing

and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**.

21. Signing of the Contract

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

CONFORME:

Authorized Signatory (Signature over printed name)

Contact No:

Name of Company/Firm

Company's Official Email Address
(where notices will be sent)

Company's Official
Contact No.



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Bids and Awards Committee
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: pcmcbac@gmail.com
Trunkline: 8588-9900 local 361/355 Telefax No.: 8924-0870

SECTION III

Bid Data Sheet

*Supply and Delivery of
Various Medical Equipment*

IB-2022-142 to 154

Bid Data Sheet

ITB Clause							
5.3	<p>For this purpose, contracts similar to the Project shall be:</p> <p>a. Supply and delivery of <i>various rehab / laboratory / hospital / medical / radiological / dental equipment</i></p> <p>b. Completed within <i>the last three (3) years</i> prior to the deadline for the submission and receipt of bids.</p>						
7.1	Subcontracting is not allowed.						
12	The Bid prices for Goods supplied from outside of the Philippines shall be quoted in Philippine Pesos.						
14.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:						
	IB No.	Qty	Unit	Item Description	Total ABC	Bid Security Amount in cash, cashier's/ manager's check, bank draft/ guarantee or irrevocable letter of credit (2% of ABC)	Bid Security Amount in Surety Bond (5% of ABC)
	IB 2022-142	1	unit	Videonystanography System	3,080,000.00	61,600.00	1,232,000.00
	IB 2022-143	1	unit	Wireless Handheld Ultrasound	1,000,000.00	20,000.00	400,000.00
	IB 2022-144	2	unit	Bed Bath Stretcher	300,000.00	6,000.00	120,000.00
	IB 2022-145	1	unit	Pediatric Orthopedic Traction bed Package with bedside table overbed table and footstool with IV pole	350,000.00	7,000.00	140,000.00
	IB 2022-146	11	unit	Stainless Medicine Cabinet Heavy Duty	3,300,000.00	66,000.00	1,320,000.00
	IB 2022-147	4	unit	Phototherapy, Blue Light	640,000.00	12,800.00	256,000.00
	IB 2022-148	2	unit	Weighing Scale for Neonates	160,000.00	3,200.00	64,000.00
	IB 2022-149	10	unit	Portable Digital Sphygmomanometer	1,100,000.00	22,000.00	440,000.00
	IB 2022-150	14	unit	Refrigerator for Medicine (+2-+8c)	840,000.00	16,800.00	336,000.00
	IB 2022-151	3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	4,500,000.00	90,000.00	1,800,000.00
	IB 2022-152	1	unit	High-End Cardiac Ultrasound with Advanced Applications & Interchangeable Probes	12,000,000.00	240,000.00	4,800,000.00
	IB 2022-153	1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	6,000,000.00	120,000.00	2,400,000.00
	IB 2022-154	1	unit	Portable ECG Machine 12 Led with review station	500,000.00	10,000.00	200,000.00

19.3	<p>Supply and Delivery of the following:</p> <table border="1"> <thead> <tr> <th data-bbox="313 256 456 318">IB Number</th> <th data-bbox="456 256 521 318">Qty</th> <th data-bbox="521 256 602 318">UOM</th> <th data-bbox="602 256 1089 318">ITEM DESCRIPTION</th> <th data-bbox="1089 256 1252 318">ABC per unit</th> <th data-bbox="1252 256 1432 318">TOTAL ABC</th> </tr> </thead> <tbody> <tr> <td colspan="6" data-bbox="313 318 1432 351">Supply and Delivery of:</td> </tr> <tr> <td data-bbox="313 351 456 383">IB 2022-142</td> <td data-bbox="456 351 521 383">1</td> <td data-bbox="521 351 602 383">unit</td> <td data-bbox="602 351 1089 383">Videonystanography System</td> <td data-bbox="1089 351 1252 383">3,080,000.00</td> <td data-bbox="1252 351 1432 383">3,080,000.00</td> </tr> <tr> <td data-bbox="313 383 456 416">IB 2022-143</td> <td data-bbox="456 383 521 416">1</td> <td data-bbox="521 383 602 416">unit</td> <td data-bbox="602 383 1089 416">Wireless Handheld Ultrasound</td> <td data-bbox="1089 383 1252 416">1,000,000.00</td> <td data-bbox="1252 383 1432 416">1,000,000.00</td> </tr> <tr> <td data-bbox="313 416 456 448">IB 2022-144</td> <td data-bbox="456 416 521 448">2</td> <td data-bbox="521 416 602 448">unit</td> <td data-bbox="602 416 1089 448">Bed Bath Stretcher</td> <td data-bbox="1089 416 1252 448">150,000.00</td> <td data-bbox="1252 416 1432 448">300,000.00</td> </tr> <tr> <td data-bbox="313 448 456 518">IB 2022-145</td> <td data-bbox="456 448 521 518">1</td> <td data-bbox="521 448 602 518">unit</td> <td data-bbox="602 448 1089 518">Pediatric Orthopedic Traction bed Package with bedside table overbed table and footstool with IV pole</td> <td data-bbox="1089 448 1252 518">350,000.00</td> <td data-bbox="1252 448 1432 518">350,000.00</td> </tr> <tr> <td data-bbox="313 518 456 550">IB 2022-146</td> <td data-bbox="456 518 521 550">11</td> <td data-bbox="521 518 602 550">unit</td> <td data-bbox="602 518 1089 550">Stainless Medicine Cabinet Heavy Duty</td> <td data-bbox="1089 518 1252 550">300,000.00</td> <td data-bbox="1252 518 1432 550">3,300,000.00</td> </tr> <tr> <td data-bbox="313 550 456 583">IB 2022-147</td> <td data-bbox="456 550 521 583">4</td> <td data-bbox="521 550 602 583">unit</td> <td data-bbox="602 550 1089 583">Phototherapy, Blue Light</td> <td data-bbox="1089 550 1252 583">160,000.00</td> <td data-bbox="1252 550 1432 583">640,000.00</td> </tr> <tr> <td data-bbox="313 583 456 615">IB 2022-148</td> <td data-bbox="456 583 521 615">2</td> <td data-bbox="521 583 602 615">unit</td> <td data-bbox="602 583 1089 615">Weighing Scale for Neonates</td> <td data-bbox="1089 583 1252 615">80,000.00</td> <td data-bbox="1252 583 1432 615">160,000.00</td> </tr> <tr> <td data-bbox="313 615 456 647">IB 2022-149</td> <td data-bbox="456 615 521 647">10</td> <td data-bbox="521 615 602 647">unit</td> <td data-bbox="602 615 1089 647">Portable Digital Sphygmomanometer</td> <td data-bbox="1089 615 1252 647">110,000.00</td> <td data-bbox="1252 615 1432 647">1,100,000.00</td> </tr> <tr> <td data-bbox="313 647 456 680">IB 2022-150</td> <td data-bbox="456 647 521 680">14</td> <td data-bbox="521 647 602 680">unit</td> <td data-bbox="602 647 1089 680">Refrigerator for Medicine (+2-+8c)</td> <td data-bbox="1089 647 1252 680">60,000.00</td> <td data-bbox="1252 647 1432 680">840,000.00</td> </tr> <tr> <td data-bbox="313 680 456 725">IB 2022-151</td> <td data-bbox="456 680 521 725">3</td> <td data-bbox="521 680 602 725">unit</td> <td data-bbox="602 680 1089 725">High-End Cardiac Monitor with Invasive Pressure Monitoring</td> <td data-bbox="1089 680 1252 725">1,500,000.00</td> <td data-bbox="1252 680 1432 725">4,500,000.00</td> </tr> <tr> <td data-bbox="313 725 456 770">IB 2022-152</td> <td data-bbox="456 725 521 770">1</td> <td data-bbox="521 725 602 770">unit</td> <td data-bbox="602 725 1089 770">High-End Cardiac Ultrasound with Advanced Applications & Interchangeable Probes</td> <td data-bbox="1089 725 1252 770">12,000,000.00</td> <td data-bbox="1252 725 1432 770">12,000,000.00</td> </tr> <tr> <td data-bbox="313 770 456 814">IB 2022-153</td> <td data-bbox="456 770 521 814">1</td> <td data-bbox="521 770 602 814">unit</td> <td data-bbox="602 770 1089 814">High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes</td> <td data-bbox="1089 770 1252 814">6,000,000.00</td> <td data-bbox="1252 770 1432 814">6,000,000.00</td> </tr> <tr> <td data-bbox="313 814 456 824">IB 2022-154</td> <td data-bbox="456 814 521 824">1</td> <td data-bbox="521 814 602 824">unit</td> <td data-bbox="602 814 1089 824">Portable ECG Machine 12 Led with review station</td> <td data-bbox="1089 814 1252 824">500,000.00</td> <td data-bbox="1252 814 1432 824">500,000.00</td> </tr> </tbody> </table>	IB Number	Qty	UOM	ITEM DESCRIPTION	ABC per unit	TOTAL ABC	Supply and Delivery of:						IB 2022-142	1	unit	Videonystanography System	3,080,000.00	3,080,000.00	IB 2022-143	1	unit	Wireless Handheld Ultrasound	1,000,000.00	1,000,000.00	IB 2022-144	2	unit	Bed Bath Stretcher	150,000.00	300,000.00	IB 2022-145	1	unit	Pediatric Orthopedic Traction bed Package with bedside table overbed table and footstool with IV pole	350,000.00	350,000.00	IB 2022-146	11	unit	Stainless Medicine Cabinet Heavy Duty	300,000.00	3,300,000.00	IB 2022-147	4	unit	Phototherapy, Blue Light	160,000.00	640,000.00	IB 2022-148	2	unit	Weighing Scale for Neonates	80,000.00	160,000.00	IB 2022-149	10	unit	Portable Digital Sphygmomanometer	110,000.00	1,100,000.00	IB 2022-150	14	unit	Refrigerator for Medicine (+2-+8c)	60,000.00	840,000.00	IB 2022-151	3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	1,500,000.00	4,500,000.00	IB 2022-152	1	unit	High-End Cardiac Ultrasound with Advanced Applications & Interchangeable Probes	12,000,000.00	12,000,000.00	IB 2022-153	1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	6,000,000.00	6,000,000.00	IB 2022-154	1	unit	Portable ECG Machine 12 Led with review station	500,000.00	500,000.00
IB Number	Qty	UOM	ITEM DESCRIPTION	ABC per unit	TOTAL ABC																																																																																						
Supply and Delivery of:																																																																																											
IB 2022-142	1	unit	Videonystanography System	3,080,000.00	3,080,000.00																																																																																						
IB 2022-143	1	unit	Wireless Handheld Ultrasound	1,000,000.00	1,000,000.00																																																																																						
IB 2022-144	2	unit	Bed Bath Stretcher	150,000.00	300,000.00																																																																																						
IB 2022-145	1	unit	Pediatric Orthopedic Traction bed Package with bedside table overbed table and footstool with IV pole	350,000.00	350,000.00																																																																																						
IB 2022-146	11	unit	Stainless Medicine Cabinet Heavy Duty	300,000.00	3,300,000.00																																																																																						
IB 2022-147	4	unit	Phototherapy, Blue Light	160,000.00	640,000.00																																																																																						
IB 2022-148	2	unit	Weighing Scale for Neonates	80,000.00	160,000.00																																																																																						
IB 2022-149	10	unit	Portable Digital Sphygmomanometer	110,000.00	1,100,000.00																																																																																						
IB 2022-150	14	unit	Refrigerator for Medicine (+2-+8c)	60,000.00	840,000.00																																																																																						
IB 2022-151	3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	1,500,000.00	4,500,000.00																																																																																						
IB 2022-152	1	unit	High-End Cardiac Ultrasound with Advanced Applications & Interchangeable Probes	12,000,000.00	12,000,000.00																																																																																						
IB 2022-153	1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	6,000,000.00	6,000,000.00																																																																																						
IB 2022-154	1	unit	Portable ECG Machine 12 Led with review station	500,000.00	500,000.00																																																																																						
20.2	<p>The Lowest Calculated Bidder shall submit the following documentary requirements within a non-extendible period of five (5) calendar days from receipt of the notification that contain the following:</p> <ol style="list-style-type: none"> 1. Latest Income and Business Tax Returns filed and paid through the BIR Electronic Filing (EFPS) within the last three (3) months 2. Certificate of Performance in letterhead of their clients indicating the contact numbers and email addresses signed by the authorized head of the Department from three (3) clients of the bidder issued within the last six (6) months prior to bid opening. <p style="margin-left: 40px;">Note: Certification issued by PCMC – Materials Management Division must be included if bidder had done business with us. Certification of which should be of same category (e.g. equipment/supplies) of project being bided.</p> 3. Duly notarized Certificate as Principal Manufacturer or Exclusive Distributor/ Authorized Dealer from the Principal Manufacturer 4. Current and Valid Certificate of Manufacturer's compliance with ISO and/or other Standards/Conformance /Safety Requirements for the equipment 5. Duly notarized Certificate from Principal Manufacturer/ Distributor that the brand has been present for at least five (5) years in the local/ international market and model being bided is currently in use in the local/international market with at least one (1) installation in the Philippines, preferably within NCR area (Please specify and submit supporting documents) 6. Certificate from the Principal Manufacturer duly notarized from country of origin: <ol style="list-style-type: none"> i. All the terms and conditions stated in the bidding documents per IRR of RA 9184 and corresponding contract for the project shall be honored by the Principal Manufacturer, including in the event that a change of dealership will occur during the duration of the contract up to the warranty and preventive maintenance period; ii. The expected useful life of the equipment under normal use (indicate normal capacity, i.e. number of patients, operating hours, other considerations); 																																																																																										

	<ul style="list-style-type: none"> iii. Guarantee on availability of all spare parts, accessories and consumables at least for the next five (5) years from testing, commissioning, acceptance and delivery; iv. That it has competence in handling and providing technical support as well as maintenance of the equipment being offered; and v. Consumer guidelines regarding disposal of the equipment (Information about how and where the used and decommissioned products/ parts can be returned for recycling and/ or disposal (e.g. buy-back program of the product after end of useful life). <p>7. Duly notarized Certificate from Bidder:</p> <ul style="list-style-type: none"> i. That parts, accessories and consumables are readily available at the authorized Philippine service center/s for a period of five (5) years after the warranty period; ii. That it has available competent in-house technical specialists in handling and providing technical support as well as maintenance of the equipment being offered; iii. That it will conduct training for proper operation and maintenance to end-users of the equipment upon delivery; and iv. That it will provide replacement/back-up unit while the delivered unit is being repaired. <p>8. The Recurring and Maintenance Costs (<i>use of Form DOBA – PCMC – RMF8 is required</i>)</p> <p>9. List of Consumables (PARTS/ACCESSORIES/SUPPLIES) [<i>use of Form DOBA – PCMC – LCF9 is required</i>]</p> <p>10. Section II. Instructions to Bidders with signature (conforme) on all pages</p> <p>11. Section III. Bid Data Sheet with signature (conforme) on all pages</p> <p>12. Section IV. General Conditions of the Contract with signature (conforme) on all pages</p> <p>13. Section V. Special Conditions of the Contract with signature (conforme) on all pages</p> <p>Failure of the Bidder declared as LCB to duly submit the requirements stated above or a finding against the veracity of such shall be ground for forfeiture of the bid security and disqualify the Bidder for award.</p>
21.2	<i>No additional contract documents relevant to the Project</i>

CONFORME:

Authorized Signatory (Signature over printed name)

Contact No:

Name of Company/Firm

Company’s Official Email Address
(where notices will be sent)

Company’s Official
Contact No.



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Bids and Awards Committee
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: pcmcbac@gmail.com
Trunkline: 8588-9900 local 361/355 Telefax No.: 8924-0870

SECTION IV

General Conditions of Contract

*Supply and Delivery of
Various Medical Equipment*

IB-2022-142 to 154

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

2. Advance Payment and Terms of Payment

2.1. Advance payment of the contract amount is provided under Annex “D” of the revised 2016 IRR of RA No. 9184.

2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

[Include the following clauses if Framework Agreement will be used:]

2.3. For a single-year Framework Agreement, prices charged by the Supplier for Goods delivered and/or services performed under a Call-Off shall not vary from the prices quoted by the Supplier in its bid.

2.4. For multi-year Framework Agreement, prices charged by the Supplier for Goods delivered and/or services performed under a Call-Off shall not vary from the prices quoted by the Supplier during conduct of Mini-Competition.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184. *[[Include if Framework Agreement will be used:] In the case of Framework Agreement, the Bidder may opt to furnish the performance security or a Performance Securing Declaration as defined under the Guidelines on the Use of Framework Agreement.]*

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project *[[Include if Framework Agreement will be used:] or Framework Agreement]* specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the **SCC, Section IV (Technical Specifications)** shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

- 6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 Revised IRR of RA No. 9184.
- 6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier’s liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

CONFORME:

Authorized Signatory (Signature over printed name)

Contact No:

Name of Company/Firm

Company’s Official Email Address
(where notices will be sent)

Company’s Official
Contact No.



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Bids and Awards Committee
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: pcmcbac@gmail.com
Trunkline: 8588-9900 local 361/355 Telefax No.: 8924-0870

SECTION V

Special Conditions of Contract

*Supply and Delivery of
Various Medical Equipment*

IB-2022-142 to 154

Special Conditions of Contract

GCC Clause	
1	<p>Delivery and Documents –</p> <p>For purposes of the Contract, “EXW,” “FOB,” “FCA,” “CIF,” “CIP,” “DDP” and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:</p> <p><i>For Goods Supplied from Abroad</i> “The delivery terms applicable to the Contract are DDP delivered to PCMC. In accordance with INCOTERMS.”</p> <p><i>For Goods Supplied from Within the Philippines,</i> “The delivery terms applicable to this Contract are delivered to PCMC. Risk and title will pass from the Supplier to PCMC upon receipt and final acceptance of the Goods at their final destination.”</p> <p>Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).</p> <p>The details of shipping and/or other documents to be furnished by the Supplier are as follows:</p> <p><i>For Goods supplied from within the Philippines:</i></p> <p>Upon delivery of the Goods to the Project Site, the Supplier shall notify the Procuring Entity and present the following documents to the Procuring Entity:</p> <ul style="list-style-type: none"> (i) Current and Valid Certificate of Manufacturer's compliance with ISO and/or other Standards/Conformance /Safety Requirements for the equipment; (ii) Original copy of the Supplier’s invoice showing Goods’ description, quantity, unit price, and total amount; (iii) Original copy of the Manufacturer’s and/or Supplier’s warranty certificate on parts and services, including accessories starting from final acceptance; (iv) Original copy of Certification of Availability of Replacement/ Back-up unit while the provided unit/s is/are being repaired. (v) Original copy of Certificate of Calibration, Testing and Adjustment (vi) Original copy of certificate of preventive maintenance during the warranty period. (vii) Original copy of certificate of attendance on conducted training to end-users and Engineering Personnel for proper operation and maintenance of the equipment. (viii) Original and two (2) copies in English Language: <ul style="list-style-type: none"> a. User's Manual (Hard copy) b. Operation's Manual (hard copy) c. Service Manual with CD Diagram (ix) Original Copy of Certification as locally manufactured <p><i>For Goods supplied from abroad:</i></p> <p>Upon shipment, the Supplier shall notify the Procuring Entity and the insurance company by cable the full details of the shipment, including Contract Number, description of the Goods, quantity, vessel, bill of lading number and date, port of loading, date of shipment, port of discharge etc. Upon delivery to the Project Site, the Supplier shall notify the Procuring Entity and present the following documents as applicable with the documentary requirements of any letter of credit issued taking precedence:</p>

- (i) Original copy of the Supplier's invoice showing Goods' description, quantity, unit price, and total amount;
- (ii) Original copy of the Manufacturer's and/or Supplier's warranty certificate on parts and services, including accessories starting from final acceptance
- (iii) Original copy of Certification of Availability of Replacement/ Back-up unit while the provided unit/s is/are being repaired.
- (iv) Original copy of Certificate of Calibration, Testing and Adjustment
- (v) Original copy of certificate of preventive maintenance during the warranty period.
- (vi) Original copy of certificate of attendance on conducted training to end-users and Engineering Personnel for proper operation and maintenance of the equipment.
- (vii) Original and two (2) copies of:
 - a. User's Manual in English language,
 - b. Operation's Manual (hard copy)
 - c. Service Manual with CD Diagram
- (viii) Copy of Proof of Payment of Import Duties from Bureau of Customs / Bill of Lading.

For purposes of this Clause the Procuring Entity's Representative at the Project Site is the Property and Supply Section / Procurement Section.

Incidental Services –

The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements:

- a. performance or supervision of on-site assembly and/or start-up of the supplied Goods;
- b. furnishing of tools required for assembly and/or maintenance of the supplied Goods;
- c. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;
- d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and
- e. training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.

The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

Spare Parts –

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

Select appropriate requirements and delete the rest.

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
 - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and

- ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in Section VI (Schedule of Requirements) and the costs thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for at least for the next five (5) years from testing, commissioning, acceptance and delivery;

Spare parts or components shall be supplied as promptly as possible, but in any case for a period of five (5) years after the warranty period;

Packaging –

The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit

The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.

The outer packaging must be clearly marked on at least four (4) sides as follows:

Name of the Procuring Entity
Name of the Supplier
Contract Description
Final Destination
Gross weight
Any special lifting instructions
Any special handling instructions
Any relevant HAZCHEM classifications

A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.

Transportation –

Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.

Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.

Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this

	<p>Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.</p> <p>The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.</p> <p>Intellectual Property Rights –</p> <p>The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.</p>
2.2	<p>The terms of payment shall be as follows : Thirty to Forty-Five (30 – 45) calendar days from submission of documentary requirements</p>
3	<p>Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security.</p>
4	<p>The inspections and tests that will be conducted are:</p> <ol style="list-style-type: none"> 1) Upon delivery, the Goods shall undergo preliminary physical inspection by the Inspection Team of the PROCURING ENTITY to ascertain the physical condition and acceptability of the Goods. 2) The supplier shall promptly replace the equivalent quantity of Goods taken as samples without cost to the PROCURING ENTITY.
5	<p>The obligation of the winning bidder for the warranty shall be covered by retention money required of under R.A. Sec. 62.1., equivalent to five percent (5%) of every progress payment.</p> <p>The said amount shall only be released after the lapse of the warranty period specified in Section VII Technical Specification; provided, however, that the Supplies delivered are free from patent and latent defects and all the conditions imposed under this Contract have been fully met.</p>

CONFORME:

Authorized Signatory (Signature over printed name)

Contact No:

Name of Company/Firm

Company's Official Email Address
(where notices will be sent)

Company's Official
Contact No.



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Bids and Awards Committee
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: pcmbac@gmail.com
Trunkline: 8588-9900 local 361/355 Telefax No.: 8924-0870

SECTION VI

Schedule of Requirements

Supply and Delivery of Various Medical Equipment

IB-2022-142 to 154

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

IB Number	Qty	UOM	ITEM DESCRIPTION	ABC per unit	TOTAL ABC
Supply and Delivery of:					
IB 2022-142	1	unit	Videonystanography System	3,080,000.00	3,080,000.00
IB 2022-143	1	unit	Wireless Handheld Ultrasound	1,000,000.00	1,000,000.00
IB 2022-144	2	unit	Bed Bath Stretcher	150,000.00	300,000.00
IB 2022-145	1	unit	Pediatric Orthopedic Traction bed Package with bedside table overbed table and footstool with IV pole	350,000.00	350,000.00
IB 2022-146	11	unit	Stainless Medicine Cabinet Heavy Duty	300,000.00	3,300,000.00
IB 2022-147	4	unit	Phototherapy, Blue Light	160,000.00	640,000.00
IB 2022-148	2	unit	Weighing Scale for Neonates	80,000.00	160,000.00
IB 2022-149	10	unit	Portable Digital Sphygmomanometer	110,000.00	1,100,000.00
IB 2022-150	14	unit	Refrigerator for Medicine (+2-+8c)	60,000.00	840,000.00
IB 2022-151	3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	1,500,000.00	4,500,000.00
IB 2022-152	1	unit	High-End Cardiac Ultrasound with Advanced Applications & Interchangeable Probes	12,000,000.00	12,000,000.00
IB 2022-153	1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	6,000,000.00	6,000,000.00
IB 2022-154	1	unit	Portable ECG Machine 12 Led with review station	500,000.00	500,000.00

Delivery Site	PCMC Requirement DELIVERY PERIOD	Bidder's Offer <i>(specify definite offer on delivery schedule)</i>
Materials and Management Division G/F PCMC, Quezon Avenue, cor. Agham Road Quezon City	Within Forty -Five to Sixty (46 to 60) calendar days upon receipt of Purchase Order/Notice to Proceed	

DELIVERY AND ACCEPTANCE

- The supplier should deliver the goods called for in the Purchase Order (PO) within the Delivery Period, as offered, upon receipt of approved Purchase Order (PO) through faxed or personally received during office hours at the Procurement Division.
- All goods delivered pursuant to the Purchase Order (PO) shall be subject to acceptance and inspection by the end-user as well as by the House Inspector and of the Resident Auditor or their representatives. Goods delivered not in conformity with specifications shall be rejected and the contractor held in default.

CONFORME:

Authorized Signatory (Signature over printed name)

Contact No:

Name of Company/Firm

Company's Official Email Address
(where notices will be sent)

Company's Official
Contact No.



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Bids and Awards Committee
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: pcmcba@gmail.com
Trunkline: 8588-9900 local 361/355 Telefax No.: 8924-0870

SECTION VII

Technical Specifications

*Supply and Delivery of
Various Medical Equipment*

IB-2022-142 to 154



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER
Bids and Awards Committee
Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: pcmcba@gmail.com
Trunkline: 8588-9900 local 361/355 Telefax No.: 8924-0870

SECTION VIII

Checklist of Technical and Financial Documents

Supply and Delivery of Various Medical Equipment

IB-2022-142 to 154

Checklist of Technical and Financial Documents

The Bidder shall submit the following **TECHNICAL COMPONENT ENVELOPE (ARRANGED, NUMBERED AND TABBED)** *[Strictly NO using of staple wire and thick materials for tabs]* as enumerated below:

I. TECHNICAL COMPONENT ENVELOPE

Class "A" Documents

Legal Documents

1. Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) in accordance with Section 8.5.2 of the IRR

Note: For the purpose of updating the Certificate of Registration and Membership, all Class “A” eligibility documents mentioned in this section supporting the veracity, authenticity and validity of the Certificate shall remain current and updated. The failure by the prospective bidder to update its Certificate with the current and updated Class “A” eligibility documents shall result in the automatic suspension of the validity of its Certificate until such time that all of the expired Class “A” eligibility documents has been updated (per GPPB Resolution No. 15-2021).

Technical Documents

2. Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid. (*Use of Form No. DOBA-PCMC-SCF3b is required*)
3. Single Largest Completed Contract (SLCC) similar to the contract to be bid (*Refer to BDS ITB Clause 5.3*), except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 Revised IRR of RA No. 9184, within three (3) years prior to bid opening (*use of Form No. DOBA-PCMC-SCF3a is required*)
4. Original copy of Bid Security (**Refer to BDS Clause 14.1**). If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission or Original copy of the Notarized Bid Securing Declaration
5. Duly accomplished and signed Production/ Delivery Schedule using the form as provided for in Section VI
6. Duly accomplished and signed Technical Specification using the form as provided for in Section VII
 - i. Certification of Availability of demo unit (*use of Form No. DOBA-PCMC-DUF11 is required*)
 - ii. Signed conforme on the attached Engineering Clearance
 - iii. Brochures/sales literature reflecting the technical specifications

7. Original duly signed Omnibus Sworn Statement (OSS); and if applicable, Original Notarized Secretary's Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder. *(Use of the Form provided is required)*

Financial Documents

8. The prospective bidder's computation of the Net Financial Contracting Capacity (NFCC) must be at least equal to the ABC to be bid *(Use of Form No. DOBA-PCMC-NFF4 is required)*

OR

a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, it must be at least equal to 10% of the ABC

Class "B" Documents

9. If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence *(Use of Form No. DOBA-PCMC-JVF6 is required)*.

OR

Duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful

Other documentary requirements under RA NO. 9184 (as applicable):

- i. [For foreign bidders claiming by reason of their country's extension of reciprocal rights to Filipinos] Certification from the relevant government office of their country stating that Filipinos are allowed to participate in government procurement activities for the same item or product
- ii. Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity

The Bidder shall submit the following **FINANCIAL COMPONENT ENVELOPE (ARRANGED, NUMBERED AND TABBED)** *[Strictly NO using of staple wire and thick materials for tabs]* as enumerated below:

II. FINANCIAL COMPONENT ENVELOPE

1. Duly accomplished and signed Financial Bid Form
2. Duly accomplished and signed Price schedule (use applicable forms)
 - i. Use the form "*For Goods offered from within the Philippines*" if bidder is offering goods from within the Philippines.

- ii. Use the form “*For Goods offered from abroad*” if bidder is offering goods from Abroad.

Note: Bidder shall return to PCMC the issued **USB Flash Drive** containing the **FOLLOWING:**

- a. soft copy of their accomplished **Technical Specifications** (in **EXCEL format**)
- b. **SCANNED copy** (in **PDF Format**) of **ALL** the required documents under Section VIII. Checklist of Technical and Financial Documents

CONFORME:

Authorized Signatory (Signature over printed name)

Contact No:

Name of Company/Firm

Company’s Official Email Address
(where notices will be sent)

Company’s Official
Contact No.

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

TECHNICAL SPECIFICATIONS
IB-2022-142

Instruction: Accomplish this form by presenting a clear statement of your offer. Please write **the specific, precise and complete** statement which complies with the required specifications. **DO NOT** write "COMPLY" or the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT		BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description
1	Unit	VIDEONYSTAGMOGRAPHY SYSTEM
		I. HARDWARE COMPONENTS
		1. Caloric Air Irrigator
		With Touchscreen (capacitive) display
		Interface type: USB
		Power Requirements:
		• Power Supply Range: 220-230VAC, 50/60Hz
		• Power Consumption: Maximum of 45 watts
		• Power Cord: at least 3 meters
		• Fuse: 2 pcs rated at 315Ma/250V
		Range and Accuracy
		• Time display accuracy of +/- 0.5 seconds
		• Stimulus duration range of 10 to 210 seconds in steps of 5 seconds
		• Temperature display accuracy of +/- 0.1°C
		• Stimulus temperature range of minimum 20°C to a maximum of 50°C
		• Air Flow: minimum of 5,000 MI/MIN
		Connections
		• USB interface for control and configuration
		• With air filter connection
		General Requirements
		• USB interface connection to PC
		• With encodr control knob
		• Has an internal loud speaker to play signal tones
		• With built-in trigger button and LED Status indicator integrated in the irrigator gun
		• With color-coded indicator for cold and warm status
		• With countdown timer duration display during irrigation
		2. Optokinetic Stimulator Unit
		Capable of generating test pattern for optokinetic test
		With 2 USB 2.0 Ports for external peripherals connections
		With HDMI/RS232 Port for display test pattern
3. Posture test		
Dimensions: at least 60x80x20mm (WxDxH)		
Weight: at least 50g		
Voltage range: Max. 4.2V DC		
Power Consumption: Max. 2.5W		
Battery at least 650mAh		
4. Frenzel Test		
• Video Frenzel is compatible with Binocular VNG goggles		
• Camera for recording patient's position movement integrated with the PC		
II. SOFTWARE FEATURES		
• Videonystagmography (VNG)		
- Capable of calibrating the visual stimulation that can adapt to any room dimensions		
- Configurable test that can specify test names, test position, point color, point size, amplitude, latency, velocity, fixation time, etc.		
- Adapt image position		

	- Capable of adjusting the Gamma, Black level, Position of the image of the eye	
	- Capable of designing a template for chronological test procedure	
	- Capable of configuring different tests, Visual Positional and Calorics measurements	
	● Visual Test:	
	Saccade - Both Eyes, Spontaneous - Both Eyes, Gaze - Both Eyes	
	Tracking - Both Eyes, Optokinetic, Dix-Hallpike	
	● Positional Test:	
	Position - Both Eyes	
	● Bithermal Caloric Test:	
	Right - Warm, Left - Warm, Right - Cold, Left-Cold	
	- Calculation of all relevant parameters	
	Automatic calculation of statistical characteristics	
	● Average SPV	
	● Nystagmus Beats	
	● Unilateral Weakness and Directional Preponderance	
	● Scherer (Right Ear, Left Ear, Unilateral Weakness, Right Beating, Left Beating, Directional Preponderance)	
	● Jongkees (Right ear, Left Ear, Unilateral weakness, Right Beating, Left Beating, Directional Preponderance)	
	● Wexler (Unilateral Weakness)	
	● Unilateral Nystagmus	
	- Automatic nystagmus detection and manual nystagmus marking	
	- Nystagmus marking settings	
	1. Reversal point - Right beating above the curve	
	2. End of the nystagmus: Right beating below the curve	
	- Air Caloric	
	Capable of displaying diagrams	
	1. SPV Graph: shows the discrete velocities marked with color-coded dots	
	2. Cumulative Amplitude: shows the cumulated amplitudes of the slow phase	
	3. Claussen - Diagram: shows nystagmus beats during different culmination phases	
	4. Haid - Stoll Diagram: shows nystagmus beats during different culmination phases	
	5. Scherer - Diagram: shows the directional preponderance and unilateral weakness	
	- Capable of video recording for all measurements	
	- With Database Administration, Patient Manager	
	● Posture test	
	- With analysis for two directions (anterior-posterior, medial-lateral) and for both single measurements (eyes open, eyes closed)	
	- Automatically calculates different indices for every measurement such as Definition RQ (Velocity) Definition RQ (Area), Stability Index, Equilibrium Score)	
	- With automatic Unterberger's stepping test	
	● Frenzel Test	
	- Capable of measuring the following tests:	
	1. Spontaneous Nystagmus	
	2. Nystagmus Suppression with Integrated fixation light	
	3. Dix-Hallpike	
	4. Position/Positioning	
	- Can record video of the eye movement during the measurement procedure. Uses two cameras for each eye	

	- Capable of adding comments and recording speech with an integrated microphone	
	- Capable of configuring the measurement duration and the fixation light	
	- Has a video explorer for each patient	
	- Wide and bright display for both eyes without the use of magnifying glasses and lighting	
The unit will operate at 230 volts 3-pin Power Plug/Cable, strictlyly at 60 hertz		
Power Consumption (<i>must be clearly stipulated</i>)		
Mechanical parts should be of heavy duty durable type		
Equipment should be maintainable and serviceable		
Other Specifications: (please indicate if applicable)		
	BRAND:	
	MODEL:	
B. ACCESSORIES		
1.	Computer Specifications	
	Operating System: Windows 10	
	CPU: Intel Core i5 or higher	
	Storage: 512 SSD with 1TB	
	RAM: 16GB	
	Display: at least 55-inches LED Monitor	
	Display aspect ration: 4:3	
	External Hard Drive: 4TB	
2.	Binocular Googles	
	Googles with mask cover	
	Interface type: USB 2.0	
	Frame rate: 100 frames per second	
	Two video cameras (left and right)	
	With built-in fixation LED	
	With adjustable mirrors (3 positions)	
	Leather cushion with velcro	
	Power Requirements:	
	Voltage: +5V USB	
	Current: Maximum of 5 Watts	
	Infrared Light: according to EN 60825-1	
	Connections: 2 pcs mini-USB Type B to Type B USB 2.0 PC	
	Cable Length: minimum of 3 meters	
3.	Foot Switch	
	Current: Maximum of 25 milli-Ampere	
	Power Consumption: Less than 1 Watt	
	Cable Length: at least 3 meters	
	Connections: 1 pc mini-USB Type B to Type B USB 2.0 PC	
4.	Uninterruptible Power Supply (UPS) - at least 1KVA	
C. WARRANTY		
1.	Warranty at least 2 years on parts and labor	
2.	Quarterly Preventive Maintenance and calibration during warranty period	
3.	System and software upgrade, free of charge within warranty period	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME

TELEPHONE / FAX NO.

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

TECHNICAL SPECIFICATIONS
IB-2022-143

Instruction: Accomplish this form by presenting a clear statement of your offer. Please write **the specific, precise and complete** statement which complies with the required specifications. **DO NOT** write "COMPLY" or the page numbers of the brochure/data sheet, etc.

QTY	UNIT	PCMC REQUIREMENT Item Description	BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
1	Unit	WIRELESS HANDHELD ULTRASOUND	
		A. MAIN SPECIFICATIONS	
		1. Patient Compatibility: Pediatric	
		2. Applications/ Software Lung preset (inclusive) Operating system compatible with software, upgradeable operating system at no cost once OS platform is not applicable to run the software <ul style="list-style-type: none"> • DICOM ready • Needle Enhancement • Advanced MSK Presets • Pulsed Wave Doppler • Elastography 	
		Scanners	
		Multipurpose HD Convex: Frequency 2-6 MHz Depth up to 40 cm Specialty: Abdominal, Lung , OB/GYN, Cardiac, Superficial, Vascular	
		Specialty HD Linear: Frequency 4-13 MHz Depth up to 11 cm Specialty: Vascular, Lung , Musculo Skeletal, Small Parts, Nerve	
		3. Probe Transducers Standard Features	
		Wireless <ul style="list-style-type: none"> • handheld point and shoot scanning device • a one-click scanning device connected to the App, allowing remote access to the scanned results via iOS and Android devices 	
		Durable and Watertight <ul style="list-style-type: none"> • compact, waterproof and fully immersible for easy disinfection and sanitation • made of magnesium shell that proves of its high quality and its heavyweight guarantees that the device would never land on the tip/head end 	
		Imaging TRANSMISSION <ul style="list-style-type: none"> • 1 to 20 MHz waveforms • Bi-polar output • Up to 20 continuous pulses • 10 to 80V peak-to-peak 	
		Beamforming Technology that provides 8x faster frame rates and 8x more data processed for higher resolution	
		Secured Storage has accessible digital storage through iCloud a secure web-based portal that allows you to save, review, manage, and share your scanned images	
		4. Imaging modes B-Mode M-Mode/ Color M-mode Color Doppler Imaging Power Doppler Imaging/Directional PDI Pulsed Wave Doppler Tissue Doppler Imaging (TDI) Continous Wave Doppler (CWD)	
		5. Device 2 units tablet latest gen OPERATING SYSTEM <ul style="list-style-type: none"> • iAndroid 5 (API 21)+ (1 unit tablet) • iOS 11.0+ (1 unit tablet) 	

	<p>WIRELESS</p> <ul style="list-style-type: none"> • Bluetooth Low Energy 4.0 • Wi-Fi 802.11n (5GHz) <p>SECURITY & ENCRYPTION</p> <ul style="list-style-type: none"> • Wi-Fi : TLS 1.2 • Bluetooth : AES128 and RSA4096 	
6.	The unit will operate at 230 volts 3-Pin Power Plug/Cable,	
7.	With built-in battery back-up system for detection and life saving equipment and no separate volt conversion	
8.	Power Consumption: <i>(must be clearly stipulated)</i>	
9.	Mechanical parts should be of heavy duty durable type	
10.	Equipment should be maintainable and serviceable	
Other Specifications: (please indicate if applicable)		
	BRAND:	
	MODEL:	
B. ACCESSORIES		
1.	Two (2) units Fan Probe attaches easily to probe transducers; Two (2) shockproof case	
2.	Two(2) units Arm Holder can hold phones and tablets sized 4"-11" weighing up to 1kg	
3.	Two (2) units charging station with built-in lithium battery which provides up to 60 min of scanning time	
4.	Two (2) pieces CROSS-BODY BAG The cross-body bag can hold two HD scanners, an HD charger, a small tablet and a bottle of gel.	
5.	One (1) unit All-in-one Ink Tank System Printer (WiFi, Printer, Scanner, Copier)	
C. CONSUMABLES TO BE PROVIDED PER UNIT		
1.	Two (2) gallons of UTZ gel with 250 ml bottle	
2.	Four (4) sets High grade level 4 PPE	
D. WARRANTY		
1.	Warranty at least three (3) years on parts and labor	
2.	International warranty: three (3) years on probes; six (6) months on accessories	
3.	Quarterly Preventive Maintenance and calibration during the	
4.	Availability of Service Engineer within 24 hours - 24/7	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

NAME OF COMPANY	ADDRESS
SIGNATURE OVER PRINTED NAME	TELEPHONE / FAX NO.

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

Invitation to Bid No. IB 2022-144

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

TECHNICAL SPECIFICATIONS

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
2	unit	Bed Bath Stretcher	
A. SPECIFICATIONS			
	1	With height adjustment through pump and with tilt angle easy for drainage	
	2	High quality stainless steel (Medical Grade)	
	3	Hydraulically height adjustment by foot pedal	
	4	With guard rail, for safety and designed for dewatering	
	5	With covered castor wheels approximately 3-5 inches in diameter	
	6	Load bearing of 100-200 kg	
	7	Equipped with four bumper wheels	
	8	Mechanical parts should be of heavy duty durable type	
	9	Equipment should be maintainable and serviceable	
		Other Specifications (Please indicate if applicable)	
		BRAND:	
		MODEL:	
B.WARRANTY			
	1	Warranty at least two (2) years on parts and two (2) years on labor	

Additional Requirements

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

AUTHORIZED REPRESENTATIVE'S INFOS:

Signature over Printed Name	Designation/Position
-----------------------------	----------------------

COMPANY'S DETAILS:

NAME OF COMPANY

ADDRESS

COMPANY'S OFFICIAL EMAIL ADDRESS	TEL. NO. / FAX NO.
----------------------------------	--------------------

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

Invitation to Bid No. IB 2022-145

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

TECHNICAL SPECIFICATIONS

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	Pediatric Orthopedic Traction bed Package with bedside table overbed table and foot stool with IV pole	
A. SPECIFICATIONS			
		1 Electrically operated and adjustable for heights Trendelenburg/Reverse	
		2 Four section Mattress	
		3 Bed Frame made of steel with powder coating	
		4 Removable head and foot panels with weightlifting bearing capacity 25 kg.with accessory pocket two (2) pockets	
		5 With movable x-ray permeable backrest or with x-ray permeable mattress	
		6 Base frame and support frame should be made of steel for long life and prevention from rusting	
		7 Equipped with for articulated half-length closed tuck away side rails polypropylene medical grade plastic materials with lock	
		8 Equipped with four large castor 5- 7 inches in diameter with central brakes and steering facility	
		9 With high density foam and water polyurethane 5-7 inches thickness	
		10 Bumping in all corners	
		11 Safe working load: 200-300 kg	
		12 One IV pole with four hooks with lifting bearig capacity 25 kg with accessory pocket	
		13 Two drainage hook under the bed frame on both sides	
		14 Unit will operate at 230 volts 3- Pin Power Plug/Cable, strictly at 60 hertz;	
		15 Power Consumption of the unit must be clearly stipulated	
		16 Mechanical parts should be of heavy duty durable type	

		17	Equipment should be maintainable and serviceable	
		Other Specifications (Please indicate if applicable)		
			BRAND:	
			MODEL:	
B. ACCESSORIES				
		1	Table size: Length approximately 700-800 mm with 380-450 mm <ul style="list-style-type: none"> • Hard plastic table material • Adjustable height approximately 645-900mm • With Heavy duty caster wheels with brakes • Heavy duty steel base 	
		2	Bedside Table <ul style="list-style-type: none"> • Drawer pull table top- one storage cabinet with horizontal shelf • With handles bars on both sides and four heavy duty ball caster wheels with lock • Metal Body: ABS plastic top or full ABS har plastic • Size: approximately L 480=W500x700mm 	
		3	Foot Stool <ul style="list-style-type: none"> • Two steps ladder • Made of stainless steel • Load capacity: approximately 350 kg • Size: approximately H 22cmxW25cmx L 35cm • Non Slip corrugated rubber top surface • Heavy duty type mechanical part 	
C.WARRANTY				
		1	Warranty at least two (2) years on parts and two (2) years on labor	
		2	Quarterly Preventive Maintenance during warranty period	

Additional Requirements

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

AUTHORIZED REPRESENTATIVE'S INFOS:

Signature over Printed Name	Designation/Position
------------------------------------	-----------------------------

COMPANY'S DETAILS:

NAME OF COMPANY

ADDRESS

COMPANY'S OFFICIAL EMAIL ADDRESS	TEL. NO. / FAX NO.
---	---------------------------

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

Invitation to Bid No. IB 2022-146

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

TECHNICAL SPECIFICATIONS

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
11	unit	Stainless Medicine Cabinet Heavy Duty	
A. SPECIFICATIONS			
	1	Stainless steel type 304 gauge 20 hairline finish	
	2	Two swing type glass door with lock	
	3	Five compartment adjustable shelf	
	4	Approximate dimension at least Length: 85cm Height: 180cm Width: 40cm	
	5	Mounted on four pcs.4 inches (100mm) diameter heavy duty mold on rubber caster wheels with stainless swivel and wheel lock (all metal fork, brake and wheel hub)	
	6	Mechanical parts should be of heavy duty durable type	
	7	Equipment should be maintainable and serviceable	
		Other Specifications (Please indicate if applicable)	
		BRAND:	
		MODEL:	
C. ACCESSORIES (Per Unit)			
	1	With Led light (strips) in every compartment	
B.WARRANTY			
	1	Warranty at least two (2) years on parts and two (2) years on labor	

Additional Requirements

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

AUTHORIZED REPRESENTATIVE'S INFOS:

Signature over Printed Name	Designation/Position
------------------------------------	-----------------------------

COMPANY'S DETAILS:

NAME OF COMPANY

ADDRESS

COMPANY'S OFFICIAL EMAIL ADDRESS	TEL. NO. / FAX NO.
---	---------------------------

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

Invitation to Bid No. IB 2022-147

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

TECHNICAL SPECIFICATIONS

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
4	unit	Phototherapy Blue Light	
A. SPECIFICATIONS			
		1 Incorporate a state- of -the art light source for the treatment of new-born jaundice	
		2 Deliver Intensive phototherapy: 35-110W/cm ² /nm	
		3 Utilize blue light emitting diodes (LEDs) with wavelength 450-470nm	
		4 Light enclosure is compact in size and light weight: maximum of 36-48cm x 28-38cm x 60-80cm	
		5 Movable and adjustable pole: 53cmx55cmx170cm at maximum height	
		6 Adjusted both horizontally and vertically and tilted over a wide angle range (up to 90 degress)	
		7 Base roll stand four wheels with lock	
		8 The unit will be operated at 230 volts, 3 Pin power plug/cable, strictly at 60 hertz	
		9 No separate volt conversion	
		10 Power Consumption: 32 watts, 2 amp	
		11 Mechanical parts should be of heavy duty durabl type	
		12 Equipment should be maintainable and serviceable	
		Other Specifications (Please indicate if applicable)	
		BRAND:	
		MODEL:	
B. ACCESSORIES (Per Unit)			
		1 One (1) unit UPS with frequency converter per unit	
		BRAND:	
		MODEL:	
		2 Will provide one (1) unit Phototherapy light meter for the total number of units	
C. CONSUMABLES			
		1 Will provide five (5) disposable eye cover/protector per unit (3 small, 2 medium)	
		2 Complete set of Led Light bulbs of 50,000 hrs.life span	

	D.WARRANTY		
	1	With one (1) year warranty on parts and services	
	2	With quarterly maintenance and calibration with certificate and sticker	
	3	Conduct proper training for operations for end-user	

Additional Requirements

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

AUTHORIZED REPRESENTATIVE'S INFOS:

Signature over Printed Name	Designation/Position
------------------------------------	-----------------------------

COMPANY'S DETAILS:

NAME OF COMPANY

ADDRESS

COMPANY'S OFFICIAL EMAIL ADDRESS	TEL. NO. / FAX NO.
---	---------------------------

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

Invitation to Bid No. IB 2022-148

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

TECHNICAL SPECIFICATIONS

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
2	unit	Weighing Scale for Neonates	
A. SPECIFICATIONS			
		1 Weight Capacity of maximum of 5 kg	
		2 LCD Display	
		3 With length measuring rod ranges from 35-80 cm	
		4 Removable and easy to clean weighing cradle	
		5 Extra-large curved weighing tray size approximately: W-20-25 inches x H-4-6 inches x D-10-11 inches	
		6 Can be used with rechargeable Batteries and with AC adapter/charger	
		7 With auto-HOLD and TARE feature	
		8 Programmable Auto Off	
		9 With functions: LB/KG Conversion lockout	
		10 Equipment parts are Heavy Duty	
		11 With utility cart 2 layers and four caster wheels with lock	
		12 With Built in battery back-up for detection and no separate volt conversion	
		13 Power Consumption of the unit must be clearly stipulated	
		14 The unit will be operated at 230 volts, 3 Pin power plug/cable, strictly at 60 hertz	
		15 Mechanical parts should be of heavy duty durable type	
		16 Equipment should be maintainable and serviceable	
		Other Specifications (Please indicate if applicable)	
		BRAND:	
		MODEL:	

	B.WARRANTY		
	1	With one (1) year warranty on accessories,parts and services	
	2	With free quarterly preventive maintenance and calibration	

Additional Requirements

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

AUTHORIZED REPRESENTATIVE'S INFOS:

Signature over Printed Name

Designation/Position

COMPANY'S DETAILS:

NAME OF COMPANY

ADDRESS

COMPANY'S OFFICIAL EMAIL ADDRESS

TEL. NO. / FAX NO.

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

Invitation to Bid No. IB 2022-149

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

TECHNICAL SPECIFICATIONS

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
10	unit	Portable Digital Sphygmomanometer	
A. SPECIFICATIONS			
		1 With Pulse rate	
		2 With air hose suited for neonate, infant, child and adult BP cuff	
		3 With BP cuff of different sizes	
		4 With at least 3 meter power cord	
		5 With stand with wheels and lock , and with basket for storage purposes	
		6 With built in battery back up system for detection and life saving equipment and no separate volt conversion	
		7 Power Consumption of the unit must be clearly stipulated	
		8 Unit will operate at 230 volts 3-Pin Power Plug/Cable, strictly at 60 hertz	
		9 Mechanical parts should be of heavy duty durable type	
		10 Equipment should be maintainable and serviceable	
		Other Specifications (Please indicate if applicable)	
		BRAND:	
		MODEL:	
B.ACCESSORIES (Per Unit)			
		1 BP cuff of different sizes	
C.CONSUMABLE TO BE PROVIDED PER UNIT			
		1 Rechargeable battery approximately 6 hours	
		2 BP cuffs of different sizes	
		3 BP Valve	
B.WARRANTY			
		1 Warranty at least two (2) years on parts and two (2) years on labor	
		2 Quarterly Preventive Maintenance	

Additional Requirements

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

AUTHORIZED REPRESENTATIVE'S INFOS:

Signature over Printed Name	Designation/Position
------------------------------------	-----------------------------

COMPANY'S DETAILS:

NAME OF COMPANY

ADDRESS

COMPANY'S OFFICIAL EMAIL ADDRESS	TEL. NO. / FAX NO.
---	---------------------------

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

Invitation to Bid No. IB 2022-150

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

TECHNICAL SPECIFICATIONS

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
14	unit	Refrigerator for Medicine (+2- +8)	
A. SPECIFICATIONS			
		1 Capacity 60L, 3 layers	
		2 External size (WxDxH)mm 495x548x683	
		3 Internal size (WxDxH)mm 415x405x455	
		4 Temperature range 2 ⁰ C-8 ⁰ C	
		5 Temperature accuracy 0.1 ⁰ C	
		6 Control System: microprocessor control	
		7 Alarm: Audible and visual alarm for high and low temperature. Door ajar	
		8 Refrigerator Type: force air refrigerator system	
		9 Refrigerant R600a	
		10 Climate Type N SN	
		11 Glass Door: Toughened with electric heating function, no condensation	
		12 Construction-contraction: Unibody design & High pressure and High Density Cyclopentane Foaming	
		13 Internal Material: cold rolled steel coated with anti bacterial powder	
		14 External cold rolled steel coated with anti bacterial powder	
		15 Material: basket 1 pc	
		16 Consumption: 95 W	
		17 Power Supply AC220V <u>+10% 60 Hz</u>	
		18 Shelves: 3 pcs cold drawn steel wire impregnated adjustable shelf	

		19	Door Lock: 1 pc	
		20	Gross Weight (kg) 49	
		21	Power Consumption of the unit must be clearly stipulated	
		22	The unit will be operated at 230 volts, 3 Pin power plug/cable, strictly at 60 hertz	
		23	Mechanical parts should be of heavy duty durable type	
		24	Equipment should be maintainable and serviceable	
		Other Specifications (Please indicate if applicable)		
			BRAND:	
			MODEL:	
	B. ACCESSORIES			
		1	Automated Voltage Regulator (AVR) 1 KVA SERVO type	
			BRAND:	
			MODEL:	
	B.WARRANTY			
		1	Warranty at least two (2) years on parts and two (2) years on labor	

Additional Requirements

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

AUTHORIZED REPRESENTATIVE'S INFOS:

Signature over Printed Name **Designation/Position**

COMPANY'S DETAILS:

NAME OF COMPANY

ADDRESS

COMPANY'S OFFICIAL EMAIL ADDRESS **TEL. NO. / FAX NO.**

PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City

TECHNICAL SPECIFICATIONS

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		<i>A. Specifications</i>	
		1. Main Equipment:	
		<ul style="list-style-type: none"> portable touch screen cardiac monitor with transducer cables for non-invasive and invasive monitoring of vital signs and intravascular pressures that can be mounted on top of anesthesia machine for intra-operative use in the operating room, transported with trolley stand and can be wall mounted for post-operative use in the surgical intensive care unit 	
		<ul style="list-style-type: none"> interface: with LAN and wireless capability enabling communication with compatible central monitoring system both from bedside and during transport 	
		<ul style="list-style-type: none"> several modes and basic parameters with optional configuration 	
		<ul style="list-style-type: none"> measurements include but not limited to: ECG with waveforms, respiratory rate with waveform, pulse oximetry / SpO2 with waveform, capnography / ETCO2 ,onitoring with mainstream sensor, non-invasive blood pressure (NIBP) with waveform, invasive pressures with waveforms, temperature monitoring with neonatal/pediatric/adult probe, non-invasive and invasive CO monitoring 	
		2. Patient Compatibility:	
		<ul style="list-style-type: none"> suitable for intra-operative and post-operative cardiac monitoring of adults, adolescents/young adults, children and neonates in the OR and and ICU 	
		3. ECG:	
		<ul style="list-style-type: none"> 3/6/12 lead ECG with adjustable waveform size and sweep speeds of 1.56, 6.25, 12.5, 25, 50 mm/s 	
		<ul style="list-style-type: none"> arrhythmia and pacing recognition 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		<ul style="list-style-type: none"> • ECG interpretation and arrhythmia analysis including ST segment analysis 	
		<ul style="list-style-type: none"> • capable of measuring 18-lead ECG 	
		<ul style="list-style-type: none"> • monitoring system standard in all lead sets 	
		<ul style="list-style-type: none"> • integral 3-trace thermal recorder providing waveform and data reports for diagnostic support continuously variable steering in 2D, color Doppler and Doppler modes 	
		4. Respiratory Rate:	
		<ul style="list-style-type: none"> • respiration measured by impedance method with waveform 	
		<ul style="list-style-type: none"> • wide ranges of respiratory rate detected 	
		<ul style="list-style-type: none"> • apnea detection 	
		5. Pulse Oximetry / SpO2	
		<ul style="list-style-type: none"> • uses Masimo or Nellcor technology for SpO2 monitoring 	
		<ul style="list-style-type: none"> • wide ranges of rate detected 	
		<ul style="list-style-type: none"> • appropriate waterproof and shock-proof SpO2 sensor for adults and older children (finger clip), infants (finger rubber glove or wraparound rubber) and neonates (multisite Y probe or wraparound rubber) 	
		<ul style="list-style-type: none"> • SpO2 Monitoring with pleth waveform as visual indicator of patient's pulse and perfusion indicator as numerical value for the pulsatile portion 	
		6. Capnography / ETCO2 Monitoring	
		<ul style="list-style-type: none"> • ET-CO2 measurement and monitoring with mainstream CO2 sensor for intubated and non-intubated patients 	
		<ul style="list-style-type: none"> • warm-up time is ≤ 5 secs 	
		<ul style="list-style-type: none"> • wide ranges of measurements detected 	
		<ul style="list-style-type: none"> • CO2 value display update cycle every 3-5 secs or when alarm is generated with abnormal values 	
		7. Non-invasive Blood Pressure (NIBP)	
		<ul style="list-style-type: none"> • non-invasive blood pressure (with MAP) with waveform measurable manually, automatically and stat using oscillometric method with selectable color display 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		<ul style="list-style-type: none"> • wide ranges of measurements detected 	
		<ul style="list-style-type: none"> • quality assurance of measurement accuracy 	
		<ul style="list-style-type: none"> • step deflation measurement with custom series mode and NIBP timestamp shows the most recent or until the next NIBP measurement 	
		<ul style="list-style-type: none"> • fast and pain-free measurement with cuff inflation time <12 secs for adults and children and <6 secs for neonates and infants 	
		8. Invasive Blood Pressure	
		<ul style="list-style-type: none"> • at least 4 channels of invasive blood pressure monitoring including ART, ABP, CVP, PAP where at least 3 waves can be simultaneously displayed on the screen monitor with selectable color display 	
		<ul style="list-style-type: none"> • wide ranges of measurements detected 	
		<ul style="list-style-type: none"> • zero calibration mode 	
		<ul style="list-style-type: none"> • universally compatible with existing transducers in the market 	
		9. Temperature Monitoring	
		<ul style="list-style-type: none"> • appropriate waterproof temperature probes for external (skin) and internal placement (intranasal or equivalent) 	
		<ul style="list-style-type: none"> • wide ranges of measurements detected 	
		<ul style="list-style-type: none"> • quality assurance of measurement accuracy 	
		10. Non-Invasive CO Monitoring	
		<ul style="list-style-type: none"> • CO measurement using non-invasive parameters (ECG, pulse wave, SpO2, blood pressure) continuously with acceptable clinical accuracy 	
		<ul style="list-style-type: none"> • wide ranges of measurements detected 	
		11. Invasive CO Monitoring	
		<ul style="list-style-type: none"> • capable of invasive CO measurement using standard thermodilution methods 	
		<ul style="list-style-type: none"> • wide ranges of measurements detected 	
		12. Display Screen and Parameters:	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		<ul style="list-style-type: none"> • large resistive touch screen at least 15 inches with color TFT LCD resolution 1280 x 800 pixel with >30 colors and at least 12 selectable display colors 	
		<ul style="list-style-type: none"> • screen color display resolution minimum 1280 x750 pixel 	
		<ul style="list-style-type: none"> • measurement parameters include: ECG (3/6/12 lead), respiration (impedance method), SpO2, SpO2-2, NIBP, IBP, temperature, cardiac output, CO2 (mainstream method), (TOF-Train of Four, BIS, CO2 	
		<ul style="list-style-type: none"> • moving or fixed waveform display mode 	
		<ul style="list-style-type: none"> • sweep speed: normal =25, 50 mm/s; slow = 1.56, 6.25, 12.5 mm/s 	
		<ul style="list-style-type: none"> • at least 15 waveforms graphical and numerical display in selectable screen layouts including simultaneous display of at least three (3) pressure readings 	
		<ul style="list-style-type: none"> • can be wall mounted with bracket easily detachable for bedside with carrying handle for easy transport from one 	
		<ul style="list-style-type: none"> • at least 72 hours full disclosure, with tabular and graphics trends including ECG tracing, arrhythmia recall and pressure recording events and printable 	
		<ul style="list-style-type: none"> • with slave monitoring or secondary monitor for OR use and display 	
		<ul style="list-style-type: none"> • with USB port for download of service logs, import/export settings, installing software 	
		13. Printer:	
		<ul style="list-style-type: none"> • integral 3-trace thermal recorder providing waveform and data reports for diagnostic support 	
		<ul style="list-style-type: none"> • printer for each unit with standard thermal paper <input type="checkbox"/> 	
		14 Alarm:	
		<ul style="list-style-type: none"> • alarm items: upper/lower limit alarms, arrhythmia alarms, interbed alarms, technical alarms 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		<ul style="list-style-type: none"> • alarm system with level priority prompt (high, medium, low message) with audible and visual notification 	
		<ul style="list-style-type: none"> • audible alarm sound with adjustable loudness 	
		<ul style="list-style-type: none"> • visible blinking alarm light from 360 degrees 	
		<ul style="list-style-type: none"> • auto and manual setting of alarm limits and with alarm suspend 	
		<ul style="list-style-type: none"> • with alarm history of at least 3 days 	
		<ul style="list-style-type: none"> • dual internal speakers 	
		15. Review:	
		<ul style="list-style-type: none"> • at least 72 hours storage capacity of trend graphs 	
		<ul style="list-style-type: none"> • at least 72 hours storage capacity of vital signs list 	
		16. Other Features:	
		<ul style="list-style-type: none"> • interbed capability with central monitor providing data viewing of each patient monitor with individual display and 10-20 bed display 	
		<ul style="list-style-type: none"> • drug calculation capability for flow rates and dosages for medication and titration 	
		<ul style="list-style-type: none"> • compact modular unit with weight <7 kgs 	
		<ul style="list-style-type: none"> • built in demo mode for end user training without the need of simulators 	
		<ul style="list-style-type: none"> • data is ready to print via central monitoring system 	
		<ul style="list-style-type: none"> • mechanical parts should be heavy duty and hospital grade 	
		<ul style="list-style-type: none"> • equipment should be maintainable 	
		<ul style="list-style-type: none"> • dashboard connectivity ready 	
		17. Connectors:	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		<ul style="list-style-type: none"> dedicated single output connectors or multifunctional output connectors to simultaneously display ten (10) minimum parameters: heart rate, ECG, respiratory rate, SpO2, ET/CO2, temperature, NIBP, at least 3 invasive pressure readings (ABP, CVP and PAP) 	
		<ul style="list-style-type: none"> with AC power connector, data port network connector, USB connector, DVI or display port connector 	
		<ul style="list-style-type: none"> thermal printer 	
		18. Accesories:	
		<ul style="list-style-type: none"> one (1) unit transport monitor: display size is 5-6 inches color TFT; non-fade, fixed waveform display mode; viewing area at least 115 x 85 mm; maximum number of 9 traces (12 traces for 12-lead ECG measurement); sweep speed of 50, 25, 12.5 or 6.25 mm/s; waveform display colors: 9 traces (12 traces for 12-lead ECG measurement) 	
		<ul style="list-style-type: none"> movable trolley with basket and easy maneuvering brakelock 	
		<ul style="list-style-type: none"> wall mount brackets 	
		<ul style="list-style-type: none"> removable printer modules and thermal printer 	
		<ul style="list-style-type: none"> CO inline sensor and injectate cable 	
		<ul style="list-style-type: none"> connector cables for but not limited to: NIBP #1, invasive pressures #3, ECG #1, pulse oximeter #1, temperature probes #2 skin and nasal/rectal, ET/CO2 #1, CO #1 	
		<ul style="list-style-type: none"> reusable waterproof, shock-proof pulse oximeter sensors (2 pcs adult, 2 pcs pedia and 2 pcs neonate) 	
		<ul style="list-style-type: none"> BP cuff dual hose (2 reusable each) for the following sizes: big adult, adolescent or small adult, child/pedia, infant, neonate 	
		<ul style="list-style-type: none"> with USB port for additional connectivity, importing and exporting protocols and exams, archiving and serviceability 	
		<ul style="list-style-type: none"> 3/5-electrode ECG and wires 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		• reusable CO sensor (at least 1)	
		• air hose tubing for NIBP cuffs	
		• power cord	
		• battery pack	
		19. Power:	
		• includes AC adapter, power cord and system battery pack	
		• operated at 100-240V, strictly ast 50/60 Hz	
		• power consumption stipulated	
		• rechargeable internal battery pack that is safe	
		• type of battery: lithium ion	
		• at least two (2) hours battery operation time for continuous monitoring	
		• with less than 4 hours recharging of battery	
		20. Durability:	
		• medical grade unit that can withstand frequent transport from OR to ICU	
		• mechanical parts should be of heavy-duty type	
		• equipment should be maintainable	
		21. Conforms to the attached PCMC Engineering Clearance Requirements	
		<i>Other Specifications: (please indicate if applicable)</i>	
		BRAND:	
		MAKE / MODEL:	
		B. Accessories	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		1 • one (1) unit transport monitor: display size is 5-6-inches color TFT; non-fade, fixed waveform display mode; viewing area at least 115 x 85 mm; maximum number of 9 traces (12 traces for 12-lead ECG measurement); sweep speed of 50, 25, 12.5 or 6.25 mm/s; waveform display colors: 9 traces (12 traces for 12-lead ECG measurement)	
		2 • movable trolley with basket and easy maneuvering brakelock	
		3 • wall mount brackets	
		4 • removable printer modules and thermal printer	
		5 • connector cables for but not limited to: NIBP #1, invasive pressures #3, ECG #1, pulse oximeter #1, temperature probes #2 skin and nasal/rectal, ETCO2 #1, CO #1	
		6 • reusable waterproof, shock-proof pulse oximeter sensors (2 pcs adult, 2 pcs pedia and 2 pcs neonate)	
		7 • BP cuff dual hose (2 reusable each) for the following sizes: big adult, adolescent or small adult, child/pedia, infant, neonate	
		8 • with USB port for additional connectivity, importing and exporting protocols and exams, archiving and serviceability	
		9 • 3/5-electrode ECG and wires	
		10 • reusable CO sensor (at least 1)	
		11 • air hose tubing for NIBP cuffs	
		12 • power cord	
		13 • battery pack	
		C. Consumables	
		1 • mainstream CO2 sensor kits and airway adapter for End-tidal Carbon Dioxide: 2 sets disposable for mainstream for each unit	
		2 • ten (10) pcs. pressure transducers (Utah type)	
		3 • three (3) packs disposable ECG electrodes	
		4 • three (3) packs disposable ECG electrodes	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
3	unit	High-End Cardiac Monitor with Invasive Pressure Monitoring	
		5 • five (5) packs (50 wipes/pack)	
		6 • five (5) pcs 1TB USB	
		7 • one (1) bottle of probe cleaner or equivalent	
		8 • six (6) rolls of thermal paper for printer	
		9 • two (2) mouth guards	
		D. Warranty	
		1 • three (3) years international warranty on parts, services and accessories	
		2 • quarterly preventive maintenance and system optimization during the warranty period	
		3 • certification from the bidder / supplier that in the event of malfunction, they must provide an immediate back up / loaner unit while the delivered unit is being repaired.	
		4 • certification guarantee from supplier on availability of all spare parts for the next 5 years (at least) from testing, commissioning, acceptance and delivery	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

NAME OF COMPANY	ADDRESS
------------------------	----------------

SIGNATURE OVER PRINTED NAME	TELEPHONE / FAX NO.
------------------------------------	----------------------------

PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City

TECHNICAL SPECIFICATIONS

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Cardiac Ultrasound with Advanced Applications and Interchangeable Probes	
		<i>A. Specifications</i>	
		1. Main Equipment:	
		<ul style="list-style-type: none"> • color Doppler cardiac ultrasound machine with high-end model and system designed to handle high volume procedures and high output work flow 	
		<ul style="list-style-type: none"> • continuous, real-time adjustment of system gain and time gain compensation achieving balanced brightness of tissues providing solid image quality 	
		<ul style="list-style-type: none"> • real-time compounding, with widescreen capability and multiple beam-steered lines of sight acquiring more information and reducing angle-generated artifacts 	
		<ul style="list-style-type: none"> • has image processing for noise and artifact reduction improving tissue and border definition 	
		<ul style="list-style-type: none"> • fully configurable, allowing end-user to select imaging capabilities, transducers, and clinical analysis to support exam needs with option to add supported capabilities at any time as well as upgrades when they become available 	
		2. Patient Compatibility:	
		<ul style="list-style-type: none"> • suitable for station, bedside and intra-procedural echocardiography of adults, adolescents/young adults, children, neonates and fetus 	
		3. Features:	
		<ul style="list-style-type: none"> • automatic scan optimization and tissue specific imaging presets 	
		<ul style="list-style-type: none"> • one-touch image optimization 	
		<ul style="list-style-type: none"> • low noise, wide dynamic range digital 	
		<ul style="list-style-type: none"> • ECG 3 leads display 	
		<ul style="list-style-type: none"> • advanced adaptive image processing 	
		<ul style="list-style-type: none"> • continuously variable steering in 2D, color Doppler and Doppler modes 	
		<ul style="list-style-type: none"> • one-button controls 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Cardiac Ultrasound with Advanced Applications and Interchangeable Probes	
		<ul style="list-style-type: none"> • post-processing for recalled images in 2D (gain), M- mode reconstruction (2D), Color (gain), Doppler (baseline/sweep speed, scale) 	
		<ul style="list-style-type: none"> • support for radiology, vascular, and cardiology interventional procedures 	
		<ul style="list-style-type: none"> • upgradeable with new modalities and analysis in the future 	
		<ul style="list-style-type: none"> • upgradeable with additional interchangeable probes 	
		<ul style="list-style-type: none"> • comprehensive measurement tools with dedicated pediatric and adult congenital cardiac analysis; fetal echo analysis; vascular analysis 	
		<ul style="list-style-type: none"> • easy transport from room-to-room with sleep mode in <30 secs and transport mode lasts at least 30 minutes before recharge 	
		4. Operating / Imaging Modes:	
		<ul style="list-style-type: none"> • Real-time/2D imaging 	
		<ul style="list-style-type: none"> • M-mode: anatomical and color 	
		<ul style="list-style-type: none"> • Chroma imaging with multiple color maps 	
		<ul style="list-style-type: none"> • Dual mode imaging for 2D and color flow 	
		<ul style="list-style-type: none"> • Color compare mode 	
		<ul style="list-style-type: none"> • Color flow Doppler 	
		<ul style="list-style-type: none"> • Pulsed wave (PW) and high-PRF PW Doppler 	
		<ul style="list-style-type: none"> • Continuous wave (CW) Doppler and high-PRF CW Doppler 	
		<ul style="list-style-type: none"> • M-Mode and Doppler review 	
		<ul style="list-style-type: none"> • Tissue Doppler Imaging (TDI), TDI M-Mode 	
		<ul style="list-style-type: none"> • Coronary PW sub-mode 	
		<ul style="list-style-type: none"> • Pulse inversion 	
		<ul style="list-style-type: none"> • Cineloop acquisition (time and beats) with long capture 	
		<ul style="list-style-type: none"> • Cineloop review with trackball control of image selection 	
		<ul style="list-style-type: none"> • Panoramic 2D imaging 	
		<ul style="list-style-type: none"> • Tissue harmonics 	
		<ul style="list-style-type: none"> • Global longitudinal strain 	
		<ul style="list-style-type: none"> • 2D Transesophageal Echocardiography (TEE) 	
		<ul style="list-style-type: none"> • 3D Transthoracic Echocardiography (TTE) capability 	
		<ul style="list-style-type: none"> • 3D Transesophageal Echocardiography (TEE) capability 	
		5. Applications / Software Preset	
		<ul style="list-style-type: none"> • Pediatric Cardiology clinical option 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Cardiac Ultrasound with Advanced Applications and Interchangeable Probes	
		• Neonatal Cardiology clinical option	
		• Fetal Cardiology clinical option	
		• Adult Cardiology clinical option	
		• Stress Echocardiography clinical option	
		• Vascular Cardiology clinical option	
		• Transcranial Doppler option	
		• Custom examination protocols	
		• Needle visualization	
		• Live plane imaging	
		6. Measurement Tools and Quantification	
		• region of interest (ROI) or its equivalent	
		• Intima Media Thickness (IMT) or its equivalent	
		• fully automated GLS tool for LV or its equivalent	
		• strain quantification or its equivalent	
		• automated 2D quantification or its equivalent	
		• automated 2D quantification for LA	
		7. Display Screen and Controls:	
		• at least 21 inches high definition flat panel LED display with wide viewing angle and automatic ambient light compensation	
		• full-color 12-inch capacitive touchscreen	
		• can be folded down for easy maneuverability and transport	
		• full articulation of extension arm for easy viewing in any angle and location	
		• full articulation of control panel and monitor allowing for proper alignment whether sitting or standing for ergonomic scanning	
		• common user experience control panel with central track ball and easy-access mode keys	
		• with quick keys and active mode	
		8. Transducers:	
		• at least three (3) with the following frequencies	
		1. matrix array cardiac transducer probe for use on adult/pedia and compatible with 2D and 3D imaging with frequency range between 1-5 MHz	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Cardiac Ultrasound with Advanced Applications and Interchangeable Probes	
		2. sector array/cardiac transducer probe for use on small pedia/neonate with frequency range between 4-12 MHz	
		3. sector array transesophageal transducer for pediatric/neonate with weight as low as 2.5 kgs with frequency range between 3-8 MHz	
		• compact pinless connectors	
		• interchangeable with existing transducer probes	
		9. Other Features:	
		• user friendly because of enhanced mobility	
		• multi-port connects to four transducers	
		• multi-cup holder supports six cups for both probes and gel bottles	
		• with integrated storage trays/shelves and drawer	
		• with integrated footrests	
		• with four sturdy wheels and swivel/brake lock control	
		• with footswitch providing hands-free freeze, acquire and print	
		• with barcode scanner allowing quick and easy entry of patient data	
		10. Data Storage and Transfer:	
		• medical grade digital video recorder	
		• DICOM networking configurable and compatible with connectivity to existing PACS system	
		• at least 500GB internal hard drive	
		• with in-built anti-virus and malware protection	
		• comes with the latest available tools for maintaining patient information security with data encryption option	
		• with 5 USB ports for importing and exporting protocols and exams, archiving and serviceability	
		• export PC format image export to removable media	
		• export PDF report to removable media	
		• at least 180 hours of HD video can be stored	
		• directly connect to ultrasound thru LAN	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Cardiac Ultrasound with Advanced Applications and Interchangeable Probes	
		• HD video using MPEG-4 or AVC-HD	
		• simultaneous record to hard drive, or to USB stick	
		11. Report:	
		• generated printout after end of exam	
		• user-configurable with templates per clinical exam	
		12. Power:	
		• operated at 100-240V, 3-pin power plug/cable, strictly ast 50/60 Hz	
		• no separate volt conversion	
		• power consumption must be clearly stipulated	
		13. Durability:	
		• medical grade unit and wheels that can withstand frequent transport to different areas of the hospital	
		• mechanical parts should be of heavy-duty type	
		• equipment should be maintainable	
		14. Conforms to the attached PCMC Engineering Clearance Requirements	
		<i>Other Specifications: (please indicate if applicable)</i>	
		BRAND:	
		MAKE / MODEL:	
		<u>B. Accessories (to be provided per unit)</u>	
		1 • two (2) 3-leads ECG cables:1 adult and 1 pediatric	
		2 • one (1) UPS 2.0 KVA - can provide at least 30 mins back up time for ultrasound machine	
		3 • one (1) internal B&W printer: internally mounted and remote controlled, medical grade, high resolution black and white thermal printer for image documentation, uses A7 size	
		4 • one (1) medical grade recorder with HDMI AND USB input and output	
		5 • one (1) TEE disinfection basin	
		6 • three (3) stainless steel trolley for UPS with wheels and brakes	
		7 • two (2) probe holders/sleeves	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Cardiac Ultrasound with Advanced Applications and Interchangeable Probes	
		8 • one (1) external hard drive 1TB	
		<i>C. Consumables (to be provided per unit)</i>	
		1 • five (5) packs disposable ECG electrodes	
		2 • ten (10) packs Disinfectant Ultrasound Wipes (50 wipes/pack)	
		3 • one (1) gallon ultrasound gel	
		4 • three (3) bottles of probe cleaner or equivalent	
		5 • one (1) box (10 pcs) of thermal paper for printer	
		6 • four (4) probe holders/sleeves	
		7 • ten (10) mouth guards	
		D. Warranty	
		1 Three (3) years international warranty on parts, services and accessories and 3 years on labor	
		2 Quarterly preventive maintenance and system optimization during the warranty period	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME

TELEPHONE / FAX NO.

PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City

TECHNICAL SPECIFICATIONS

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	
		A. Specifications	
		1. Main Equipment:	
		<ul style="list-style-type: none"> • color Doppler portable cardiac ultrasound machine with high-end model and system designed to make portable exams easy and efficient occupying less space in a confined area 	
		<ul style="list-style-type: none"> • continuous, real-time adjustment of system gain and time gain compensation achieving balanced brightness of tissues providing solid image quality 	
		<ul style="list-style-type: none"> • all-digital compact broadband beamformer 	
		<ul style="list-style-type: none"> • has ultrafine 2D focusing with advanced tuning and signal processing 	
		<ul style="list-style-type: none"> • has image processing for noise and artifact reduction improving tissue and border definition 	
		<ul style="list-style-type: none"> • fully configurable, allowing end-user to select imaging capabilities, transducers, and clinical analysis to support exam needs with option to add supported capabilities at any time as well as upgrades when they become available 	
		<ul style="list-style-type: none"> • carrying weight is <7.5 kgs 	
		<ul style="list-style-type: none"> • with ergonomic carrying handle 	
		2. Patient Compatibility:	
		<ul style="list-style-type: none"> • suitable for bedside and intra-procedural echocardiography of adults, adolescents/young adults, children, neonates and fetus 	
		3. Features:	
		<ul style="list-style-type: none"> • automatic scan optimization and tissue specific imaging presets 	
		<ul style="list-style-type: none"> • one-touch image optimization 	
		<ul style="list-style-type: none"> • low noise, wide dynamic range digital 	
		<ul style="list-style-type: none"> • ECG 3 leads display 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	
		<ul style="list-style-type: none"> • advanced adaptive image processing 	
		<ul style="list-style-type: none"> • continuously variable steering in 2D, color Doppler and Doppler modes 	
		<ul style="list-style-type: none"> • one-button controls 	
		<ul style="list-style-type: none"> • post-processing for recalled images in 2D (gain), M- mode reconstruction (2D), Color (gain), Doppler (baseline/sweep speed, scale) 	
		<ul style="list-style-type: none"> • support for radiology, vascular, and cardiology interventional procedures 	
		<ul style="list-style-type: none"> • upgradeable with new modalities and analysis in the future 	
		<ul style="list-style-type: none"> • upgradeable with additional interchangeable probes 	
		<ul style="list-style-type: none"> • comprehensive measurement tools with dedicated pediatric and adult congenital cardiac analysis; fetal echo analysis; vascular analysis 	
		4. Operating / Imaging Modes:	
		<ul style="list-style-type: none"> • Real-time/2D imaging 	
		<ul style="list-style-type: none"> • M-mode: anatomical and color 	
		<ul style="list-style-type: none"> • Chroma imaging with multiple color maps 	
		<ul style="list-style-type: none"> • Dual mode imaging 	
		<ul style="list-style-type: none"> • Color compare mode 	
		<ul style="list-style-type: none"> • Invert and color invert 	
		<ul style="list-style-type: none"> • Duplex for simultaneous 2D and Doppler 	
		<ul style="list-style-type: none"> • Pulsed wave (PW) and high-PRF PW Doppler 	
		<ul style="list-style-type: none"> • Continuous wave (CW) Doppler and high-PRF CW Doppler 	
		<ul style="list-style-type: none"> • M-Mode and Doppler review 	
		<ul style="list-style-type: none"> • Tissue Doppler Imaging (TDI), TDI M-Mode 	
		<ul style="list-style-type: none"> • Pulse inversion 	
		<ul style="list-style-type: none"> • Cineloop acquisition (time and beats) with long capture 	
		<ul style="list-style-type: none"> • Cineloop review with trackball control of image selection 	
		<ul style="list-style-type: none"> • Color power angiography 	
		<ul style="list-style-type: none"> • Tissue harmonics 	
		<ul style="list-style-type: none"> • 2D Transesophageal Echocardiography (TEE) 	
		<ul style="list-style-type: none"> • 3D Transthoracic Echocardiography (TTE) capability 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	
		• Global longitudinal strain capability	
		• 3D Transesophageal Echocardiography (TEE) capability	
		5. Applications / Software Preset	
		• Pediatric Cardiology clinical option	
		• Neonatal Cardiology clinical option	
		• Fetal Cardiology clinical option	
		• Adult Cardiology clinical option	
		• Stress Echocardiography clinical option	
		• Vascular Cardiology clinical option	
		• Transcranial Doppler option	
		• Needle visualization	
		• Live plane imaging	
		6. Display Screen and Controls:	
		• at least 15 inches high resolution display with wide viewing angle and automatic ambient light compensation	
		• laptop style with alphanumeric QWERTY keyboard	
		• with at least eight (8) time gain compensation controls and two (2) lateral gain compensation controls	
		• common user experience control panel with central track ball and easy-access mode keys	
		• with quick keys and active mode	
		8. Transducers:	
		• at least two (2) with the following frequencies	
		1. matrix array transesophageal probe measuring 1cm diameter x 1m length and mechanically rotatable from 0-180° for use on adult/pedia with weight of >30kgs and compatible with 2D and 3D imaging with frequency range between 2-7 MHz	
		2. sector array/cardiac transducer probe for use on small pedia/neonate with frequency range between 4-12 MHz	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	
		<ul style="list-style-type: none"> • with available neonatal sector array transesophageal probe measuring <0.55x90cm and mechanically rotatable from 0-180° for use on neonates with weight at least 2.5kgs with frequency range between 3-8 MHz 	
		<ul style="list-style-type: none"> • interchangeable with existing transducer probes 	
		9. Other Features:	
		<ul style="list-style-type: none"> • user friendly because of enhanced mobility 	
		<ul style="list-style-type: none"> • highly mobile cart with hardware module that supports switching among up to three imaging transducers 	
		<ul style="list-style-type: none"> • multi-port connects to three (3) transducers 	
		<ul style="list-style-type: none"> • multi-cup holder supports four (4) holders for both probes and gel bottles 	
		<ul style="list-style-type: none"> • with integrated transducer connector holder, gel holders and cable management 	
		<ul style="list-style-type: none"> • with integrated storage tray/shelf 	
		<ul style="list-style-type: none"> • with integrated footrests 	
		<ul style="list-style-type: none"> • with four (4) sturdy wheels with four (4) locking casters 	
		<ul style="list-style-type: none"> • with rear handle and micro-positioning grips 	
		10. Data Storage and Transfer:	
		<ul style="list-style-type: none"> • with external DVD RW drive 	
		<ul style="list-style-type: none"> • capability for storage of images or export to DICOM, JPEG and AVI for PC compatibility 	
		<ul style="list-style-type: none"> • medical grade digital video recorder 	
		<ul style="list-style-type: none"> • all-digital compact broadband beamformer 	
		<ul style="list-style-type: none"> • DICOM networking configurable and compatible with connectivity to existing PACS system 	
		<ul style="list-style-type: none"> • at least 500GB internal hard drive 	
		<ul style="list-style-type: none"> • with two (2) USB ports for additional connectivity, importing and exporting protocols and exams, archiving and serviceability 	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	
		<ul style="list-style-type: none"> • export PC format image export to removable media 	
		<ul style="list-style-type: none"> • export PDF report to removable media 	
		<ul style="list-style-type: none"> • directly connect to ultrasound thru LAN 	
		<ul style="list-style-type: none"> • HD video using MPEG-4 or AVC-HD 	
		<ul style="list-style-type: none"> • simultaneous record to hard drive or to USB stick 	
		11. Power:	
		<ul style="list-style-type: none"> • includes AC adapter, power cord and system battery pack 	
		<ul style="list-style-type: none"> • operated at 100-240V, 3-pin power plug/cable, strictly ast 60 Hz 	
		<ul style="list-style-type: none"> • no separate volt conversion 	
		<ul style="list-style-type: none"> • power consumption must be clearly stipulated 	
		12. Durability:	
		<ul style="list-style-type: none"> • medical grade unit and wheels that can withstand frequent transport to different areas of the hospital 	
		<ul style="list-style-type: none"> • mechanical parts should be of heavy-duty type 	
		<ul style="list-style-type: none"> • equipment should be maintainable 	
		13. Conforms to the attached PCMC Engineering Clearance Requirements	
		<i>Other Specifications: (please indicate if applicable)</i>	
		BRAND:	
		MAKE / MODEL:	
		<i>B. Accessories (to be provided per unit)</i>	
	1	One (1) pc travel case: padded wheeled travel case with front zipper pockets and telescoping handle; includes a retractable stacking extension for transporting additional supplies; a customized transducer and gel carrying bag for up to four (4) transducers	
	2	Two (2) 3-leads ECG cables:1 adult and 1 pediatric	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	High-End Portable TEE Cardiac Ultrasound with Interchangeable Probes	
		3 One (1) UPS 2.0 KVA - can provide at least 30 mins back up time for ultrasound machine	
		4 One (1) external B&W printer: medical grade, high resolution black and white thermal printer for image documentation, uses A7 size	
		5 One (1) medical grade recorder with HDMI AND USB input and output	
		6 One (1) gel warmer 220-230V power requirement with at least 2 containers	
		7 One (1) TEE disinfection basin	
		8 One (one) stainless steel trolley for UPS with wheels and brakes	
		9 Two (2) probe holders/sleeves	
		10 One (1) external hard drive 1TB	
		<u>C. Consumables (to be provided per unit)</u>	
		1 Three (3) packs disposable ECG electrodes	
		2 ten (10) packs Disinfectant Ultrasound Wipes (50 wipes/pack)	
		3 One (1) gallon ultrasound gel	
		4 One (1) bottle of probe cleaner or equivalent	
		5 One (1) box (10 pcs) of thermal paper for printer	
		6 Ten (10) pieces mouth guards	
		<u>D. Warranty</u>	
		1 Three (3) years international warranty on parts, services and accessories and 3 years on labor	
		2 Quarterly preventive maintenance and system optimization during the warranty period	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME

TELEPHONE / FAX NO.

PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City

TECHNICAL SPECIFICATIONS

Instruction: Accomplish this form by presenting a clear statement of your offer. It should be specific, precise and complete which demonstrates how your statement complies with the required specifications. Please do not write "comply" of the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	Portable ECG Machine - 12 lead with Review Station, Wireless	
		A. Specifications	
		1. Main Equipment:	
		• Simultaneous recording and printing of 12 leads	
		• With real time LCD display of ECG tracing	
		• Keyboard for patient data input	
		• USB port for export of ECG tracings to portable devices/computer	
		• With rechargeable battery	
		2. Review Station / Accessories	
		• Branded Desktop	
		• at least 5th Gen Intel i5 Processors 4GB DDR3RAM	
		• 1TB HDD	
		• Media Card Reader	
		• Windows 7 Pro/Windows 10 32/65 bit 8x, DVD-RW ODD Drive	
		• At least 23" LED Monitor	
		• SFF (Small Form Factor)	
		• Mouse/Keyboard - Branded or same brand of desktop	
		• 800va UPS	
		3 With storage memory of at least 100 ECGs.	
		4 With User Manual in English	
		5 ECG tracings can be saved in PDF format.	
		6 Preferably with adult and pediatric ECG analysis	
		16 The unit will be operated at 230 Volts 3-Pin Power Plug / cable, strictly at 60 hertz	
		17 No separate volt conversion	
		18 Power consumption of the unit (<i>must be clearly stipulated</i>)	
		19 Mechanical parts should be of heavy-duty type	
		20 Equipment should be maintainable	

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	Portable ECG Machine - 12 lead with Review Station, Wireless	
		<i>Other Specifications: (please indicate if applicable)</i>	
		BRAND:	
		MAKE / MODEL:	
		<u>B. Accessories (to be provided per unit)</u>	
		1 One (1) Unit 1KVA UPS with frequency converter or its equivalent	
		2 With Adult and Pediatric chest bulbs / electrodes - 2 sets of each (pieces/set)	
		3 Table for computer	
		4 Branded (HP) Heavy duty laser jet printer (to include starting cartridge ink)	
		5 Stainless ECG Machine Trolley, with four castor wheels, with one drawer and one shelf underneath. Fitted with an IV hook rod	
		<u>C. Consumables (to be provided per unit)</u>	
		1 With at least 3 rolls/pads ECG thermal paper	
		<u>D. Warranty</u>	
		• Warranty of at least 3 years on parts and 3 years on services and labor	
		• Quarterly preventive maintenance and calibration with certificate and sticker.	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

NAME OF COMPANY	ADDRESS
-----------------	---------

SIGNATURE OVER PRINTED NAME	TELEPHONE / FAX NO.
-----------------------------	---------------------