











MINUTES OF PRE-BID CONFERENCE

Name of Project / ABC : **Supply and Delivery of One (1) unit Radiant Warmer (Re-Bid)**
Total ABC = Php 1,600,000.00

IB No. : **IB-2020-112**

Date / Time : **October 9, 2020 / 2:00 P.M.**

Venue : **Procurement Division's Waiting Area (Video Conference)**

Present were:		
Name	Designation	Signature
1. Maria Rosario S. Cruz, MD	BAC Chairperson	
2. Antonio E. Domasian, Jr., RMT	BAC Vice-Chairperson	
3. Emma A. Mariano, CPA, MGM-ESP	BAC Member	
4. Maria Cristina Estrelin c. Natonton, RN	BAC Member	
5. James Anthony F. Paredes, RN	BAC Member	
6. Ellen D. Teves, MHSA	BAC Member	
7. Ester L. Reinoso, MHSA	TWG Chairperson	
8. Engr. Neil Baribe	TWG Member	
9. Aileen M. Saremo	BAC Secretariat	
10. Jen Leslie Teraytay	BAC Secretariat	

PROSPECTIVE BIDDERS:

- | | | |
|--------------------------------------|---|--------------------|
| 1. Healthsolutions Enterprises, Inc. | - | Gilbert Anonuevo |
| 2. Bluesky Trading | - | Easter Joy Bucalon |
| 3. Dynamed Healthcare, Inc. | - | Franz Blaza |
| 4. Undecim Medical Supplies | - | Joyce Ann Sanosa |

I. Business Matters:

- The pre-bid was called to order at 2:28 P.M. It was presided by the BAC Chairperson, Maria Rosario S. Cruz, MD and discussed by BAC Member, Ms. Maria Cristina Estrelin C. Natonton, RN;

2. Reading of the details of the issued Invitation to Bid No. IB-2020-112 for the bidding of the above mentioned project; and
3. Presentation and clear explanation on the requirements in the Bid Data Sheet, Special Conditions of the Contract, Schedule of Requirements and the bidder's queries about the requirements, specifications and other conditions.

II. Noted comment/reaction from the representatives of the BAC, TWG and bidders.

PARTICULAR / PCMC Requirement	Issues Raised by Prospective Bidders	BAC/End-user's Remark/Comments/ Reactions/Explanation
SECTION VII. TECHNICAL SPECIFICATIONS		
IB-2020-112 : Radiant Warmer		
Bed tilt capable (30° - 60°)	It's too inclined, the average of tilting is 10° to 12° only, and the baby might slide on that range.	Noted, the BAC will refer this matter to the end-user. For Bid Bulletin
With APGAR score timer	Can it be with APGAR timer only? The timer is not particular on APGAR and the score is from the doctor	Noted
Built-in T-piece resuscitator (O2 source : one pipe-in hose for oxygen and compressed air; and two regulators for oxygen and compressed air)	We do not have a built-in T-piece resuscitator in our warmer, can we offer separate T-piece Resuscitator from the unit to be placed on the tray?	The BAC will refer this matter to the end-user. For bid bulletin
Built-in suction machine with liner	We offer reusable receptacle/jar only	To verify with the end-user, for bid bulletin
Built-in weighing scale (can weigh ≥ 350g)	Can I give a weighing scale that is not integrated to the unit? You can choose either built-in weighing scale or with built-in cassette tray	The BAC think that the end-user wants to have a built-in weighing scale to avoid frequent transfer of the patient
With built-in battery back-up	We do not have a back-up battery, but with UPS.	As long as it has UPS
Heating Element	Is this additional from the one included in the package? Is the heating element a thermal blanket or a tube that provides heat?	Yes To clarify with the end-user
Section VI. Schedule of Requirements		
Delivery Period: Forty-Five to Sixty (45 to 60) calendar days	Can it be sixty to ninety (60 to 90) calendar days?	Okay

III. Other Matters:

Emphasis/clarification on the following:


1. As indicated in the Bid Data Sheet ITB Clause 10, request for clarification(s) on any part of the Bidding Documents or for an interpretation must be in writing and submitted to the BAC of the Procuring Entity concerned at least ten (10) calendar days before the deadline set for the submission and receipt of bids.

There having no other remaining topics, the pre-bid conference was adjourned at 3:01 P.M.

Prepared by:

 10-12-23
Aileen M. Saremo
J.O., BAC Secretariat

Checked and Reviewed by:


ZENAIDA V. TALAGTAG, MGM-ESP
Head, BAC Secretariat 