



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100

website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph

Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

NOTICE OF AWARD
NOA-2019-129

VARIANCE TRADING CORPORATION

AGGG-CTII Bldg. Cor. Matapat St.,

Brgy. Pinyahan, District 4,

Quezon City

Tel. No.: 952-3075 / Fax No.: 952-9233

Dear Sir / Madam:

This is to inform you that based on the result of the conducted Public Bidding for the project stated below under Invitation to Bid No. IB-2019-103 as per BAC Resolution No. R2019-09-317, your proposal was found to be responsive.

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST (Php)
4	unit	Supply and Delivery of Suction Machine Brand/Model: Bicakilar Turkey/ Vaculine Maxi (see Annex "A" for detailed specifications)	Php998,000.00

You are hereby required to provide on or before 22 SEP 2019 the Performance Security in either of the following form:

FORM OF PERFORMANCE SECURITY	AMOUNT OF PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price)
a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank	Five percent (5%) Php49,900.00
b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%) Php299,400.00



**NOTICE OF AWARD
NOA-2019-129
VARIANCE TRADING CORPORATION**

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

Very truly yours,


JULIUS A. LEZCIONES, MD, PhD, DPA, CESO III
Executive Director

Conforme:

This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto.

Authorized Signatory (Signature over printed name)

Designation

Date

VARIANCE TRADING CORPORATION

QTY	UNIT	ITEM DESCRIPTION
4	unit	Supply and Delivery of Suction Machine Brand/Model: Bicakcilar Turkey/ Vaculine
		Total Cost = Php 998,000.00 (Php 249,500.00 per unit)
		A. Specifications:
		1 With 2 pcs of 3L canister
		2 With canister system suitable for use
		3 0-680/710 mmHg vacuum range (at sea level)
		4 With Vacuum diversion knob
		5 Output: 50L/min
		6 Function of One suction bag remains while other is being discarded
		7 Ergonomic control panel design
		8 Hydrophobic bacterial air filter and clear view tight seal lid
		9 With basket in the rear for spare suction catheters
		10 Polycarbonate canister capable with steam sterilization
		11 With operation status indication lamp
		12 With Silent, dry vacuum pump
		13 Motor speed of 1400 rpm
		14 With a sturdy body construct with durable polyster dye
		15 With 4 swiveling castor and 2 with brakes on the wheels
		16 With a dimensions of 485mm x 650mm x 985 mm
		17 Weight of 35kg
		18 Operational on 230 volts and 60 Hz
		19 No separate conversion of voltage
		20 Power consumption: 0.9A, 145 W
		21 Heavy duty type on the mechanical part
		22 Equipment is maintainable
		B. Accessories per unit
		1 AVR, compatible to the equipment power requirement and with proper grounding when needed
		2 Suction bottle, one (1) each for 3L
		3 Five (5) pieces Adaptor
		4 Will provide rinse solution receptacle and two (2) separate suction line holders
		C. Consumables per unit
		1 Twenty (20) pieces Bacterial filter
		2 Suction liner, two (2) each
		D. Other Specifications
		➤ Metal support for the canister made of stainless steel against weight

QTY	UNIT	ITEM DESCRIPTION
		➤ Metal support for the canister made of stainless steel against horizontal movements or fixed installation
		➤ Rubber paddings against scratches of the canister
		E. Warranty
		➤ With warranty certificate for at least two (2) years warranty for parts and labor including accessories
		➤ With quarterly preventive maintenance during the warranty period.
		F. With duly notarized certificate from Distributor - Variance Trading Corporation
		<i>Certifies that:</i>
		i. Parts and accessories and consumables are readily available at the authorized Philippine service center/s for a period (5) years after the warranty period.
		ii. With available competent in-house technical specialists in handling and providing technical support as well as maintenance of the equipment.
		iii. Will conduct on-site training for end-users and maintenance personnel on basic preventive maintenance and proper operation of the unit.
		iv. It will provide replacement/back-up unit while the delivered unit is being repaired.
		G. With duly notarized certificate from Principal Manufacturer - Bicakcilar Tibbi Cihazlar Sanayi ve Ticaret A.S.
		<i>Certifies that:</i>
		i. That the Principal Manufacturer shall honor all the terms and conditions stated in the bidding documents per IRR of RA 9184 and corresponding contract for the project , including in the event that a change of dealership will occur during the duration of the contract up to the warranty and preventive maintenance period
		ii. The expected useful life of the equipment is at least five (5) years under normal use which is 10 patients and 8 operating hours;
		iii. Guarantee on availability of all spare parts, accessories and consumables is at least five (5) years from testing, commissioning, acceptance and delivery;
		iv. That it has competence in handling and providing technical support as well as maintenance of the equipment offered; and
		v. In the case that the equipment is due for disposal, International Standards and Guidelines should be followed. This type of medical equipment has no hazardous waste so it can be thru hospital garbage disposal.
		H. DELIVERY PERIOD: Sixty to Ninety (60 to 90) calendar days

Conforme:

Authorized Signatory (Signature over printed name)

Designation/Position

Date _____

Annex A (NOA-2019-129)
Suction Machine