



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

NOTICE OF AWARD
NOA-2019-159

ZUELLIG PHARMA CORP.
JVA INTERPHIL LABORATORIES INC.
KM 14 West Service Road SSH Corner Edison Avenue,
Brgy. Sun Valley, Parañaque City
Tel. No. (02) 908-2222
Fax No. (02) 325-0641

Dear Sir / Madam:

This is to inform you that based on the result of the conducted Public Bidding for the project stated below under Invitation to Bid No. IB-2019-153 as per BAC Resolution No. R2019-12-447, your proposal was found to be responsive.

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST (Php)
1	Set	HARMONIC (ULTRASONIC) AND ENSEAL (ADVANCE BIPOLAR) TECHNOLOGY IN ONE GENERATOR Brand: ETHICON JOHNSON & JOHNSON, GEN 11 <i>(see Annex "A" for detailed specifications)</i>	Php 2,350,000.00

You are hereby required to provide on or before 31 DEC 2019 the Performance Security in either of the following form:



FORM OF PERFORMANCE SECURITY	AMOUNT OF PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price)
a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank	Five percent (5%) Php117,500.00
b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%) Php705,000.00



NOTICE OF AWARD
NOA-2019-159
ZUELLIG PHARMA CORP. JVA INTERPHIL LABORATORIES INC.

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

Very truly yours,


JULIUS A. LECCIONES, M.D., PhD, DPA, CESO III
Executive Director 

Conforme:

This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto.

Authorized Signatory (Signature over printed name)

Designation

Date

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST
1	set	HARMONIC (ULTRASONIC) AND ENSEAL (ADVANCE BIPOLAR) TECHNOLOGY IN ONE GENERATOR Brand: ETHICON JOHNSON & JOHNSON, GEN 11	Php2,350,000.00
		A. Specifications:	
		Precisely directed mechanical energy: a blade vibrating at 55,500 Hz creates cellular friction, causing amino acids in protein to unwind the and hydrogen bonds to break down. The amino acid morp into a new protein shape, forming a sticky cogulum tha seals vessels up to and including 5 mm as well as lymphatics	
		Hemostasis is achieved with minimal lateral thermal damage and minimal smoke (mist or vapor) is produced. Minimal thermal spread: precise cutting with minimal thermal damage allows for safer dissection near vital structures, compared with electrosurgery laser	
		Ethicon offers the most complete advanced energy portfolio for open and laparoscopic procedures . Seals vessels up to and including 5mm as well as lymphatics.	
		Positive Temperature Coefficient (PTC) provides optimal temperature control. <ul style="list-style-type: none"> • A polymer compound within jaws uses PTC Technology to regulate energy flow. • Maintains consistent tissue temperature of 100 °C to minimize thermal spread from heat production. Designed to minimize sticking	
		Efficiency improved with no need for stepped sealing and diving; one step to both seal divides vs. four steps vs. other tissue sealers	
		Degree of protection againts electric shock; type CF applied part	
		Class of protection againts electric shock: Class 1	
		Ingress protection G11 enclosure : IP21	
		Ingress protection G11 enclosure : IP68	
		Output: 230 VAC, 60 Hz	
		Ambient Operating Conditions: Temperatures: 59°F - 80°F, 15°C - 27°C	
		Transport and storage Conditions: Temperatures: -31°F - 192.2°F - 80°F, 35°C ±54°C	
		Humidity 10% - 95 % non condensing	
		Weight Generator : 6 kg or below	
		Footswitch : 4 kg or below	
		The unit is operated at 230 volts, strictly at 60 hertz	

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST
1	set	HARMONIC (ULTRASONIC) AND ENSEAL (ADVANCE BIPOLAR) TECHNOLOGY IN ONE GENERATOR Brand: ETHICON JOHNSON & JOHNSON, GEN 11	Php2,350,000.00
		No separate voltage conversion	
		Power consumption must be clearly stipulated	
		Mechanical parts is of heavy-duty type	
		Equipment is main tainable	
		<i>B. Additional Accessories per unit</i>	
		<ul style="list-style-type: none"> • One (1) Unit AVR 	
		<i>C. Warranty:</i>	
		<ul style="list-style-type: none"> • Two (2) years warranty on parts, service, and accessories • Quarterly Preventive Maintenance and Calibration during the warranty period. 	
		<i>E. With duly notarized certificate from Distributor - ZUELLIG PHARMA CORPORATION</i>	
		<i>Certifies that:</i>	
		i. Parts and accessories and consumables are readily available at the authorized Philippine service center/s for a period of five (5) years after the warranty period.	
		ii. Has available competent in-house technical specialists in handling and providing technical support as well as maintenance of the equipment.	
		iii. Will conduct training for proper orientation and maintenance to end-users of the equipment upon delivery.	
		iv. Will provide replacement/back-up unit while the delivered unit is being repaired.	
		<i>F. With duly notarized certificate from Principal Manufacturer - JOHNSON & JOHNSON</i>	
		<i>Certifies that:</i>	
		i. The Principal Manufacturer shall honor the terms and conditions agreed upon by its distributor, Zuellig Pharma Corporation in the event that a change of dealership occurs during the duration of the contract up to the warranty and preventive maintenance period;	
		ii. The expected useful life of ETHICON GEN11 Generator is at least 10 years depending on the usage of care	

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST
1	set	HARMONIC (ULTRASONIC) AND ENSEAL (ADVANCE BIPOLAR) TECHNOLOGY IN ONE GENERATOR Brand: ETHICON JOHNSON & JOHNSON, GEN 11	Php2,350,000.00
		iii. Guarantee on availability of all spare parts, accessories and consumables at least for the next five years from testing, commissioning, acceptance and delivery of the machine.	
		iv. That it has the competence to handle and provide Technical Support as well as maintenance of the equipment being offered; and	
		v. The Generator G11 and accessories must not be disposed of at the end of life with other waste.	
		To recycle waste equipment, obtain instructions from Ethicon Customer Service Department (call 1-877-ETHICON, +-513-337-8901 (English) or contact the local sales representative to discuss local waste disposal solutions and processes .	
		The Generator G11 poses disposal risk similar to consumer electronics such computers. There are no radioactive substances, batteries, or hazardous liquids that may leak in the Generator G11	
		vi. That the brand has been present for at least five (5) years in the local/ international market, and that the GEN11 Harmonic & Enseal Advanced Energy Generator is currently used in the international and Philippine Market with at least 2 installations in Metro Manila.	
		G. DELIVERY PERIOD: Forty Five (45) - Sixty (60) calendar days	

Conforme:

Authorized Signatory (Signature over printed name)

Designation/Position

Date _____