



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunkline: 588-9900 DirectLine: 924-0836 Fax No: 924-0840

NOTICE OF AWARD
NOA-2020-105

EVERYDAY ENTERPRISE

Angus Complex Bldg.,
81-A McArthur Highway,
Brgy. Matina Crossing Talomo District,
Davao City
Email Add: everyday.enterprise@yahoo.com.ph

Dear Sir / Madam:

This is to inform you that based on the result of the conducted Public Bidding for the project stated below under Invitation to Bid No. IB-2020-032 as per BAC Resolution No. R2020-07-295, your proposal was found to be responsive.

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST (Php)
1	unit	Supply and Delivery of Operating Table Brand: TECHARTMED ET300C (see Annex "A" for detailed specifications)	Php580,000.00


You are hereby required to provide on or before 08 AUG 2020 the Performance Security in either of the following form:

FORM OF PERFORMANCE SECURITY	AMOUNT OF PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price)
a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank	Five percent (5%) Php29,000.00
b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%) Php174,000.00

**NOTICE OF AWARD
NOA-2020-105
EVERYDAY ENTERPRISE**

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

Very truly yours,


JULIUS A. LECCIONES, MD, PhD, DPA, CESO III
Executive Director

Conforme:

This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto.

Authorized Signatory (Signature over printed name)

Designation

Date

ANNEX "A"
NOA-2020-105
EVERYDAY ENTERPRISE

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST
1	unit	Supply and Delivery of Operating Table Brand: TECHARTMED ET300C	Php580,000.00
		A. Specifications:	
		1 Electro-hydraulic driven system	
		2 X-ray and C-arm compatible	
		3 4-section design to meet different requirements of surgical positions	
		4 Stable hydraulic systems with square column, high load-bearing and low	
		5 Stainless steel body structure with detachable latex foam (slow rebound)	
		6 Adjustments: Lifting, forth/back tilting, up-folding of back board and hip board	
		7 Load capacity: 200 kg	
		8 Control of position: Electro Hydraulic with a wire handset	
		9 Minimum Table Length: 2070 mm	
		10 Minimum Table Width: 550 mm	
		11 Lowest Table Level: 690 mm	
		12 Lowest Table Level: 1,040 mm	
		14 Forward/Backward Leaning Angle: $\geq 30^\circ / \geq 25^\circ$	
		15 Left/Right Leaning Angle: $\geq 15^\circ / \geq 15^\circ$	
		16 Max Height of Cone: > 110mm	
		17 Upward/Downward for Back Plate: $\geq 75^\circ / \geq 20^\circ$	
		18 Upward/Downward for Head Plate: $\geq 45^\circ / \geq 90^\circ$	
		19 Upward/Downward/Outward for Leg Plates: $\geq 15^\circ / \geq 90^\circ / \geq 90^\circ$	
		20 The unit complies with PCMC's power source at 230 volts, strictly at 60 hertz	
		21 No separate volt conversion	
		22 Power consumption: 1.0 kW	
		23 Mechanical parts made of heavy-duty type	
		24 Equipment is maintainable	
		B. Accessories Per Unit	
		1 AVR (compatible to the equipment power requirement) and proper grounding when needed	
		C. Warranty	
		➤ With at least two (2) years warranty on services, parts and accessories	
		➤ With quarterly preventive maintenance and calibration during the warranty period	
		D. With duly notarized certificate from Distributor - Everyday Enterprise	
		<i>Certifies that:</i>	
		i. Parts and accessories and consumables are readily available at the authorized Philippine service center/s for a period (5) years after the warranty period.	
		ii. With available competent in-house technical specialists in handling and providing technical support as well as maintenance of the equipment.	

ANNEX "A"
NOA-2020-105
EVERYDAY ENTERPRISE

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST
		iii. Will conduct training for proper orientation and maintenance to end-users of the equipment upon delivery.	
		iv. It will provide replacement/back-up unit while the delivered unit is being repaired during.	
		E. With duly notarized certificate from Principal Manufacturer - Ningbo Techart Medical Equipment Co., Ltd.	
		<i>Certifies that:</i>	
		i. That the Principal Manufacturer shall honor all the terms and conditions stated in the bidding documents per IRR of RA 9184 and corresponding contract for the project including in the event that a change of dealership will occur during the duration of the contract up to the warranty and preventive maintenance period;	
		ii. The expected useful life of the equipment under normal use is up to ten (10) years;	
		iii. Guarantee on availability of all spare parts, accessories and consumables for the next five (5) years from testing, commissioning, acceptance and delivery;	
		iv. That it has competence in handling and providing technical support as well as maintenance of the equipment offered; and	
		v. Consumer guidelines regarding disposal of the equipment (see attached guidelines on disposal)	
		F. DELIVERY PERIOD: Ninety (90) calendar days	

Conforme:

Authorized Signatory (Signature over printed name)

Designation/Position

Date _____

CONSUMER GUIDELINES ON DISPOSAL

To: **WHOM IT MAY CONCERN**

When **TECHARTMED** operating tables and their parts reach the end of their useful life, reusing is the last significant stage in their life cycle. Over 90% of the weight of this surgical table is steel and different metals, making it profoundly recyclable for a mechanical item. Reusing must be completed by an expert organization or office, and the surgical table or its parts must never be disposed of with mixed waste.

PRECEDING RECYLING

Expel the oils from the hydraulic system. These materials are harmful to nature and must be isolated and sent to a particular waste removal office.

DISMANTLING AND RECYCLING

The next step is to remove and dismantle systems and components from the main frame, so that they can be reused as raw materials for other products or disposed of appropriately. Some of the materials, such as plastics, can also be incinerated to generate energy.

The frames are then cut into pieces and sorted by type, after which the materials are moved on for processing, recycling or recovering the materials as appropriate, using different technologies. These processes greatly reduce the amount of waste materials sent to landfill sites.

Signed:

Name: **Hong Yi**

Title: **Director**

Duly authorized to sign this certification on behalf of:
NINGBO TECHART MEDICAL EQUIPMENT CO., LTD



Dated on 28TH day of July, 2020