



PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City  
Tel. No. 8588-9900 local 224/226

REQUEST FOR QUOTATION  
No. RFQ-2021-336

Date : \_\_\_\_\_  
Name of Supplier : \_\_\_\_\_  
Address : \_\_\_\_\_  
Telephone No. : \_\_\_\_\_

Please quote your lowest price as per specifications per item listed below thru :

Alternative Mode of Procurement - (Negotiated Procurement : Small-Value) on or before August 18, 2021  
PhilGEPS Reference No. \_\_\_\_\_ posted on August 14, 2021

Please fax your quotation at 8588-9997 or email at pcmcproc@gmail.com c/o Mr. Danny Rodriguez

QTY	UNIT	PCMC REQUIREMENT Item Description	OFFER TO COMPLY WITH THE REQUIREMENTS
2	unit	Oxygen Blender	<b>BRAND:</b>
			<b>MAKE/MODEL:</b>
		ABC per unit : 120,000.00	<b>UNIT COST :</b>
		Total ABC : 240,000.00	<b>TOTAL COST :</b>
		<b>A. Specifications</b>	
1		Provides pressurized gas at a precise air/ O2 concentration	
2		Audible supply failure alarm	
3		Includes two outlet ports with check valves to prevent glow when hoses are disconnected	
4		Oxygen concentration control adjustment range: 21-100%	
5		Overall flow range: 2-120 L/min	
6		Auxillary outflow range: 2-90 L/min	
7		Bleed flow: 10-12 L/min when activated	
8		Maximum flow for 21% and 100% control knob settings: > 120 L/min	
9		Supply pressures: 30-75 PSI air and O2	
10		Pressure drop: less than 6 PSI + IPSI	
11		Accuracy: +3% of setting, full range	
12		Alarm/ bypass activation: when inlet gas pressure differ by 20 PSI or more	
13		Alarm sound generator: reed alarm	
14		Alarm sound intensity: 80dB, 1 foot minimum	
15		Alarm/ bypass reset: When inlet gas pressures differential < 15 PSI	
16		Weight: 2¼ lbs	
17		Gas outlet connector: type "DISS" O2	
18		Gas inlet connector: type "DISS" air and O2	
		<b>B. Accessories Per Unit</b>	
1		The unit shall include internally and externally mounted filters for cleaning the air intake	
2		All user-removable filters shall be cleanable. Cleaning instructions shall be included	
3		For two or more simultaneous pediatric patients: 1 x flowmeter stand with minimum range from 0 to 2 L/min;  Kink-resistant oxygen tubing with standard connectors (15 m each)  4 x infant cannula with 2 m kink-resistant oxygen tubing with standard connectors  4 x neonatal cannula with 2 m kink-resistant oxygen tubing with standard connectors	
		<b>C. Consumables to be Provided Per Unit:</b>	
1		Rechargeable battery/ alkaline batteries	
2		Spare main power cable	
3		Replacement sets of spare fuse (if non-resettable fuses are used)	

QTY		UNIT		PCMC REQUIREMENT	OFFER TO COMPLY WITH THE REQUIREMENTS
				Item Description	
				<b>D. Warranty:</b>	
				With at least two (2) years warranty on parts and services including quarterly preventive maintenance and calibration	
				<b>D. Delivery Period</b>	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

**Documentary Requirements**

- Mayor's/Business Permit
- PhilGEPS Reg. Number
- Income / Business Tax Return (For ABC's above Php500,000.00)
- Omnibus Sworn Statement (For ABC's above Php50,000.00) [Use of the Form provided is required]

**Documentary Requirements upon delivery of the Goods**

- Original copy of the Supplier's invoice showing Goods' description, quantity, unit price, and total amount
- Original copy of the Manufacturer's and/or Supplier's warranty certificate on parts and services, including accessories starting from final acceptance
- Original copy of Certificate of Calibration, Testing and Adjustment, if applicable
- Original copy of certificate of preventive maintenance during the warranty period
- Original copy of certificate of attendance on conducted training to end-users and Engineering Personnel for proper operation and maintenance of the equipment, if applicable
- Original and two (2) copies in English Language:
  - a. User's Manual (Hard copy)
  - b. Operator's Manual (Hard copy)
  - c. Service Manual with CD Diagram
- Copy of Proof of Payment of Import Duties from Bureau of Customs / Bill of Lading, if imported

**Posting of Warranty Security:**

The obligation for the warranty shall be covered by retention money in an amount equivalent to at least five percent (5%) of the Contract Price deductible to final payment.

The said amount shall only be released after the lapse of the warranty period specified above; provided, however, that the Supplies delivered are free from patent and latent defects and all the conditions imposed under this Contract have been fully met.

HSPR-PCMC-RQF3  
170314 Rev 1

\_\_\_\_\_  
Signature over Printed Name

*\*For Certificate of Creditable Tax Withheld at Source (BIR Form N. 2307) and Certificate of Final Tax Withheld at Source (BIR Form No. 2306) please submit your latest/updated BIR Certificate of Registration together with your quote.*