



PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City  
Tel. No. 8588-9900 local 224/226

REQUEST FOR QUOTATION  
No. RFQ-2021-367

Date : \_\_\_\_\_  
Name of Supplier : \_\_\_\_\_  
Address : \_\_\_\_\_  
Telephone No. : \_\_\_\_\_

Please quote your lowest price as per specifications per item listed below thru :

Alternative Mode of Procurement - (Negotiated Procurement : Small-Value) on or before September 10, 2021  
PhilGEPS Reference No. \_\_\_\_\_ posted on September 7, 2021

Please fax your quotation at 8588-9997 or email at danilonrodriguez@gmail.com or pcmcproc@gmail.com c/o Mr. Danny Rodriguez

QTY	UNIT	PCMC REQUIREMENT Item Description	OFFER TO COMPLY WITH THE REQUIREMENTS
1	unit	<b>BiPAP Machine</b>	<b>BRAND:</b>
			<b>MAKE/MODEL:</b>
		ABC per unit : 900,000.00	<b>UNIT COST :</b>
		Total ABC : 900,000.00	<b>TOTAL COST :</b>
		<b>A. Specifications</b>	
1		Patient Compatibility	
		Pediatric patient 5 kgs and above	
		<b>MODES</b>	
		A/C-PC • Assist control (pressure control)	
		A/C-VC • Assist control (volume control)	
		CPAP • Continuous positive airway pressure	
		PSV • Pressure support ventilation	
		S/T • Spontaneous/timed ventilation	
		SIMV-PC • Synchronized intermittent mandatory ventilation (pressure control)	
		SIMV-VC • Synchronized intermittent mandatory ventilation (volume control)	
		MPV-PC • Mouthpiece ventilation (pressure control)	
		MPV-VC • Mouthpiece ventilation (volume control)	
2		Adjustable Settings	
		Tidal volume: 0-2000ml	
		Respiratory Rate: 0 - 90 breaths/min	
		Peak Flow: 0-200 L/min	
		Inspiratory pressure: 0-90 cmH20	
		I:E ratio: 9.9:1 to 1:9.9	
		PEEP/CPAP: 0-20 CmH20	
		FiO2: 21-100%	
3		Monitor	
		At least 8" screen display with high resolution screen color display	
		Touchscreen LCD/LED	
		Displays manual calibration system	
4		Alarm Display	
		Adjustable alarm limits	
		Adjustable Audiovisual alarms for available measured and monitored parameters , including circuit disconnection and/or occlusion, apnea and no trigger	
5		Accessories	
		Heater base - manual temperature controlled with universal chamber	
		Flow sensor, oxygen sensor, HEPA filter, tubing holder	

QTY	UNIT	PCMC REQUIREMENT Item Description	OFFER TO COMPLY WITH THE REQUIREMENTS
	6	Gas Source	
		Medical Oxygen and Medical Compressed Air gauges and fittings should be compatible with the existing:	
		a. Suspended type of gas source outlet	
		b. DISS (Diameter Index Safety System) concealed wall-mounted outlets	
	7	Power Requirement	
		230 VAC / 60 Hertz	
	8	AVR	
		1000 VA	
	9	Training	
		Conduct training and orientation for proper operation and maintenance inclusive of manual	
		<b>B. Consumables to be provided per unit</b>	
	1	Ventilator Set Tubing - five (5) sets each standard size for neonate, pediatric & adult tubings	
		<b>C. Warranty</b>	
	1	Availability of Service Engineer within 24 hours - 24/7	
	2	Quarterly Preventive Maintenance and calibration during the warranty period	
	3	Provides service unit when installed unit is under repair	
	4	International warranty 3 years on parts and services	
	5	Availability of parts in the market within the next 5 years	
		<b>D. Delivery Period</b>	
		30 days	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

#### Documentary Requirements

- Mayor's/Business Permit
- PhilGEPS Reg. Number
- Income / Business Tax Return (For ABC's above Php500,000.00)
- Omnibus Sworn Statement (For ABC's above Php50,000.00) [Use of the Form provided is required]

#### Documentary Requirements upon delivery of the Goods

- Original copy of the Supplier's invoice showing Goods' description, quantity, unit price, and total amount
- Original copy of the Manufacturer's and/or Supplier's warranty certificate on parts and services, including accessories starting from final acceptance
- Original copy of Certificate of Calibration, Testing and Adjustment, if applicable
- Original copy of certificate of preventive maintenance during the warranty period
- Original copy of certificate of attendance on conducted training to end-users and Engineering Personnel for proper operation and maintenance of the equipment, if applicable
- Original and two (2) copies in English Language:
  - a. User's Manual (Hard copy)
  - b. Operator's Manual (Hard copy)
  - c. Service Manual with CD Diagram
- Copy of Proof of Payment of Import Duties from Bureau of Customs / Bill of Lading, if imported

QTY	UNIT	PCMC REQUIREMENT Item Description	OFFER TO COMPLY WITH THE REQUIREMENTS
-----	------	--------------------------------------	---------------------------------------

**Posting of Warranty Security:**

The obligation for the warranty shall be covered by retention money in an amount equivalent to at least five percent (5%) of the Contract Price deductible to final payment.

The said amount shall only be released after the lapse of the warranty period specified above; provided, however, that the Supplies delivered are free from patent and latent defects and all the conditions imposed under this Contract have been fully met.

*rx 9/6/21*

HSPR-PCMC-RQF3  
170314 Rev 1

\_\_\_\_\_  
Signature over Printed Name

*\*For Certificate of Creditable Tax Withheld at Source (BIR Form N. 2307) and Certificate of Final Tax Withheld at Source (BIR Form No. 2306) please submit your latest/updated BIR Certificate of Registration together with your quote.*