



Republic of the Philippines  
DEPARTMENT OF HEALTH  
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

**BID BULLETIN NO. BB-2024-039**

Date : October 22, 2024  
TO : ALL PROSPECTIVE BIDDERS  
FROM : BIDS AND AWARDS COMMITTEE  
SUBJECT : AMENDMENT AS STATED  
Name of Project : Supply and Delivery of One (1) unit Endoscopic Video System for Pulmonary inclusive with Tower and Bronchoscope with Suction Channel (2.8mm and 3.7mm) under IB-2024-095  
ABC = 15,000,000.00


This Bid Bulletin No. **BB-2024-039** outlines the amendments on the following which shall form an integral part of the Bidding Documents for above projects as follows:

Under Section VII. Technical Specifications

FROM	TO
<b>I. SPECIFICATIONS</b>	<b>I. SPECIFICATIONS</b>
<b>2. Monitor</b>	<b>2. Monitor</b>
Dimension (WxHxD) : 630-637 x 390-396 x 78 - 80 mm	Dimension (WxHxD) : 635-660 x <u>400-510</u> x <u>150 x 320</u> mm
Weight : 7 - 8 kgs	Weight : 7 - <u>10</u> kgs
<b>5. Device</b>	<b>5. Device</b>
Size : 2.5 mm	Size : <u>2.8</u> mm

PARTICULARS	CLARIFICATION
<b>I. SPECIFICATIONS</b>	<b>I. SPECIFICATIONS</b>
<b>5. Device</b>	<b>5. Device</b>
Angulation range : Up 180°, Down 130° Right 120° Left 120°	Angulation range : Up 180°, Down 130° Right 120° Left 120°
<b>III. CONSUMABLES PER UNIT</b>	<b>III. CONSUMABLES PER UNIT</b>
100 pc Swivel	100 pc Swivel <u>adaptor</u>

For information and guidance of all concerned.

  
FRANCIS S. DELA CUESTA, RN, MAN  
Chairperson, Bids and Awards Committee

