



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
 website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
 Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

BID BULLETIN NO. BB-2024-048

Date : November 13, 2024
 TO : ALL PROSPECTIVE BIDDERS
 FROM : BIDS AND AWARDS COMMITTEE
 SUBJECT : AMENDMENT AS STATED

Name of Project/s : **IB 2024 - 109 :One (1) Lot Supply, Delivery, Installation, Testing and Commissioning of High-End X-Ray Fluoroscopy Dual Tube, Ceiling Mounted with Workstation with High-End Ultrasound Machine**
 ABC= Php 112,000,000.00

This Bid Bulletin No. **BB-2024-048** outlines the amendments in the bidding documents for the above project as follows:

	FROM	TO
	<u>A. Specifications</u>	<u>A. Specifications</u>
	<u>15. Requirements for the Main Unit per Engineering Clearance</u>	<u>15. Requirements for the Main Unit per Engineering Clearance</u>
15	Unit will operate at 230 Volts 3-Pin Power Plug Cable, strictly at 60 Hertz	The unit will be operated at least 230 volts, strictly at 60 hertz or Manufacturer's recommendation
	<u>B. Accessories</u>	
12.9	Histogram-based dose regulation – enables consistently high image quality	Remove
	5. Two (2) units Branded Workstation hardware	
5.1	Workstation : Model	Remove
5.1.2	650VA UPS with license software for workstation	with license software for workstation
5.1.4	External Video Card: At least 8GB	External Video Card: At least 4GB
5.1.5	Memory: 96 GB Memory	Memory: 32GB Memory or manufacturers specifications
5.1.7	Power Supply: at least 950W	Power Supply: At least 500 W or manufacturers recommendation with Engineering Clearance
	5.2 Reading Station	
5.2.2	650VA UPS with license software for workstation	with license software for workstation
5.2.4	External Video Card: At least 8GB	External Video Card: At least 4GB
5.2.5	Memory: 96 GB Memory	Memory: 32GB Memory or manufacturers specifications
5.2.7	Power Supply: at least 950W	Power Supply: At least 500 W or manufacturers recommendation with Engineering Clearance

BID BULLETIN NO. BB-2024-048

5.2	MODEL	Remove
	One (1) unit Disc Publisher	
9.1	with license for workstation and desktop operation (MS Office)	Remove
9.20	External Video Card: at least 8GB	Remove
9.23	Graphical Processing Unit: Manufacturers's standard with UPS	Remove
9.24	650VA UPS with license software for workstation	Remove
9.25	CPU: at least 6 cores, 1.9 GHz	Remove
9.26	External Video Card: At least 8GB	Remove
9.27	Memory: 96 GB Memory	Remove
9.28	Windows 10 and must be upgradeable	Remove
10	One (1) unit Mid Range Ultrasound unit for Radiology with Shearwave	One (1) unit High-end ultrasound Machine (Brand should be of the same brand with the Fluoroscopy Machine)
	10.28 APPLICATIONS/ SOFTWARE:	
10.28.12	Accessories: 10.28.12 The system images and volumes from other modalities like Computed Topography (CT), Magnetic Resonance Imaging (MRI), Mammography, X-Ray and Ultrasound into the ultrasound system for simultaneous viewing alongside the real time ultrasound image	10.28.12 The system images and volumes from other modalities like Computed Tomography (CT), Magnetic Resonance Imaging (MRI), and Mammography.
	<u>2. Requirements for the Main Unit per Engineering Clearance (Ultrasound)</u>	
2.3	Should provide built in TVSS	at least 60 hertz or manufacturer's recommendation with engineering clearance
22	Accessories : Customized Cabinets	Five (5) Customized Built-in Cabinets
24	Accessories CCTV with two-way voice system	24. One (1) unit system four (4) CCTV with two-way voice system
6	Room renovation including repainting and other civil electrical and air conditioning works including selected rooms at the radiology division such as reception must be based on the existing size of the area.	Remove, double entry

For information and guidance of all concerned.

(sgd)

FRANCIS S. DELA CUESTA, RN, MAN
Chairperson, Bids and Awards Committee