

Republic of the Philippines DEPARTMENT OF HEALTH PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100 website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

BID BULLETIN NO. BB-2025-014

Date

November 13, 2024

TO

FROM

ALL PROSPECTIVE BIDDERS

BIDS AND AWARDS COMMITTEE

SUBJECT

AMENDMENT/ADDENDUM/CLARIFICATION AS STATED

One (1) Lot Supply and Delivery of reagents/consumables under Reagent Tie-up

Name of Project

Agreement for three (3) years for Automated Analyzer for Urine Electrolytes and

Ionized Calcium Determination Invitation to Bid No. IB-2025-029

Total ABC = Php 480,000.00

This Bid Bulletin No. BB-2025-014 outlines the amendments and additional requirements in the bidding documents for above project as follows:

I. AMENDMENTS

Under Terms of Reference (TOR)

	ITEM DESCRIPTION	
	FROM	ТО
2.1	Installation and Commissioning of the machine	2.1 Installation and Commissioning of the machine
peripherals		peripherals
	1.3 Pay all related expenses and requirements on LIS nnectivity.	DELETE

II. ADDITIONAL REQUIREMENT

Under Terms of Reference (TOR)

2.2 Supply and Delivery of Reagents and Consumables

2.2.8 The winning bidder shall provide all necessary reagents and supplies required for the equipment validation, which will be conducted by the PCMC Pathology Division upon delivery of the machine. The validation studies must include 20 tests for each of the following criteria:

- > Linearity.
- > Precision (within run and between run)
- > Accuracy (within run and between run)
- > Carryover
- > Limit of blank
- > Lower and higher limit of quantification
- > Contamination
- > Matrix of interferences



BID BULLETIN NO. BB-2025-014

- > Comparison bias
- > Reference range

For information and guidance of all concerned.

FRANCIS S. DELA CUESTA, RN, MAN

Chairperson, Bids and Awards Committee