

# Republic of the Philippines DEPARTMENT OF HEALTH PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100

website: <a href="mailto:www.pcmc.gov.ph">www.pcmc.gov.ph</a> email: <a href="mailto:officeofthedirector@pcmc.gov.ph">officeofthedirector@pcmc.gov.ph</a> Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

## BID BULLETIN NO. BB-2025-016

Date

: No

**November 13, 2024** 

TO

ALL PROSPECTIVE BIDDERS

**FROM** 

.

**BIDS AND AWARDS COMMITTEE** 

**SUBJECT** 

Name of Project

AMENDMENT/ADDENDUM/CLARIFICATION AS STATED

One (1) Lot Supply and Delivery of reagents/consumables under Reagent Tie-up

Agreement for three (3) years for Blood Gas

Invitation to Bid No. IB-2025-021 Total ABC = Php 2,889,780.80

This Bid Bulletin No. <u>BB-2025-016</u> outlines the additional requirement in the bidding documents for above project as follows:

#### ADDITIONAL REQUIREMENT

#### **Under Terms of Reference (TOR)**

### 2.2 Supply and Delivery of Reagents and Consumables

2.2.7 The winning bidder shall provide all necessary reagents and supplies required for the equipment validation, which will be conducted by the PCMC Pathology Division upon delivery of the machine. The validation studies must include 20 tests for each of the following criteria:

- > Linearity,
- > Precision (within run and between run).
- > Accuracy (within run and between run)
- > Carryover
- > Limit of blank
- > Lower and higher limit of quantification
- > Contamination
- > Matrix of interferences
- > Comparison bias
- > Reference range

For information and guidance of all concerned.

FRANCIS S. DELA CUESTA, RN, MAN

Chairperson, Bids and Awards Committee