



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100

website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph

Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

BID BULLETIN NO. BB-2025-017

Date : November 13, 2024
TO : ALL PROSPECTIVE BIDDERS
FROM : BIDS AND AWARDS COMMITTEE
SUBJECT : AMENDMENT/ADDENDUM/CLARIFICATION AS STATED
Name of Project : **One (1) Lot Supply and Delivery of reagents/consumables under Reagent Tie-up Agreement for three (3) years for Coagulation Invitation to Bid No. IB-2025-022**
Total ABC = Php 7,160,319.30

This Bid Bulletin No. **BB-2025-017** outlines the amendments and additional requirements in the bidding documents for above project as follows:

I. AMENDMENTS

A. Under Section VII. Technical Specifications

ITEM DESCRIPTION	
FROM	TO
4. With onboard refrigerated reagent storage within refrigerator temperature (<10C) to maintain the stability and longevity of the loaded reagents	4. With onboard refrigerated reagent storage of <15°C

II. ADDITIONAL REQUIREMENT


Under Terms of Reference (TOR)

2.2 Supply and Delivery of Reagents and Consumables

2.2.8 The winning bidder shall provide all necessary reagents and supplies required for the equipment validation, which will be conducted by the PCMC Pathology Division upon delivery of the machine. The validation studies must include 20 tests for each of the following criteria:

- > Linearity
- > Precision (within run and between run)
- > Accuracy (within run and between run)
- > Carryover
- > Limit of blank
- > Lower and higher limit of quantification
- > Contamination
- > Matrix of interferences
- > Comparison bias
- > Reference range

For information and guidance of all concerned.


FRANCIS S. DELA CUESTA, RN, MAN
Chairperson, Bids and Awards Committee

