

PRICE SCHEDULE

PROCURING ENTITY : PHILIPPINE CHILDREN'S MEDICAL CENTER						NAME OF BIDDER :			
INVITATION TO BID NO. : IB-2024-114									
1	2	3	4	5	6	7	8	9	10
Qty	Item	Description (Brand / Make / Model)	Country of Origin	Manufacturer	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 1 x 6)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 1 x 9)
1 Unit	VIDEO LARYNGOSCOPE								
ABC = Php 3,500,000.00									
<ul style="list-style-type: none"> Bids will be valid for one hundred twenty (120) days and it shall remain binding upon us and may be accepted at any time before the expiration of that period; 						TERMS OF PAYMENT (For discounts being offered, if there's any. Otherwise, state "NONE"):			
<ul style="list-style-type: none"> PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government. 						NAME AND SIGNATURE OF AUTHORIZED REPRESENTATIVE			

PRICE SCHEDULE

PROCURING ENTITY : PHILIPPINE CHILDREN'S MEDICAL CENTER						NAME OF BIDDER :				
INVITATION TO BID NO. : 2024-114										
1	2	3	4	5	6	7	8	9	10	11
Qty	Item	Description (Brand / Make / Model)	Country of Origin	Manufacturer	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	Sales and other taxes payable if Contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Price, per unit (col 6+7+8+9)	Total Price delivered Final Destination (col 10) x (col 1)
1 Unit	VIDEO LARYNGOSCOPE									
ABC = Php 3,500,000.00										
<ul style="list-style-type: none"> Bids will be valid for one hundred twenty (120) days and it shall remain binding upon us and may be accepted at any time before the expiration of that period; 						TERMS OF PAYMENT (For discounts being offered, if there's any. Otherwise, state "NONE"):				
<ul style="list-style-type: none"> PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government. 						NAME AND SIGNATURE OF AUTHORIZED REPRESENTATIVE				

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

SECTION VII: TECHNICAL SPECIFICATIONS
IB 2024-114

Instruction:

1. **FILL-OUT** this form by including a **CONCISE DESCRIPTION** of your offer. Please write the **SPECIFIC, PRECISE** and **COMPLETE** statement which conforms with the required specifications.
2. **DO NOT** write "**COMPLY**" **OR DO NOT** write **page numbers** of the brochure/data sheet, etc.

PCMC REQUIREMENT			BIDDER'S OFFER TO COMPLY WITH THE REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	VIDEO LARYNGOSCOPE	
		A. Specifications of the Main Equipment	
	1	Monitor (1 unit)	
		1. Interactive FULL HD touch screen that can also be used while wearing OR gloves	
		2. Picture-in-Picture and Side-by-Side modes	
		3. Choice between Click-to-Record, Auto-Record or Time Machine mode	
		4. Integrated microphone allows audio comments to be added to your videos	
		5. Increased diversity of visualization through maximum flexibility of single-use and reusable products	
		6. Timer for documenting intubation time for the purposes of training or studying	
		7. Screen resolution minimum 1920 x 1200 pixels	
		8. Screen size 8" - 10 inches	
		9. CMOS Technology	
		10. Two (2) camera input with USB and HDMI port	
		11. Optimized user interface with touchscreen	
		12. At least 16 GB SD card, can capture video and image in real time and data transfer from	
		13. SD card to USB Flash drive	
		14. Splash proof according to IP54	
		15. Shock resistant ABS plastic housing	
		16. Intelligent power management with rechargeable Li-ion batteries	
		17. With VESA 75 mounting option	

PCMC REQUIREMENT			BIDDER'S OFFER TO COMPLY WITH THE REQUIREMENTS
QTY	UNIT	Item Description	
		18. Compatible to use with video laryngoscopes, flexible intubation endoscopes, rigid intubation endoscopes, and semi-rigid intubation endoscopes	
		19. Mechanical parts should be of heavy-duty type	
		20. Equipment should be maintainable	
		BRAND:	
		MAKE/MODEL:	
		2 CMOS technology Blades (1 unit each)	
		1. Reusable Macintosh Blade Size 3	
		• For use with the monitor	
		• Image and video capture via the button on blade	
		• Option for compatibility to use with handheld monitor for future upgrade	
		2. Reusable Hyper angulated D-Blade for difficult intubation, Adult	
		• With catheter introduction size 16-18 Fr	
		• For use with the monitor	
		• Image and video capture via the button on the blade	
		• Option for compatibility to use with handheld monitor for future upgrade	
		3 Portable Monitor for CMOS Endoscopes (2 units) - handheld	
		1. Screen size 3.5" LCD monitor let perform rapid secure intubations in seconds	
		2. OTI (Open for Intubate) Display ready for immediate use after starting the monitor	
		3. Monitor movable via two rotation axes	
		4. Real time documentation of images and video sequences recorded to internal memory and retrieved via USB data cable	
		5. Always ready for use - Exchangeable and rechargeable batteries	
		6. IPX8 Completely watertight; fully immersible in High Level Disinfectant	
		7. Compatible with semi-rigid or optical stylet intubation options	
		8. 1 hour operation time, 2 hour charging time of battery	
		9. Charging Unit, for one rechargeable battery with power supply and mains adaptor for EU, UK, USA and AUS, power supply at least 100 - 240 VAC, 50/60 Hz, suitable for wipe disinfection (1 unit)	

PCMC REQUIREMENT			BIDDER'S OFFER TO COMPLY WITH THE REQUIREMENTS
QTY	UNIT	Item Description	
		10. Battery, rechargeable Li-Ion battery for power supply, optimal operating time of 60 min, wipe disinfection	
		Other Specifications: (please indicate if applicable)	
		BRAND:	
		MAKE/MODEL:	
		B. Accessories to be provided	
		1. IV Stand for the monitor (1 unit)	
		2. Five antistatic castors	
		3. Storage tray	
		4. With blade holder and cable winder	
		5. Connecting cable for the use of video laryngoscope, length at least 200cm (1 unit)	
		C. Consumables to ber Provided per unit	
		1. 10 packs of disinfecting wipes non-alcohol base	
		D. Warranty	
		1. With at least (2) years warranty on services, parts, and accessories	
		2. Quarterly Preventive Maintenance Services and Calibration during the warranty period	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

COMPANY'S DETAILS:

NAME OF COMPANY

ADDRESS

EMAIL ADDRESS **TEL. NO. / FAX NO.**

AUTHORIZED REPRESENTATIVE'S INFOS:

Signature over Printed Name **Designation/Position**

Email Address **Mobile No.**