



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

BID BULLETIN NO. BB-2024-041

Date : **October 22, 2024**
TO : **ALL PROSPECTIVE BIDDERS**
FROM : **BIDS AND AWARDS COMMITTEE**
SUBJECT : **AMENDMENT AS STATED**
IB 2024-091 : Supply, Delivery, Installation, Testing and Commissioning of
Name of Project/s : **One (1) Washer Disinfector (Turnkey)**
(ABC=Php 8,000,000.00)


1. Under Section III. Bid Data Sheet

| FROM | TO |
|--|---|
| ITB. Clause 20.2 The Lowest Calculated Bidder shall submit the following documentary requirements within a non-extendible period of five (5) calendar days from receipt of the notification that contains the following | |
| 7. Current and Valid Certificate of Manufacturer's compliance with ISO and/or other Standards/Conformance /Safety Requirements for the equipment | Current and Valid Certificate of Manufacturer's Compliance with the recognized environmental standards such as ENERGY STAR, ISO14001, or its equivalent. |

1. Under Section VII. Technical Specifications

| FROM | TO |
|--|---|
| <u>A. Specifications</u> | <u>A. Specifications</u> |
| 1. Chamber material made up of stainless steel medical grade 304L or 316L. | 1. Chamber material made up of stainless steel medical grade 304L or 316L. Chamber material capacity of 200 - 220 liters |
| 3. Door configuration: single manual drop down glass / automated vertical sliding ; at least 600 interior and 3mm exterior glass | 3. Door configuration: |
| <ul style="list-style-type: none">Automated Double door vertical sliding doorDouble pane tempered glass | <ul style="list-style-type: none">Automated Double door vertical sliding doorDouble pane tempered glass |
| 9. Cycle time is at least 30 minutes minimum and 40 minutes maximum including drying time | 9. Cycle time is at least 25 minutes minimum and 40 minutes maximum including drying time |
| 14. Compliance with the recognized environmental standards such as ENERGY STAR, ISO14001, or its equivalent. | delete |

For information and guidance of all concerned.


FRANCIS S. DELA CUESTA, RN, MAN
Chairperson, Bids and Awards Committee