



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

BID BULLETIN NO. BB-2024-042

Date : **October 22, 2024**
TO : **ALL PROSPECTIVE BIDDERS**
FROM : **BIDS AND AWARDS COMMITTEE**
SUBJECT : **AMENDMENT AS STATED**

Name of Project	IB NUMBER	NAME OF PROJECT
	IB-2024-088	Supply, Delivery, Installation, Testing & Commissioning of One (1) lot Duplex Anaesthetic Gas Scavenging System
	IB-2024-089	Supply, Delivery, Installation, Testing & Commissioning of One (1) lot Centralized Medical Air Plant (Duplex System)
	IB-2024-090	Supply, Delivery, Installation, Testing & Commissioning of One (1) lot Centralized Medical Vacuum Plant (Triplex System)

This Bid Bulletin No. **BB-2024-042** outlines the amendments on the following which shall form an integral part of the Bidding Documents for above projects as follows:

Under Section VII. Technical Specifications

QUESTIONS	CLARIFICATION
IB-2024-088 Supply, Delivery, Installation, Testing & Commissioning of One (1) lot Duplex Anaesthetic Gas Scavenging System	
The required specifications for Medical Vacuum Pump pertains to Anesthetic Gas Scavenging System (AGSS). There is difference between the two under regulating standards.	An AGSS is designed to capture and remove excess anesthetic gases at their source, thereby preventing them from entering the surrounding environment. Waste Anesthetic Gases are collected directly from the patient breathing system or anesthetic circuit, the collected gases are then transferred to a scavenging system, the scavenged gases are released outside the building into the atmosphere and finally be vented outside whereas systems employs filters or other technologies to neutralize harmful substances before disposal. On the contrary, as for best safety compliance, anesthetic gas scavenging system is not terminated to medical vacuum plant.

PhilHealth Accredited



BID BULLETIN NO. BB-2024-042

PARTICULARS	CLARIFICATION
IB-2024-089 Supply, Delivery, Installation, Testing & Commissioning of One (1) lot Centralized Medical Air Plant (Duplex System)	
Must be NFPA99 completely packaged medical-grade centralized air plant shall be duplex (two compressors), reciprocating, 100% oil-less compressor system, stack mounted, dry and stand alone system applicable for hospital only	Reciprocating Compressor are less maintenance, no oil to change and no tip seals to change.
Section III. Bid Data Sheet ITB Clause 5.3 (contracts similar to the project)	Medical Gas Piping Project or a piping distribution network is different from Medical Gas System Plant, where it includes centralized medical grade breathing gas facility.
Section III. Bid Data Sheet ITB Clause 20.2 (9. Certificate from the Principal Manufacturer duly notarized from country of origin)	The Certificate from Manufacturer duly notarized from country of origin demonstrates that the Medical Air Plant and Medical Vacuum Plant offered are of quality, safe, factory-tested and compliance to NFPA99 Standards

For information and guidance of all concerned.


FRANCIS S. DELA CUESTA, RN, MAN
 Chairperson, Bids and Awards Committee 