

PRICE SCHEDULE

(For Goods Offered from Abroad)

PROCURING ENTITY : PHILIPPINE CHILDREN'S MEDICAL CENTER						NAME OF BIDDER :				
INVITATION TO BID NO. : IB-2024-111										
1	2	3	4	5	6	7	8	9	10	
Qty	Unit	Item	Description (Brand / Make / Model)	Country of Origin	Manufacturer	Unit price CIF port of entry (specify port) or CIP named place (specify border point or place of destination)	Total CIF or CIP price per item (col. 1 x 6)	Unit Price Delivered Duty Unpaid (DDU)	Unit price Delivered Duty Paid (DDP)	Total Price delivered DDP (col 1 x 9)
1	unit	Supply and Delivery of Syringe Infusion Pump <i>Total ABC = 145,000.00</i>								
<ul style="list-style-type: none"> • Bids will be valid for one hundred twenty (120) days and it shall remain binding upon us and may be accepted at any time before the expiration of that period; 						TERMS OF PAYMENT (For discounts being offered, if there's any. Otherwise, state "NONE"):				
<ul style="list-style-type: none"> • PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government. 						NAME AND SIGNATURE OF AUTHORIZED REPRESENTATIVE				

PRICE SCHEDULE

(For Goods Offered from within the Philippines)

PROCURING ENTITY : PHILIPPINE CHILDREN'S MEDICAL CENTER							NAME OF BIDDER :				
INVITATION TO BID NO. : IB-2024-111											
1	2	3	4	5	6	7	8	9	10	11	
Qty	Unit	Item	Description (Brand / Make / Model)	Country of Origin	Manufacturer	Unit Price EXW per item	Transportation and all other costs incidental to delivery, per item	sales and other taxes payable if Contract is awarded, per item	Cost of Incidental Services, if applicable, per item	Price, per unit (col 6+7+8+ 9)	Total Price delivered Final Destination (col 10) x (col 1)
1	unit	Supply and Delivery of Syringe Infusion Pump <i>Total ABC = Php145,000.00</i>									
<ul style="list-style-type: none"> • Bids will be valid for one hundred twenty (120) days and it shall remain binding upon us and may be accepted at any time before the expiration of that period; 							TERMS OF PAYMENT (For discounts being offered, if there's any. Otherwise, state "NONE");				
<ul style="list-style-type: none"> • PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government. 							<i>NAME AND SIGNATURE OF AUTHORIZED REPRESENTATIVE</i>				

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

**TECHNICAL SPECIFICATIONS
IB-2024-111**

Instruction: Accomplish this form by presenting a clear statement of your offer. Please write the specific, precise and complete statement which complies with the required specifications. **DO NOT write "COMPLY"** or the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
1	unit	Syringe Infusion Pump	
		A. Specifications of the Main Equipment	
		1 Intuitive syringe infusion pump	
		2 Suitable for neonates, pediatrics and adults	
		3 Configurable data system integrated with the following but not limited to: delivery rate, current infusion, VTBI, total volume, syringe size, syringe brand, pressure limit, batter capacity, drugs, remaining time, alarms	
		4 Basic Features:	
		- 6 modes: rate, time, body weight, ramp up/down, sequential, micro-infusion	
		- bolus rate: automatic/manual bolus range = 0.10 - 1500ml/h (depending on syringe size)	
		- automatic calculation of the delivery rate based on different modes	
		- Infusion pause / standby: preset parameters remain stored	
		- data storage: history records at least 2000 and data can be transmitted to PC and saved as EXCEL format to print out	
		- titration function: change flow rate during infusion with increment 0.01ml/h	
		5 Standard parameters:	
		- mode	
		- flow rate = 5ml: 0.1-100ml/h, 10ml: 0.1-200ml/h, 20ml: 0.1-400ml/h, 30ml: 0.1-600ml/h, 50/60ml: 0.1-1500ml/h	
		- increment: 0.01ml(0.1-99.99ml/h), 0.1ml (100-999.9ml/h, 1ml(1000-1500ml/h)	
		- present volume (VTBI): 0.10-9999.99ml increment 0.01ml)	
		- present time: 00:00:01-99:59:59(h:m:s) adjustable	
		- accumulated volume: 0.01-9999.99ml (increment 0.01ml)	
		- KVO: 0.1-5.0ml/h adjustable, step 0.1ml/h, default is 1.0 ml/h	
		- purge: 0.10-1500ml/h (depending on syringe size)	
		6 Display Screen: at least 3.5 inches TFT color LCD with adjustable brightness	
		7 Dimensions: lightweight < 2.5kgs and stackable to slot structure	
		8 Syringe Selection: universal, not equipment exclusive and can be used with any brand of 5, 10, 20, 30, 50/60ml syringes	

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

**TECHNICAL SPECIFICATIONS
IB-2024-111**

Instruction: Accomplish this form by presenting a clear statement of your offer. Please write the specific, precise and complete statement which complies with the required specifications. **DO NOT write "COMPLY"** or the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
		9 Alarm: • visual and adjustable acoustic alarms • triggers: occlusion, battery empty, VTBI done, syringe empty, syringe disengaged, KVO finish, system malfunction, standby time expired, system abnormal, syringe near empty, reminder, no battery, battery low, VTBI near done, AC power disconnection • pre-alarm: 1-30mins before syringe is empty; 3 mins before battery is completed; 30 mins as low battery	
		10 pre-configured syringe brands	
		11 user-defined syringe brands and size after calibration • automatic recognition of syringe size	
		12 defibrillation-proof	
		13 water-proof	
		14 Mobility: convenient for transport with slot structure storage system and bedrail moun	
		15 Battery and Operation Time	
		16 Chargeable lithium ion internal battery: at least 4 hours continuous monitoring	
		17 Hospital grade equipment	
		18 Easily cleaned with standard sterilizing materials	
		19 The unit will be operated at 230 volts, 3-Pin Power Plug/Cable, strictly at 60 hertz <i>(If the unit/s is 50/60Hz, provide UPS with frequency converter when needed)</i>	
		20 Power consumption <i>(must be clearly stipulated)</i>	
		21 Mechanical parts should be of heavy-duty type	
		22 Equipment should be maintainable	
		Other Specifications: <i>(please indicate if applicable)</i>	
		BRAND	
		MAKE/MODEL	
		B. Accessories to be provided	
		1 power cord	
		2 bedrail mount	
		3 stackable storage system/slot structure	

PHILIPPINE CHILDREN'S MEDICAL CENTER
Quezon Avenue, Quezon City

**TECHNICAL SPECIFICATIONS
IB-2024-111**

Instruction: Accomplish this form by presenting a clear statement of your offer. Please write the specific, precise and complete statement which complies with the required specifications. **DO NOT write "COMPLY"** or the page numbers of the brochure/data sheet, etc.

PCMC REQUIREMENT			BIDDERS' OFFER TO COMPLY WITH THE PCMC REQUIREMENTS
QTY	UNIT	Item Description	
		4 1 IV medical pole for each unit, chrome plated steel tubing with round surface, well-balanced, stable no-tip cast aluminum base, easy rolling wheel ball with nylon casters, height range at least 50-90 inches, twist lock for easy height adjustment, 4 hooks and 5 wheels	
		C. Warranty	
		1 At least 2 years on parts, services and accessories	
		2 Quarterly preventive maintenance on warranty period	
		3 available loaner when under repair for > 1 week on warranty period	

PCMC has the right to reject any or all bids without offering any reason, waive any required formality and award the contract to any bidder whose proposals as evaluated by PCMC is the most advantageous to the government.

NAME OF COMPANY	ADDRESS
SIGNATURE OVER PRINTED NAME	TELEPHONE / FAX NO.