



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100

website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph

Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

NOTICE OF AWARD

NOA-2024-175

ROBUSTAN INC.

Unit 906 South Center Tower
2206 Market St. Madrigal Business Park
Muntinlupa City

Dear Sir / Madam:

This is to inform you that based on the result of the conducted Competitive Bidding for the project stated below under Invitation to Bid No. IB-2024-076 as per BAC Resolution No. R2024-12-707, your proposal was found to be responsive.

ITEM DESCRIPTION	TOTAL COST
Supply and Delivery of Two (2) units Heavy Duty Defibrillator <i>(see attached Annex A for the detailed offer)</i>	Php1,400,000.00 <i>(Php700,000.00/per unit)</i>

You are hereby required to provide on or before 12 JAN 2025 the Performance Security in either of the following form:

FORM OF PERFORMANCE SECURITY	AMOUNT OF PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price)
a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank	Five percent (5%) Php70,000.00
b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%) Php420,000.00



*Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC.
These unsanctioned requests are unlawful and will not be tolerated*



**NOTICE OF AWARD
NOA-2024-175
ROBUSTAN INC.**

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

Very truly yours,


SONIA B. GONZALEZ, MD, MSChSM, MPM
Executive Director 

Conforme:

This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto.

Authorized Signatory (Signature over printed name)

Designation

Date

*Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC.
These unsanctioned requests are unlawful and will not be tolerated*



QTY	UNIT	ITEM DESCRIPTION
2	Unit	Supply and Delivery of Two (2) units Heavy Duty Defibrillator
		<i>Brand: Bexen Cardio</i>
		<i>Make/Model: Reanibex 800</i>
		<i>Total Amount Awarded = Php 1,400,000.00 (Php 700,000.00/unit)</i>
		A. Specifications of the Main Equipment
		1. Screen size - 8.4" diagonal Weight: unit with no options: : 7.1 kg
		2. AC Supply: Input: 100 - 240 VAC, 50/60 Hz, 2,5 A DC Supply: 10-16 VDC, 10 A
		3. All alarm tones are generated internally by the REANIBEX 800. Infringement of alarms may be indicated by tones, voice messages and visual indications.
		4 Slide the adult paddle electrode until the paediatric paddle appears
		5 The REANIBEX 800 offers a choice between two recorders with different paper widths. With the 50 mm recorder, two waveforms may be printed simultaneously at most. Optionally, the REANIBEX 800 may be equipped with a 100 mm paper recorder that may be used to print up to 3 waveforms simultaneously
		6 12V NiMH battery with 3 Amp/h capacity
		7 Energy Selection: 1 - 2 - 3 - 4 - 5 - 6 - 7 - 8 - 9 - 10 - 15 - 20 - 30 - 50 - 70 - 100 - 125 - 150 - 175 - 200 - 250 - 300 - 360 Joules.
		8 Monitored by means of a 3, 5 and 10 lead cable, external reusable paddles and multifunction disposable electrodes
		9 Manual and Automated Biphasic Defibrillator up to 360 Joules. 5- lead patient cable; SPO2, NIBP
		10 A new and fully charged battery operating at 25 °C delivers approximately 150 shocks at 360 Joules, approximately 140 minutes of monitoring with ECG, SpO2, CO2 and non-invasive blood-pressure every 15 minutes, and approximately 150 minutes of stimulation in noninvasive pacing mode at 100 mA and 100 bpm.
		11 Input: 100 - 240 VAC, 50/60 Hz, 2,5 A Output: 15 V, 9,3 A, max 140 W and will provide UPS and AVR
		12 AC Supply: Input: 100 - 240 VAC, 50/60 Hz, 2,5 A Output: 15 V, 9,3 A, max 140 W DC Supply: 10-16 VDC, 10 A
		13 Mechanical parts is of heavy-duty type
		14 Equipment is maintainable
		B. Accessories to be provided per unit
		1 Stainless steel trolley (304) with one (1) level drawer or basket design and with polyurethane rigid caster wheels atleast 2 inches.
		2 Adult and pediatric paddles
		3 ECG cable with lead wires (clip type)
		4 Main cable at least two and half (2.5) meters in length
		C. Consumables to ber Provided per unit
		1 Thermal paper (20 rolls)
		2 Electrodes (adult and pedia) 100pcs each

QTY	UNIT	ITEM DESCRIPTION	
2	Unit	Supply and Delivery of Two (2) units Heavy Duty Defibrillator	
		3	Defibrillator gel (1 large tube)
		D. Warranty	
		1	Two (2) years warranty on parts and services
		2	Quarterly Preventive Maintenance Services and Calibration during the warranty period.
		E. With duly notarized certificate from Bidder - Robustan Inc.	
		<i>Certifies that:</i>	
		i.	Parts, accessories, and consumables are readily available at the authorized Philippine service center for a period of five (5) years after the warranty period;
		ii.	It has available competent in-house technical specialists in handling and providing technical support as well as maintenance of the equipment being offered;
		iii.	It will conduct training for proper operation and maintenance to the End-users of the equipment upon delivery; and
		iv.	It will provide replacement/back-up unit while the delivered unit is being repaired.
		F. With duly notarized certificate from Principal Manufacturer - Bexen Cardio Osatu S. Coop	
		<i>Certifies that:</i>	
		i.	All the terms and conditions stated in the bidding documents per IRR of RA 9184 and the corresponding contract for the contract for the project shall be honored by the Principal Manufacturer, including in the event that a change of dealership will occur during the duration of the contract up to the warranty and preventive maintenance period
		ii.	The defibrillators, of the brand BEXENCARDIO, models REANIBEX-100, REANIBEX-200, REANIBEX300, REANIBEX500, REANIBEX700 and REANIBEX800, after Ten (10) years from its manufacture, are out of its useful lifetime, so we cannot guarantee the correct functionality for which it was designed.
		iii.	Guarantee on the availability of all spare parts, accessories and consumables at least for the next ten (10) years from testing, commissioning, acceptance and delivery; "Robustan Inc.", #906 South Center Tower, 2206 Market St., Madrigal Business Park, Alabang, Muntinlupa City, Metro Manila, Philippines authorized to EXCLUSIVELY promote, market, sell and supply our products and spare parts within the territory of the Republic of the PHILIPPINES.
		iv.	It has competence in handling and providing technical support as well as maintenance of the equipment being offered; and
		v.	The Manufacturer, Bexen Cardio, has proper plan for the consumer guidelines regarding disposal of the equipment by giving information on recycling the product
		H. Delivery Period:	
		1.	Within Forty-five to Sixty (45 to 60) Calendar days upon receipt of Purchase Order/Notice to Proceed

Conforme:

Authorized Signatory (Signature over printed name)

Designation/Position

Date _____