



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

NOTICE OF AWARD
NOA-2024-179

OMNIBUS BIO-MEDICAL SYSTEMS, INC.

4th Floor Wilson St., Greenhills,
San Juan City
Tel. No.: 8727-1058 / 8722-4605
Contact No.: 0917-6810-3083
Email Add: saies@omnibus.ph

Dear Sir / Madam:

This is to inform you that based on the result of the conducted Public Bidding for the project stated below under Invitation to Bid No. IB-2024-096 as per BAC Resolution No. R2024-12-711, your proposal was found to be responsive.

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST (Php)
1	unit	Supply and Delivery of Three-Header Binocular Microscope with Digital Camera and Wi-Fi Capabilities Brand/Model: Olympus BX53 3-Header with SC180 Camera (see Annex "A" for detailed specifications)	Php1,399,000.00

You are hereby required to provide on or before 06 JAN 2025 the Performance Security in either of the following form:

FORM OF PERFORMANCE SECURITY	AMOUNT OF PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price)
a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank	Five percent (5%) Php69,950.00
b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%) Php419,700.00



**Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC.
These unsanctioned requests are unlawful and will not be tolerated.**



NOTICE OF AWARD
NOA-2024-179
OMNIBUS BIO-MEDICAL SYSTEMS, INC.

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

Very truly yours,


SONIA B. GONZALEZ, MD, MScHSM, MPM
Executive Director 

Conforme:

This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto.

Authorized Signatory (Signature over printed name)

Designation

Date

**Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC.
These unsanctioned requests are unlawful and will not be tolerated.**

ANNEX "A"
NOA-2024-179
OMNIBUS BIO-MEDICAL SYSTEMS, INC.

QTY	UNIT	ITEM DESCRIPTION
1	unit	Supply and Delivery of Three-Header Binocular Microscope with Digital Camera and Wi-Fi Capabilities Brand/Model: Olympus BX53 3-Header with SC180 Camera Total Cost = Php 1,399,000.00
		i. Specifications:
	1	Optical System: UIS2 optical system
	2	Illumination System: Built-in transmitted illumination system, fixed field diaphragm, high color reproducibility with long life LED power
	3	Focusing: Stage height movement (coarse movement stroke: 15mm), stroke per rotation for coarse adjustment knob 36.8mm, torque adjustment for coarse adjustment knob, fine focus knob (minimum adjustment gradations: 2.5 micrometer)
	4	Revolving nosepiece: Fixed quintuple nosepiece with inward tilt and connecting cable)
	5	Stage: Wire movement mechanical fixed stage, (WxD): 211 mm x 154 mm, Traveling range (XxY) 76 mm x 52 mm, Single and double specimen holder, sheet holder, specimen position scale, stage XY movement stopper
	6	Observation tube: Trinocular tube, type (anti-fungal), Eyepiece (anti-fungal), Tube inclination, Light path selector, Interpupillary distance adjusting range
	7	Condenser: Abbe condenser NA 1.25 with oil immersion. Built-in aperture iris diaphragm. Swing out condenser
	8	The microscope is designed to be easily upgraded for brightfield, simple polarization, fluorescence, phase contrast and dark field observation methods
	9	These high quality objectives offer excellent flatness up to FN 22 in transmitted brightfield observation (Plan Achromat-type and anti-fungal treated)
	10	Flourescence Light Source: LED reflected flourescence light source/ illuminator can be easily added as an upgrade
	11	Voltage Electric Current: AC 100-240V 50/60Hz 0.4A
	12	Software for Live View for uploading, annotating, sharing and embedding of microscopic high-resolution slides (lifetime subscription), upgradable, can interface with 45" UHD TV monitor with provision for bracket
	13	Camera type: Color CMOS Resolution: 4912 x 3684 (18 megapixel) Digital high-performance microscope camera for real time scanning with 4K UHD Resolution, C-mount camera attachment
	14	Upgradable to additional headers
	15	With flourescence microscope capability
	16	Antivirus software: one time purchase (lifetime)

ANNEX "A"
NOA-2024-179
OMNIBUS BIO-MEDICAL SYSTEMS, INC.

QTY	UNIT	ITEM DESCRIPTION
		17 Microsoft Office: Professional Plus 2024 License
		18 The unit will be operated at 230 volts, 3-pin power plug/cable, strictly at 60 hertz
		19 No separate volt conversion
		20 Power consumption: 24 W
		21 All mechanical components are built to withstand years of heavy use (heavy duty and durable type)
		22 Equipment is designed for easy maintenance and service
		II. Accessories PER UNIT
	1	One (1) unit 1000VA UPS
		Brand: K-Star
		Model: UA100
		Note: Will provide safety electrical devices compatible to the equipment power requirement) and proper grounding when needed
	2	Three (3) Microscope dust fabric cover (480x478mm)
	3	Three (3) Microscope eye piece dust cover
	4	Personal Computer set:
		Latest operating system software: Windows 11
		CPU, 8GB RAM and 512 GB SSD storage
		27" monitor
		keyboard and mouse
		III. Consumables PER UNIT
	1	Five (5) packs [300 sheets/pack] Microscope lens paper for cleaning purposes
		IV. Warranty:
		➤ Three (3) years warranty on parts, accessories and services)
		➤ With semi-annual preventive maintenance
		V. With duly notarized certificate from Distributor - Omnibus Bio-Medical Systems, Inc.
		Certifies that:
		i. Parts, accessories and consumables are readily available at the authorized Philippine service center/s for a period of ten (10) years after the warranty period.
		ii. Guarantees availability of competent in-house technical specialists in handling and providing technical support as well as maintenance of the equipment.
		iii. Will conduct training for proper orientation and maintenance to end-users of the equipment upon delivery.
		iv. Will provide replacement/back-up unit while the delivered unit is being repaired during the warranty period

ANNEX "A"
NOA-2024-179
OMNIBUS BIO-MEDICAL SYSTEMS, INC.

QTY	UNIT	ITEM DESCRIPTION
		VI. With duly notarized certificate from Principal Manufacturer - Evident Scientific Singapore Pte. Ltd.
		<i>Certifies that:</i>
		i. That the Principal Manufacturer shall honor all the terms and conditions stated in the bidding documents per IRR of RA 9184 and the corresponding contract for the project, including in the event that change of dealership will occur during the duration of the contract up to the warranty and preventive maintenance period;
		ii. The expected useful life of the equipment under normal use is more than ten (10) years;
		iii. Guarantee availability of spare parts, accessories and consumables at least for the next five (5) years from testing, commissioning, acceptance and delivery;
		iv. That it has competence in handling and providing technical support as well as maintenance of the equipment;
		v. Consumer guidelines regarding disposal:
		> The product must not be disposed of as unsorted municipal waste but should be collected separately.
		VII. DELIVERY PERIOD: Thirty (30) calendar days upon receipt of Purchase Order/Notice to Proceed.

Conforme:

Authorized Signatory (Signature over printed name)

Designation/Position

Date _____