



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

**NOTICE OF AWARD**  
**NOA-2025-083**

**HELANDE KIDNEY CARE INC.**  
3rd Floor Jafer Bldg., No. 118 West Ave.,  
Quezon City  
Tel. No.: 8927-9516  
Email Add: [ascarandang@helande.ph](mailto:ascarandang@helande.ph)

Dear Sir / Madam:

This is to inform you that based on the result of the conducted Public Bidding for the project stated below under Invitation to Bid No. IB-2025-052 as per BAC Resolution No. R2025-03-169, your proposal was found to be responsive.

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST (Php)
1	unit	<b>Supply and Delivery of Blood purification machine (Hemoperfusion machine double pump)</b> <b>Brand/Model: Future F20 Jafron Biomedical Co., Ltd.</b> <b>(see Annex "A" for detailed specifications)</b>	<b>Php2,395,000.00</b>

You are hereby required to provide on or before 22 MAR 2025 the Performance Security in either of the following form:

FORM OF PERFORMANCE SECURITY	AMOUNT OF PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price)
a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank	Five percent (5%)
b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank	<b>Php119,750.00</b>
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%) <b>Php718,500.00</b>



**Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC.  
These unsanctioned requests are unlawful and will not be tolerated.**



**NOTICE OF AWARD**  
**NOA-2025-083**  
**HELANDE KIDNEY CARE INC.**

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

Very truly yours,

  
**MARIA EVA I. JOPSON, MD, MSChSM, MPM**  
**OIC, Executive Director** 

**Conforme:**

This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto.

\_\_\_\_\_  
Authorized Signatory (Signature over printed name)

\_\_\_\_\_  
Designation

\_\_\_\_\_  
Date

**Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC.  
These unsanctioned requests are unlawful and will not be tolerated.**



QTY	UNIT	ITEM DESCRIPTION
1	unit	<b>Supply and Delivery of Blood purification machine (Hemoperfusion machine double pump)</b> <b>Brand/Model: Future F20 Jafron Biomedical Co., Ltd.</b> <b>Total Cost = Php 2,395,000.00</b>
		<b>I. Specifications:</b>
	1	Dimension:
		Height: 156cm
		Width: 67cm
		Depth: 53cm
		Weight: ~48kg
	2	Blood pump speed, 15~400 ml/min
	3	Filtrate pump speed: 15~200 ml/min
	4	Heparin pump speed: 0.5 ~ 15ml/min
	5	Warmer range: 35~39 °C
	6	Arterial pressure measuring range: -500 to 500mm Hg
	7	Venous pressure measuring range: -500 to 500mm Hg
	8	Extramembranous pressure measuring range: -500 to 500mm Hg
	9	Tansmembrane pressure measuring range: -500 to 500mm Hg
	10	Alarm:
		Bubble detection principle: Ultrasound Minimum bubble colume: 20µL (at a flow rate of 200ml/min)
		Blood leakage detection principle: Optical Alarm limit: ≤0.35mL/min (if HCT ≥32%)
	11	Type: Rotatable touch LCD screen Size: 12.1"
	12	Sliding block rotating wheels
	13	The data that can be stored and reviewed includes pressure data, heater temperature data, and alarm lists. It can store data within 120 hours.
	14	Power Specifications: Input voltage: AC 220V±22V,50Hz±1Hz Input power: ≤350VA
	15	Battery: Rechargeable lithium battery: 1 Use time: After being fully charged, it supports the running time of the blood pump > 30 min
	16	No separate volt conversion
	17	Power consumption: < 350VA (per brochure)
	18	Mechanical part is of heavy-duty durable type
	19	Equipment is maintainable and serviceable



QTY	UNIT	ITEM DESCRIPTION
		<b>II. Accessories PER UNIT</b>
	1	One (1) unit AVR (capacity will be based on the rating of the unit)
		<b>Brand/Model: STAVOL ST-3000VA</b>
	2	One (1) unit 1000 VA UPS
		<b>Brand/Model: AWP AID1000ECO</b>
		Will provide safety electrical devices compatible to the equipment power requirement) and proper grounding when needed
		<b>III. Consumables PER UNIT</b>
	1	Extracorporeal blood circuit
	2	Plasma separator/ plasma filter
		<b>IV. Warranty:</b>
	>	Three (3) years warranty on services, parts and accessories
	>	With software upgrade (free of charge)
	>	Quarterly preventive maintenance and calibration
		<b>V. With duly notarized certificate from Distributor - HELANDE KIDNEY CARE INC.</b>
		<i>Certifies that:</i>
		i. Parts, accessories and consumables are readily available at the authorized Philippine service center/s for a period of five (5) years after the warranty period.
		ii. Guarantees availability of competent in-house technical specialists in handling and providing technical support as well as maintenance of the equipment.
		iii. Will conduct training for proper orientation and maintenance to end-users of the equipment upon delivery.
		iv. It will provide replacement/back-up unit while the delivered unit is being repaired during the warranty period
		<b>VI. With duly notarized certificate from Principal Manufacturer - JAFRON BIOMEDICAL CO. LTD.</b>
		<i>Certifies that:</i>
		i. That the Principal Manufacturer shall honor all the terms and conditions stated in the bidding documents per IRR of RA 9184 and the corresponding contract for the project, including in the event that change of dealership will occur during the duration of the contract up to the warranty and preventive maintenance period;
		ii. The expected useful life of the equipment under normal use is eight (8) years
		iii. Guarantee availability of spare parts, accessories and consumables at least for the next five (5) years from testing, commissioning, acceptance and delivery;
		iv. That it has competence in handling and providing technical support as well as maintenance of the equipment;
		v. Consumer guidelines regarding disposal:

ANNEX "A"  
NOA-2025-083  
**HELANDE KIDNEY CARE INC.**

QTY	UNIT	ITEM DESCRIPTION
		<i>When the equipment reaches its end of life, please contact us for the recycling.</i>
		<b>V. DELIVERY PERIOD:</b> Thirty to Forty-Five (30 to 45) calendar days upon receipt of Purchase Order/Notice to Proceed.

*Conforme:*

\_\_\_\_\_  
Authorized Signatory (Signature over printed name)

\_\_\_\_\_  
Designation/Position

Date \_\_\_\_\_

