



Republic of the Philippines
DEPARTMENT OF HEALTH
PHILIPPINE CHILDREN'S MEDICAL CENTER

Quezon Avenue, Quezon City 1100
website: www.pcmc.gov.ph email: officeofthedirector@pcmc.gov.ph
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

NOTICE OF AWARD
NOA-2025-092

IDS MEDICAL SYSTEM PHILIPPINES, INC

Unit 1010, 10th Flr, BTTC Centre
No. 288 Ortigas Ave., cor. Roosevelt St.
Greenhills, San Juan City
Tel No.: 8737-9898

Dear Sir / Madam:

This is to inform you that based on the result of the conducted Public Bidding for the project stated below under Invitation to Bid No. **IB-2025-063** as per BAC Resolution No **R2025-03-194** , your proposal was found to be responsive.

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST (Php)
1	unit	SUPPLY AND DELIVERY OF MRI COMPATIBLE ANESTHESIA MACHINE with MONITOR for CANCER CENTER Brand: GE HEALTHCARE AESTIVA MRI + CARESCAPE B650 (see Annex "A" for detailed specifications)	Php7,880,000.00

You are hereby required to provide on or before 12 APR 2025 the Performance Security in either of the following form:

FORM OF PERFORMANCE SECURITY	PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price)
a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank	Five percent (5%) <u>Php394,000.00</u>
b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%) <u>Php2,364,000.00</u>

Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC. These unsanctioned requests are unlawful and will not be tolerated



NOTICE OF AWARD
NOA-2025-092
IDS MEDICAL SYSTEM PHILIPPINES, INC

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

Very truly yours,


MARIA EVA I. JOPSON, MD MScHSM, MPM
OIC- Executive Director 

Conforme:

This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto.

Authorized Signatory (Signature over printed name)

Designation

Date

Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC. These unsanctioned requests are unlawful and will not be tolerated

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST
1	unit	SUPPLY AND DELIVERY OF MRI COMPATIBLE ANESTHESIA MACHINE with MONITOR for CANCER CENTER	Php7,880,000.00
		BRAND: GE HealthCare	
		MAKE / MODEL: Aestiva MRI + CARESCAPE B650	
		<u>A. Specifications</u>	
		1. 15" full colour display utilising surface acoustic wave touch screen technology	
		2. Innovative patient breathing system / Integrated sensing mechanism, CO2 bypass capability	
		3. Dual air flow tube for resolution of low flows	
		4. Tool-free installation Selectatec® manifold interlocks and isolates vaporizers	
		5. Control Lighting: Front bezel indicators: Green light, Amber light, Red light	
		6. Gas Supply: Pipeline input, Cylinder , Primary regulator input,	
		7. Pipeline input range: 240 kPa to 600 kPa/35 psig to 88 psig	
		8. Pipeline connections: DISS-male, DISS-female, DIN 13252, AS4059, F90-116, or NIST (ISO 5359). All fittings available for O2, N2O and Air, and contain pipeline filter and check valve	
		9. Cylinder input: Pin indexed: Pin indexed in accordance with CGA-V-1 or DIN (nut and gland); contains input filter and check valve	
		10. Modes of ventilation: VC, PC, SIMV , PSV PRO	
		11. Tidal volume range: 5-1500ml	
		12. Minute volume range: 0 to 99.9 L/min	
		13. Respiratory Rate: 4 to 100 breaths per minute for Volume Control	
		14. I:E ratio: 2:1 to 1:8 (increments of 0.5)	
		15. Validated for use in MRI environments up to 300 Gauss, active shielded 1.5T and 3T magnets	
		16. With storage drawers : Drawer (1 standard)* — locking (international dimensions)	
		17. Single foot lever locks and unlocks two front casters	
		18. Responsive location of common gas outlet	
		19. Anesthetic agent delivery system compatible to vaporizer Isoflurane and Sevoflurane, 2 positions : 2 tool-free installation Selectatec® manifold interlocks and isolates vaporizers	
		20. Integrated Adjustable Pressure Limiting (APL) valve	
		21. integrated sensing mechanism, CO2 bypass capability	

QTY	UNIT	ITEM DESCRIPTION		TOTAL COST
1	unit	SUPPLY AND DELIVERY OF MRI COMPATIBLE ANESTHESIA MACHINE with MONITOR for CANCER CENTER		Php7,880,000.00
		22.	Quick power up for fast paced environment: Power input: USA/Canada/Mexico: 120 Vac, 60 Hz, 10A Euro: 220-240 Vac, 50 Hz, 6A	
		II.	ANESTHESIA MONITOR	
		1.	Multi-parameter modular type of monitor - ECG, SpO2, NIBP, RR, temperature (with 2 module slots) BIS module, with advanced airway modules for complete respiratory monitoring with gas exchange measurement	
		2.	Central monitoring / station ready	
		3.	Touch screen Resistive technology, Trim Knob™ Trim Knob control with clockwise, counter-clockwise and push functionality	
		4.	Alarm system with three-level priority prompt, with alarm volume	
		5.	Provision for the use of Oxygen tank pin index and host connector for oxygen tank	
		6.	Back up water trap located at the back of the Anesthesia Machine	
		III.	POWER SPECIFICATION	
		1.	Power: 100-240 Vac. 50/60Hz (Auto volt)	
		2.	Backup power: Demonstrated battery time under typical operating conditions is 90+ minutes when anaesthesia machine is fully charged. Battery time under extreme conditions is 30 minutes with monitor.	
		3.	Batter: Ideal for transport	
		IV.	Other Requirement per Engineering Clearance	
		1.	The unit will be operated at 230 Volts 3-Pin Power Plug / cable, strictly at 60 hertz	
		2.	With built-in battery back up system for detection and life saving equipment.	
		3.	No separate volt conversion	
		4.	Power consumption : 2400Watts - P26.72 Meralco cost/Hour	
		5.	Mechanical parts are heavy-duty durable type	
		6.	Equipment is maintainable and serviceable	
		B. Accessories (to be provided per unit)		
		1.	Sampling line	
		2.	D fend water trap pro	
		3.	5 leads ECG	
		4.	BP Cuff must have the following:	
			4.1 Small Adult	
			4.2 Large Adult	

QTY	UNIT	ITEM DESCRIPTION		TOTAL COST
1	unit	SUPPLY AND DELIVERY OF MRI COMPATIBLE ANESTHESIA MACHINE with MONITOR for CANCER CENTER		Php7,880,000.00
			4.3 Pediatric	
			4.4 Medium Size	
			4.5 Thigh size, skin and general	
		5.	Temperature: Esophageal and rectal probe, bis headbox and censor	
		6.	Sidestream CO2	
		7.	Provide back-up battery for transport	
		8	TVSS (Brand/Model: PROTEC PRO T)	
		9	2KVA UPS (Brand/Model: Abrelex Ares)	
		C. Warranty		
		1.	Two (2) years warranty for the main equipment parts	
		2.	Five (5) years warranty on service	
		3.	Quarterly preventive maintenance and calibration of the machine during warranty period	
		E. With duly notarized certificate from Distributor - IDS MEDICAL SYSTEMS PHILIPPINES, INC.		
		Certifes that:		
		i. Parts, accessories and consumables are readily available at the authorized Philippine service center/s for a period of five (5) years after the warranty period provided that IDSMED remains the authorized distributor of the subject product.		
		ii. Has available competence in-house technical specialist in handling and providing technical support as well as maintenance of the equipment being offered; and		
		iii. That it will conduct training for proper operation and maintenance to end-users of the equipment upon delivery; and		
		iv. Will provide replacement/back-up unit while the delivered unit is being repaired.		
		F. With duly notarized certificate from Principal Manufacturer- GENERAL ELECTRIC PHILIPPINES, INC.		
		Certifies that:		
		i. All the terms and conditions stated in bidding documents per IRR of RA 9184 and corresponding contract for the project will be honored including in the event that a change of distributorship will occur during the duration of the contract		
		ii. The expected useful /service life of the equipment under normal use is ten (10) Years		

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST
1	unit	SUPPLY AND DELIVERY OF MRI COMPATIBLE ANESTHESIA MACHINE with MONITOR for CANCER CENTER	Php7,880,000.00
		iii. Guarantee on availability of all spare parts, accessories and consumables at least for the next five (5) years from testing, commissioning, acceptance and delivery;	
		iv. That it has competence in handling and providing technical support as well as maintenance of the equipment being offered; and	
		v. Consumer to dispose of the equipment and its parts in accordance with in-country regulatory requirements. GE does not dispose of parts or equipment for consumers; consumer is responsible to dispose of the used non-returnable parts/equipment/any other waste generated during service work	
		DELIVERY PERIOD: Forty Five (45) Calendar Days upon receipt of Purchase Order / Notice to Proceed.	

Conforme:

Authorized Signatory (Signature over printed name)

Designation/Position

Date _____

