



Republic of the Philippines  
DEPARTMENT OF HEALTH  
**PHILIPPINE CHILDREN'S MEDICAL CENTER**

Quezon Avenue, Quezon City 1100  
website: [www.pcmc.gov.ph](http://www.pcmc.gov.ph) email: [officeofthedirector@pcmc.gov.ph](mailto:officeofthedirector@pcmc.gov.ph)  
Trunk Line: 8588-9900 to 20 Direct Line: 8924-6601

**NOTICE OF AWARD**

**NOA-2025-106**

**M2M MEDICAL TRADING INC**

I Coronado St., San Isidro Vista Verde  
Executive Village, Cainta, Rizal  
Tel No : +6309089283587  
Email Add: [jcatapusan@m2mmedical.com](mailto:jcatapusan@m2mmedical.com)

Dear Sir / Madam:

This is to inform you that based on the result of the conducted Competitive Bidding for the project stated below under Invitation to Bid No. IB-2025-068 as per BAC Resolution No. R2025-03-227, your proposal was found to be responsive.

QTY	UNIT	ITEM DESCRIPTION	TOTAL COST (Php)
1	unit	Supply and Delivery of Defibrillator with 3 sets of Paddle for Adult, Pedia and Neonate Brand/Model: NIHON KOHDEN TEC5631 (see Annex "A" for detailed specifications)	Php1,498,000.00

You are hereby required to provide on or before 25 APR 2025 the Performance Security in either of the following form:

FORM OF PERFORMANCE SECURITY	AMOUNT OF PERFORMANCE SECURITY (Equal to Percentage of the Total Contract Price)
a) Cash or cashier's/ manager's check issued by a Universal or Commercial Bank	Five percent (5%) <b>Php74,900.00</b>
b.) Bank draft/guarantee or irrevocable letter of credit issued by a Universal or Commercial Bank: Provided, however, that it shall be confirmed or authenticated by a Universal or Commercial Bank, if issued by a foreign bank	
c) Surety bond callable upon demand issued by a surety or insurance company duly certified by the Insurance Commission as authorized to issue such security.	Thirty percent (30%) <b>Php449,400.00</b>



*Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC.  
These unsanctioned requests are unlawful and will not be tolerated*



**NOTICE OF AWARD**  
**NOA-2025-106**  
**M2M MEDICAL TRADING INC**

Failure to provide the Performance Security shall constitute sufficient ground for cancellation of the award and imposition of penalties/ forfeiture corresponding to the Bid Security posted.

Very truly yours,

  
**MARIA EVA I. JOPSON, MD, MScHSM, MPM**  
Officer-In-Charge, Executive Director 

-----

**Conforme:**

This is to certify that the company has authorized me to accept this award, sign all related documents and hold the company bound by rules and laws applicable thereto.

\_\_\_\_\_  
Authorized Signatory (Signature over printed name)

\_\_\_\_\_  
Designation

\_\_\_\_\_  
Date

*Kindly disregard and report any form of solicitation using the name of the Executive Director and/or the PCMC.  
These unsanctioned requests are unlawful and will not be tolerated*

UNIT	ITEM DESCRIPTION	
unit	<b>Supply and Delivery of One (1) Unit Defibrillator with Three (3) sets of Paddle for Adult, Pedia and Neonate</b>	
	Brand: NIHON KOHDEN DEFIBRILLATOR	
	Make/Model: TEC5631	
	Total Amount Awarded = Php 1,498,000.00	
	<b>A. Specifications of the Main Equipment</b>	
	1.	Multiple Modes: Manual defibrillation, Cardioversion, AED, SYNC, Non-Invasive Pacing, Heart Rate and ECG Monitoring, Non- Invasive Blood Pressure and CO2 Monitoring
	2.	Display Parameters: ECG and size, HR, Pulse Rate, Respiration Rate, NIBP, SPO2, ETC02)
	3.	Display Indications: Analyzing, Charging, Charge, Time of discharge after power ON, Same message as voice in AED mode, SYNC, AED mode, AED mode (Child), Fixed mode, Demand mode, Pacing spike, Pacing rate, Pacing current, (AC operation), Lead select, Battery mark (remaining operation time), Alarm, TIME, Connection strength for Bluetooth, Recorder status, SD card status)
	4.	Display: 6.5 inches, Color TFT LCD, Viewing area 5 x 4 inches (WxH) Resolution at least 640 x 480 pixels (HxV), Luminance at least 1000 cd/m2, Number of tracings at least 4
	5.	Biphasic, truncated exponential constant power
	6.	Intuitive multi-step operation design
	7.	Dimension: Compact and portable with height 12.2 x 11.3 x 9.5 inches (WxHxD) and weight 6.9 kgs (including battery and external paddle
	8.	External Paddles: Removable and with adult and pediatric paddles
	9.	Built-in Printer
	10.	Parameters
		3 and 6 lead ECG
		Heart rate count range: 15-300 bpm (defibrillation mode and monitor mode) and 15-220 bpm (pacing mode)
		Pulse rate count range: 30 to 300 bpm
		Baseline recovery time less than 3 seconds after 270J defibrillation)
		Non-invasive pacing: Demand or fixed mode selectable)
		Pacing pulse rejection available)
		QRS detection level input conversion 0.5 mV or more (at x1 sensitivity)
		Sensitivity control 1/8, 1/4, 1/2, 1, 2, 4, 8, AUTO)
		AED with Adult/Child mode available with voice)
		Upgradable Unit
	11.	<b>Defibrillation</b>
		Output Energy External Paddle: 2, 3, 5, 7, 10, 15, 20, 30, 50, 70, 100, 150, 200 up to 270J)
		Output Energy Internal Paddle: 2, 3, 5, 7, 10, 15, 20, 30 up to 50J)



UNIT	ITEM DESCRIPTION	
unit	<b>Supply and Delivery of One (1) Unit Defibrillator with Three (3) sets of Paddle for Adult, Pedia and Neonate</b>	
		Disposable Pacing Pad: 2, 3, 5, 7, 10, 15, 20, 30, 50, 70, 100, 150, 200 up to 270J)
		AED mode (Adult): 1st 150J, 2nd 200J, 3rd 200J)
		AED mode (Child): 1st 50J, 2nd 70J, 3rd 70J)
		AED mode (Child): 1st 50J, 2nd 70J, 3rd 70J)
		Charging Time: Less than 4 seconds at 200J and 5 seconds at 270J (both AC and fully charged battery operation)
	12	Recorder
		Recorded Waveforms: ECG and when available, ECG+SPO2, ECG+CO2)
		Recording Modes: Manual recording, Real time recording, Delayed recording, Event recording, Trend report, Self test result, Basic check result, Operation history/Device condition history, Automatic recording alarm report, 3 Defibrillation recording, Periodic list report)
		Thermal printer: 50 mm roll)
	13	Battery and Operation Time
		Ni-MH battery, voltage 12.0V (9.0 to 18.0V), capacity 2800mAh)
		With fully charged new battery at 20C : min 100 discharges at 270J or min 180 minutes continuous monitoring or min 120 minutes pacing)
		Charging time not >3 hours (with AC and power off)
	14	Quality and Maintenance
		Hospital grade equipment
		Generic thermal printer paper
		Easily cleaned with standard sterilizing materials
	15	Quality and Maintenance
	16	The unit will be operated at 100-240 volts, 3-Pin Power Plug/Cable, strictly at 60 hertz. (If the unit/s is 50/60Hz, provide UPS with frequency converter when needed)
	17	With built-in battery for detection and life saving equipment and no separate voltage conversion)
	18	Power consumption : 150 Watts
	19	Mechanical parts should be of heavy-duty type
	20	Equipment is maintainable and serviceable
<b>B. Consumables to be provided</b>		
	1.	Thermal paper roll: 20 rolls)
	2.	Electrodes: Adult 50 pcs, Pedia 100 pcs)
	3.	SD Card 1TB
<b>C. Accessories to be provided</b>		
	1	ECG connector and leads clip type ( 2 sets)
	2	NIBP hose, NIBP cuff for infant, pedia and adult
	3	Spo2 Sensor for infant, pedia and adult)
	4	Mainstream ETCO2 Sensor for instubated and non-intubated patient, lightweight not morethan 5g
	5	Starter Kit:
		1 set AED cable
		5 pcs adult pad
		5 pcs pedia pad
	6	Internal paddle electrode, 35 mm dia, 1 pair/set
	7	Internal paddle electrode, 75 mm dia, 1 pair/set
	8	Internal paddle electrode, 55 mm dia, 1 pair/set

UNIT	ITEM DESCRIPTION	
unit	<b>Supply and Delivery of One (1) Unit Defibrillator with Three (3) sets of Paddle for Adult, Pedia and Neonate</b>	
	9	ECG Cable
	10	Hospital grade trolley cart with die cast aluminum base with height of 110 cms, sturdy slide-in/mount/angled support custom made for the size and weight of the defibrillator machine, balanced 4 wheel roll stand, silent caster wheels with brakes and metal basket)
<b>D. Warranty</b>		
	1.	Two (2) years on parts and services
	2	Quarterly preventive maintenance on warranty period
<b>E. With duly notarized certificate from Bidder - M2M MEDICAL TRADING INC.</b>		
	<b>Certifies that:</b>	
	i.	Parts, accessories, and consumables are readily available at the authorized Philippine service center for a period of five (5) years after the warranty period;
	ii.	It has available competent in-house technical specialists in handling and providing technical support as well as maintenance of the equipment being offered;
	iii.	It will conduct training for proper operation and maintenance to the End-users of the equipment upon delivery; and
	iv.	It will provide replacement/back-up unit while the delivered unit is being repaired.
<b>F. With duly notarized certificate from Principal Manufacturer - NIHON KOHDEN</b>		
	<b>Singapore PTE LTD</b>	
	<b>Certifies that:</b>	
	i.	Guaranteed that all the terms and conditions stated in the bidding documents per IRR of RA 9184 and the corresponding contract for the contract for the project shall be honored by the Principal Manufacturer, including in the event that a change of dealership will occur during the duration of the contract up to the warranty and preventive maintenance period.
	ii.	The expected useful life of the equipment under normal use is 6 years
	iii.	Guarantee on the availability of all spare parts, accessories and consumables at least for the next five (5) years from testing, commissioning, acceptance and delivery;
	iv.	That M2M Medical Trading has competence in handling and providing technical support as well as maintenance of the equipment being offered; and
	v.	Dispose of Nihon Kohden products according to your local laws and your facility's guidelines for waste disposal. Otherwise it may affect the environment. If there is a possibility that the product may have been contaminated with infection, dispose of it as medical waste according to your local laws and your facility's guidelines for medical waste. Otherwise, it may cause infection.
<b>G. Delivery Period:</b>		
	1.	Within Thirty (30) Calendar days upon receipt of the Purchase Order/Notice to Proceed

Conforme:

\_\_\_\_\_  
Authorized Signatory (Signature over printed name)

\_\_\_\_\_  
Designation/Position

Date \_\_\_\_\_

*CM*